

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102621\
 Data File : BF125924.D
 Acq On : 26 Oct 2021 9:19
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 26 11:59:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 25 18:49:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	113	0.00
2	1,4-Dioxane	40.000	37.983	5.0	113	0.00
3	Pyridine	40.000	38.474	3.8	108	0.00
4	n-Nitrosodimethylamine	40.000	37.769	5.6	107	0.00
5 S	2-Fluorophenol	80.000	74.785	6.5	111	0.00
6	Aniline	40.000	37.942	5.1	108	0.00
7 S	Phenol-d6	80.000	71.356	10.8	110	0.00
8	2-Chlorophenol	40.000	38.037	4.9	110	0.00
9	Benzaldehyde	40.000	37.107	7.2	111	0.00
10 C	Phenol	40.000	38.273	4.3	111	0.00
11	bis(2-Chloroethyl)ether	40.000	37.712	5.7	109	0.00
12	1,3-Dichlorobenzene	40.000	37.502	6.2	109	0.00
13 C	1,4-Dichlorobenzene	40.000	37.173	7.1	109	0.00
14	1,2-Dichlorobenzene	40.000	35.181	12.0	108	0.00
15	Benzyl Alcohol	40.000	37.862	5.3	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.462	6.3	108	0.00
17	2-Methylphenol	40.000	38.449	3.9	111	0.00
18	Hexachloroethane	40.000	38.376	4.1	112	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.040	4.9	112	0.00
20	3+4-Methylphenols	40.000	35.672	10.8	108	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	112	0.00
22	Acetophenone	40.000	36.970	7.6	108	0.00
23 S	Nitrobenzene-d5	80.000	80.046	-0.1	110	0.00
24	Nitrobenzene	40.000	40.144	-0.4	110	0.00
25	Isophorone	40.000	38.600	3.5	111	0.00
26 C	2-Nitrophenol	40.000	43.960	-9.9	111	0.00
27	2,4-Dimethylphenol	40.000	38.182	4.5	110	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.720	5.7	111	0.00
29 C	2,4-Dichlorophenol	40.000	38.795	3.0	110	0.00
30	1,2,4-Trichlorobenzene	40.000	37.965	5.1	109	0.00
31	Naphthalene	40.000	37.562	6.1	110	0.00
32	Benzoic acid	40.000	38.894	2.8	111	0.00
33	4-Chloroaniline	40.000	38.106	4.7	111	0.00
34 C	Hexachlorobutadiene	40.000	38.431	3.9	111	0.00
35	Caprolactam	40.000	38.539	3.7	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.669	3.3	111	0.00
37	2-Methylnaphthalene	40.000	37.282	6.8	110	0.00
38	1-Methylnaphthalene	40.000	38.009	5.0	111	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.469	3.8	108	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.426	-3.6	110	0.00
42 S	2,4,6-Tribromophenol	80.000	78.373	2.0	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.678	-1.7	111	0.00
44	2,4,5-Trichlorophenol	40.000	39.454	1.4	107	0.00
45 S	2-Fluorobiphenyl	80.000	78.644	1.7	108	0.00
46	1,1'-Biphenyl	40.000	38.394	4.0	108	0.00
47	2-Chloronaphthalene	40.000	39.626	0.9	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102621\
 Data File : BF125924.D
 Acq On : 26 Oct 2021 9:19
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 26 11:59:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 25 18:49:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.484	-11.2	112	0.00
49	Acenaphthylene	40.000	38.408	4.0	108	0.00
50	Dimethylphthalate	40.000	38.709	3.2	107	0.00
51	2,6-Dinitrotoluene	40.000	42.936	-7.3	111	0.00
52 C	Acenaphthene	40.000	38.644	3.4	110	0.00
53	3-Nitroaniline	40.000	41.318	-3.3	108	0.00
54 P	2,4-Dinitrophenol	40.000	38.858	2.9	109	0.00
55	Dibenzofuran	40.000	37.977	5.1	108	0.00
56 P	4-Nitrophenol	40.000	41.609	-4.0	105	0.00
57	2,4-Dinitrotoluene	40.000	44.262	-10.7	109	0.00
58	Fluorene	40.000	40.761	-1.9	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.907	0.2	110	0.00
60	Diethylphthalate	40.000	39.422	1.4	109	0.00
61	4-Chlorophenyl-phenylether	40.000	35.882	10.3	109	0.00
62	4-Nitroaniline	40.000	40.212	-0.5	106	0.00
63	Azobenzene	40.000	39.239	1.9	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.814	-2.0	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.439	1.4	108	0.00
67	4-Bromophenyl-phenylether	40.000	38.945	2.6	106	0.00
68	Hexachlorobenzene	40.000	39.156	2.1	108	0.00
69	Atrazine	40.000	39.048	2.4	102	0.00
70 C	Pentachlorophenol	40.000	41.135	-2.8	104	0.00
71	Phenanthrene	40.000	38.235	4.4	104	0.00
72	Anthracene	40.000	38.005	5.0	104	0.00
73	Carbazole	40.000	37.233	6.9	101	0.00
74	Di-n-butylphthalate	40.000	38.671	3.3	101	0.00
75 C	Fluoranthene	40.000	36.300	9.3	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	91	0.00
77	Benzydine	40.000	43.026	-7.6	104	0.00
78	Pyrene	40.000	41.162	-2.9	95	0.00
79 S	Terphenyl-d14	80.000	82.043	-2.6	98	0.00
80	Butylbenzylphthalate	40.000	44.080	-10.2	97	0.00
81	Benzo(a)anthracene	40.000	39.528	1.2	93	0.00
82	3,3'-Dichlorobenzidine	40.000	40.514	-1.3	96	0.00
83	Chrysene	40.000	39.663	0.8	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.844	-9.6	101	0.00
85 c	Di-n-octyl phthalate	40.000	46.796	-17.0	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.695	-14.2	113	0.00
88	Benzo(b)fluoranthene	40.000	36.575	8.6	97	0.00
89	Benzo(k)fluoranthene	40.000	40.636	-1.6	113	0.00
90 C	Benzo(a)pyrene	40.000	39.920	0.2	105	0.00
91	Dibenzo(a,h)anthracene	40.000	46.174	-15.4	115	0.00
92	Benzo(g,h,i)perylene	40.000	46.958	-17.4	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102621\
Data File : BF125924.D
Acq On : 26 Oct 2021 9:19
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 26 11:59:48 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 25 18:49:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0