

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102622\
 Data File : BF130788.D
 Acq On : 26 Oct 2022 17:48
 Operator : CG\JU
 Sample : N5281-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-5S

Quant Time: Oct 27 09:04:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 27 08:55:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.563	152	138714	20.000	ng	0.00
21) Naphthalene-d8	7.845	136	566382	20.000	ng	# 0.00
39) Acenaphthene-d10	9.592	164	299083	20.000	ng	0.00
64) Phenanthrene-d10	11.074	188	523084	20.000	ng	# 0.00
76) Chrysene-d12	13.710	240	327652	20.000	ng	# 0.00
86) Perylene-d12	15.098	264	243218	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.181	112	836732	104.017	ng	0.00
7) Phenol-d6	6.222	99	1026247	105.442	ng	0.00
23) Nitrobenzene-d5	7.139	82	615622	57.660	ng	0.00
42) 2,4,6-Tribromophenol	10.386	330	256650	90.234	ng	0.00
45) 2-Fluorobiphenyl	8.922	172	1112095	56.665	ng	0.00
79) Terphenyl-d14	12.657	244	1162810	59.806	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102622\
Data File : BF130788.D
Acq On : 26 Oct 2022 17:48
Operator : CG\JU
Sample : N5281-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-5S

Quant Time: Oct 27 09:04:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 27 08:55:38 2022
Response via : Initial Calibration

