

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102624\
Data File : BF140069.D
Acq On : 26 Oct 2024 19:37
Operator : RC/JU
Sample : P4547-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BP-F-21

Quant Time: Oct 28 01:10:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.886	152	148843	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	596055	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	343363	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	568427	20.000	ng	0.00
76) Chrysene-d12	14.045	240	274542	20.000	ng	0.00
86) Perylene-d12	15.527	264	300909	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	655948	68.991	ng	0.00
7) Phenol-d6	6.504	99	857229	69.614	ng	0.00
23) Nitrobenzene-d5	7.445	82	550237	51.163	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	233948	72.842	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1017396	48.970	ng	0.00
79) Terphenyl-d14	12.992	244	870683	51.677	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102624\
Data File : BF140069.D
Acq On : 26 Oct 2024 19:37
Operator : RC/JU
Sample : P4547-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BP-F-21

Quant Time: Oct 28 01:10:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

