

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/26/2024	BNA_F	<u>BF140049.D</u>
Compound Name	Response	Retention Time
DDT	384498	13.574
DDD	9181	13.274
DDE	1042	12.939
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10223	394721	2.59