

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102720\
 Data File : BF122137.D
 Acq On : 27 Oct 2020 11:36
 Operator : JU/CG
 Sample : L4469-01RE
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FY-WATERRE

Quant Time: Oct 27 12:13:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 26 17:27:32 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.710	152	107564	20.00	ng	0.00
21) Naphthalene-d8	7.998	136	432916	20.00	ng	0.00
39) Acenaphthene-d10	9.745	164	228617	20.00	ng	0.00
64) Phenanthrene-d10	11.233	188	434936	20.00	ng	0.00
76) Chrysene-d12	13.880	240	373662	20.00	ng	0.00
86) Perylene-d12	15.310	264	304450	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.346	112	95000	13.04	ng	0.01
7) Phenol-d6	6.375	99	115897	12.35	ng	0.00
23) Nitrobenzene-d5	7.281	82	91048	10.79	ng	0.00
42) 2,4,6-Tribromophenol	10.545	330	18692	6.61	ng	0.00
45) 2-Fluorobiphenyl	9.069	172	100888	6.49	ng	0.00
79) Terphenyl-d14	12.816	244	116017	5.92	ng	0.00
Target Compounds						
10) Phenol	6.387	94	103295	10.67	ng	84
50) Dimethylphthalate	9.469	163	99699	5.96	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102720\
Data File : BF122137.D
Acq On : 27 Oct 2020 11:36
Operator : JU/CG
Sample : L4469-01RE
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FY-WATERRE

Quant Time: Oct 27 12:13:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 26 17:27:32 2020
Response via : Initial Calibration

