

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102720\
 Data File : BF122134.D
 Acq On : 27 Oct 2020 9:28
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 27 10:50:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 26 17:27:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	90	0.00
2	1,4-Dioxane	40.000	37.616	6.0	86	0.00
3	Pyridine	40.000	38.036	4.9	85	0.00
4	n-Nitrosodimethylamine	40.000	39.216	2.0	88	0.00
5 S	2-Fluorophenol	80.000	79.935	0.1	92	0.00
6	Aniline	40.000	37.080	7.3	86	0.00
7 S	Phenol-d6	80.000	74.850	6.4	87	0.00
8	2-Chlorophenol	40.000	39.473	1.3	91	0.00
9	Benzaldehyde	40.000	40.111	-0.3	90	0.00
10 C	Phenol	40.000	38.020	4.9	89	0.00
11	bis(2-Chloroethyl)ether	40.000	38.317	4.2	87	0.00
12	1,3-Dichlorobenzene	40.000	38.474	3.8	89	0.00
13 C	1,4-Dichlorobenzene	40.000	38.763	3.1	89	0.00
14	1,2-Dichlorobenzene	40.000	38.867	2.8	89	0.00
15	Benzyl Alcohol	40.000	41.174	-2.9	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.227	4.4	87	0.00
17	2-Methylphenol	40.000	37.527	6.2	87	0.00
18	Hexachloroethane	40.000	40.005	-0.0	91	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.454	-1.1	95	0.00
20	3+4-Methylphenols	40.000	40.835	-2.1	96	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	40.046	-0.1	95	0.00
23 S	Nitrobenzene-d5	80.000	80.211	-0.3	93	0.00
24	Nitrobenzene	40.000	39.602	1.0	91	0.00
25	Isophorone	40.000	39.842	0.4	93	0.00
26 C	2-Nitrophenol	40.000	40.620	-1.5	91	0.00
27	2,4-Dimethylphenol	40.000	38.425	3.9	89	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.722	0.7	92	0.00
29 C	2,4-Dichlorophenol	40.000	40.336	-0.8	93	0.00
30	1,2,4-Trichlorobenzene	40.000	39.014	2.5	91	0.00
31	Naphthalene	40.000	39.161	2.1	92	0.00
32	Benzoic acid	40.000	35.122	12.2	82	0.01
33	4-Chloroaniline	40.000	39.977	0.1	93	0.00
34 C	Hexachlorobutadiene	40.000	39.950	0.1	91	0.00
35	Caprolactam	40.000	41.452	-3.6	96	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.648	-1.6	94	0.00
37	2-Methylnaphthalene	40.000	39.444	1.4	93	0.00
38	1-Methylnaphthalene	40.000	39.218	2.0	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.390	4.0	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.428	11.4	77	0.00
42 S	2,4,6-Tribromophenol	80.000	81.731	-2.2	97	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.772	0.6	95	0.00
44	2,4,5-Trichlorophenol	40.000	38.754	3.1	93	0.00
45 S	2-Fluorobiphenyl	80.000	74.877	6.4	94	0.00
46	1,1'-Biphenyl	40.000	38.404	4.0	94	0.00
47	2-Chloronaphthalene	40.000	38.201	4.5	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102720\
 Data File : BF122134.D
 Acq On : 27 Oct 2020 9:28
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 27 10:50:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 26 17:27:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.851	-2.1	97	0.00
49	Acenaphthylene	40.000	37.568	6.1	91	0.00
50	Dimethylphthalate	40.000	39.217	2.0	95	0.00
51	2,6-Dinitrotoluene	40.000	40.040	-0.1	94	0.00
52 C	Acenaphthene	40.000	38.651	3.4	95	0.00
53	3-Nitroaniline	40.000	39.939	0.2	96	0.00
54 P	2,4-Dinitrophenol	40.000	36.779	8.1	92	0.00
55	Dibenzofuran	40.000	38.419	4.0	94	0.00
56 P	4-Nitrophenol	40.000	40.186	-0.5	94	0.00
57	2,4-Dinitrotoluene	40.000	40.704	-1.8	97	0.00
58	Fluorene	40.000	39.568	1.1	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.115	-5.3	97	0.00
60	Diethylphthalate	40.000	40.738	-1.8	98	0.00
61	4-Chlorophenyl-phenylether	40.000	40.057	-0.1	98	0.00
62	4-Nitroaniline	40.000	37.419	6.5	90	0.00
63	Azobenzene	40.000	40.162	-0.4	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.862	5.3	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.786	3.0	97	0.00
67	4-Bromophenyl-phenylether	40.000	39.425	1.4	96	0.00
68	Hexachlorobenzene	40.000	38.947	2.6	96	0.00
69	Atrazine	40.000	41.440	-3.6	101	0.00
70 C	Pentachlorophenol	40.000	33.410	16.5	85	0.00
71	Phenanthrene	40.000	38.403	4.0	96	0.00
72	Anthracene	40.000	38.345	4.1	96	0.00
73	Carbazole	40.000	38.740	3.1	96	0.00
74	Di-n-butylphthalate	40.000	40.401	-1.0	99	0.00
75 C	Fluoranthene	40.000	38.105	4.7	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	38.258	4.4	93	0.00
78	Pyrene	40.000	40.517	-1.3	96	0.00
79 S	Terphenyl-d14	80.000	81.740	-2.2	99	0.00
80	Butylbenzylphthalate	40.000	42.721	-6.8	99	0.00
81	Benzo(a)anthracene	40.000	40.473	-1.2	97	0.00
82	3,3'-Dichlorobenzidine	40.000	38.893	2.8	90	0.00
83	Chrysene	40.000	39.095	2.3	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.814	-7.0	99	0.00
85 c	Di-n-octyl phthalate	40.000	41.729	-4.3	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	88	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.382	1.5	88	0.01
88	Benzo(b)fluoranthene	40.000	41.216	-3.0	87	0.00
89	Benzo(k)fluoranthene	40.000	43.686	-9.2	94	0.00
90 C	Benzo(a)pyrene	40.000	41.834	-4.6	89	0.00
91	Dibenzo(a,h)anthracene	40.000	39.663	0.8	88	0.01
92	Benzo(g,h,i)perylene	40.000	39.656	0.9	89	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102720\
Data File : BF122134.D
Acq On : 27 Oct 2020 9:28
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 27 10:50:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 26 17:27:32 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0