

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102722\
 Data File : BF130807.D
 Acq On : 27 Oct 2022 19:45
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF102722

Quant Time: Oct 28 04:09:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 28 04:00:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	41.330	-3.3	95	0.00
3	Pyridine	40.000	40.690	-1.7	94	0.00
4	n-Nitrosodimethylamine	40.000	41.832	-4.6	97	0.00
5 S	2-Fluorophenol	80.000	78.660	1.7	95	0.00
6	Aniline	40.000	40.073	-0.2	94	0.00
7 S	Phenol-d6	80.000	78.210	2.2	95	0.00
8	2-Chlorophenol	40.000	40.355	-0.9	95	0.00
9	Benzaldehyde	40.000	38.042	4.9	100	0.00
10 C	Phenol	40.000	40.316	-0.8	95	0.00
11	bis(2-Chloroethyl)ether	40.000	40.282	-0.7	95	0.00
12	1,3-Dichlorobenzene	40.000	40.075	-0.2	94	0.00
13 C	1,4-Dichlorobenzene	40.000	39.784	0.5	94	0.00
14	1,2-Dichlorobenzene	40.000	40.039	-0.1	94	0.00
15	Benzyl Alcohol	40.000	41.082	-2.7	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.560	-1.4	95	0.00
17	2-Methylphenol	40.000	40.546	-1.4	95	0.00
18	Hexachloroethane	40.000	42.243	-5.6	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.267	-0.7	97	0.00
20	3+4-Methylphenols	40.000	40.583	-1.5	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	41.399	-3.5	97	0.00
23 S	Nitrobenzene-d5	80.000	95.371	-19.2	103	0.00
24	Nitrobenzene	40.000	44.873	-12.2	99	0.00
25	Isophorone	40.000	41.725	-4.3	96	0.00
26 C	2-Nitrophenol	40.000	44.382	-11.0	109	0.00
27	2,4-Dimethylphenol	40.000	42.516	-6.3	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.328	-3.3	96	0.00
29 C	2,4-Dichlorophenol	40.000	42.670	-6.7	97	0.00
30	1,2,4-Trichlorobenzene	40.000	41.343	-3.4	96	0.00
31	Naphthalene	40.000	41.336	-3.3	96	0.00
32	Benzoic acid	40.000	44.827	-12.1	114	0.00
33	4-Chloroaniline	40.000	41.393	-3.5	95	0.00
34 C	Hexachlorobutadiene	40.000	42.689	-6.7	99	0.00
35	Caprolactam	40.000	42.598	-6.5	96	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.628	-6.6	96	0.00
37	2-Methylnaphthalene	40.000	41.133	-2.8	96	0.00
38	1-Methylnaphthalene	40.000	41.684	-4.2	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.396	-1.0	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.692	-6.7	97	0.00
42 S	2,4,6-Tribromophenol	80.000	89.852	-12.3	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.999	-5.0	97	0.00
44	2,4,5-Trichlorophenol	40.000	41.829	-4.6	97	0.00
45 S	2-Fluorobiphenyl	80.000	77.862	2.7	99	0.00
46	1,1'-Biphenyl	40.000	40.002	-0.0	96	0.00
47	2-Chloronaphthalene	40.000	40.275	-0.7	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102722\
 Data File : BF130807.D
 Acq On : 27 Oct 2022 19:45
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF102722

Quant Time: Oct 28 04:09:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 28 04:00:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.359	-8.4	107	0.00
49	Acenaphthylene	40.000	40.918	-2.3	97	0.00
50	Dimethylphthalate	40.000	40.814	-2.0	97	0.00
51	2,6-Dinitrotoluene	40.000	43.702	-9.3	105	0.00
52 C	Acenaphthene	40.000	41.116	-2.8	98	0.00
53	3-Nitroaniline	40.000	48.905	-22.3	107	0.00
54 P	2,4-Dinitrophenol	40.000	45.063	-12.7	116	0.00
55	Dibenzofuran	40.000	40.447	-1.1	97	0.00
56 P	4-Nitrophenol	40.000	48.604	-21.5	104	0.00
57	2,4-Dinitrotoluene	40.000	44.356	-10.9	107	0.00
58	Fluorene	40.000	40.320	-0.8	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.836	-7.1	99	0.00
60	Diethylphthalate	40.000	42.496	-6.2	101	0.00
61	4-Chlorophenyl-phenylether	40.000	40.449	-1.1	98	0.00
62	4-Nitroaniline	40.000	46.851	-17.1	101	0.00
63	Azobenzene	40.000	41.682	-4.2	98	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.216	-8.0	115	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.498	-1.2	97	0.00
67	4-Bromophenyl-phenylether	40.000	41.239	-3.1	97	0.00
68	Hexachlorobenzene	40.000	40.629	-1.6	98	0.00
69	Atrazine	40.000	42.001	-5.0	97	0.00
70 C	Pentachlorophenol	40.000	44.723	-11.8	100	0.00
71	Phenanthrene	40.000	40.609	-1.5	98	0.00
72	Anthracene	40.000	40.181	-0.5	96	0.00
73	Carbazole	40.000	40.202	-0.5	97	0.00
74	Di-n-butylphthalate	40.000	41.211	-3.0	99	0.00
75 C	Fluoranthene	40.000	40.166	-0.4	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzidine	40.000	35.289	11.8	83	0.00
78	Pyrene	40.000	42.505	-6.3	98	0.00
79 S	Terphenyl-d14	80.000	83.940	-4.9	99	0.00
80	Butylbenzylphthalate	40.000	44.422	-11.1	99	0.00
81	Benzo(a)anthracene	40.000	41.720	-4.3	99	0.00
82	3,3'-Dichlorobenzidine	40.000	40.197	-0.5	95	0.00
83	Chrysene	40.000	39.778	0.6	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.572	-6.4	98	0.00
85 c	Di-n-octyl phthalate	40.000	42.977	-7.4	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.964	-14.9	101	0.00
88	Benzo(b)fluoranthene	40.000	40.737	-1.8	99	0.00
89	Benzo(k)fluoranthene	40.000	41.130	-2.8	97	0.00
90 C	Benzo(a)pyrene	40.000	41.519	-3.8	96	0.00
91	Dibenzo(a,h)anthracene	40.000	45.614	-14.0	99	0.00
92	Benzo(g,h,i)perylene	40.000	46.836	-17.1	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102722\
Data File : BF130807.D
Acq On : 27 Oct 2022 19:45
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102722

Quant Time: Oct 28 04:09:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 28 04:00:21 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0