

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102816\
 Data File : BF090907.D
 Acq On : 28 Oct 2016 13:54
 Operator : UM/SJ
 Sample : H5411-01 5X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-01-5.5-6.0

Manual Integrations
 APPROVED

sohil
 11/1/2016 7:13:40 PM

Quant Time: Nov 01 14:44:45 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 28 15:18:12 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.76	152	198918	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	777839	20.00	ng	0.00
38) Acenaphthene-d10	9.80	164	374794	20.00	ng	-0.01
63) Phenanthrene-d10	11.29	188	502947	20.00	ng	0.00
75) Chrysene-d12	13.93	240	448039	20.00	ng	0.00
86) Perylene-d12	15.33	264	465498	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	171789	15.10	ng	-0.01
7) Phenol-d6	6.40	99	245395	15.99	ng	-0.02
23) Nitrobenzene-d5	7.32	82	146139	12.30	ng	-0.01
41) 2,4,6-Tribromophenol	10.59	330	26497	6.88	ng	-0.01
44) 2-Fluorobiphenyl	9.13	172	206522	9.68	ng	-0.01
78) Terphenyl-d14	12.87	244	106032	5.96	ng	-0.01
Target Compounds						
49) Dimethylphthalate	9.53	163	66563	2.66	ng	Qvalue # 96
70) Phenanthrene	11.31	178	441513	17.24	ng	97
71) Anthracene	11.36	178	112396	4.17	ng	97
74) Fluoranthene	12.50	202	485870	17.32	ng	91
77) Pyrene	12.73	202	446059	15.73	ng	96
80) Benzo(a)anthracene	13.92	228	241590m	10.16	ng	
82) Chrysene	13.95	228	216255m	8.99	ng	
85) Indeno(1,2,3-cd)pyrene	16.68	276	97433m	4.57	ng	
87) Benzo(b)fluoranthene	14.93	252	264645m	8.45	ng	
88) Benzo(k)fluoranthene	14.95	252	111437m	4.50	ng	
89) Benzo(a)pyrene	15.28	252	219041	8.14	ng	# 86
91) Benzo(g,h,i)perylene	17.08	276	92575	4.32	ng	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF102816\
Data File : BF090907.D
Acq On : 28 Oct 2016 13:54
Operator : UM/SJ
Sample : H5411-01 5X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-01-5.5-6.0

Manual Integrations
APPROVED

sohil
11/1/2016 7:13:40 PM

Quant Time: Nov 01 14:44:45 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 28 15:18:12 2016
Response via : Initial Calibration

