

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102817\
 Data File : BF099923.D
 Acq On : 28 Oct 2017 17:58
 Operator : SJ/JU
 Sample : I6077-05 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FILL-PILE-2-COMP-2

Manual Integrations
 APPROVED

Sohil
 10/30/2017 6:20:29 PM

Quant Time: Oct 30 03:48:04 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 27 15:58:57 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	102661	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	406601	20.00	ng	-0.01
38) Acenaphthene-d10	9.82	164	170670	20.00	ng	-0.02
63) Phenanthrene-d10	11.32	188	243819	20.00	ng	-0.02
75) Chrysene-d12	13.96	240	173892	20.00	ng	-0.03
86) Perylene-d12	15.43	264	197568	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	118099	18.06	ng	0.00
7) Phenol-d6	6.43	99	154896	19.98	ng	0.00
23) Nitrobenzene-d5	7.36	82	84982	13.46	ng	-0.01
41) 2,4,6-Tribromophenol	10.62	330	22376	14.39	ng	-0.02
44) 2-Fluorobiphenyl	9.14	172	191584	17.32	ng	-0.02
78) Terphenyl-d14	12.89	244	105585	13.27	ng	-0.03
Target Compounds						
49) Dimethylphthalate	9.54	163	28447	2.105	ng	Qvalue # 99
70) Phenanthrene	11.34	178	201650	14.655	ng	98
71) Anthracene	11.39	178	43108	3.086	ng	96
74) Fluoranthene	12.53	202	254196	17.776	ng	96
77) Pyrene	12.76	202	215643	16.309	ng	98
80) Benzo(a)anthracene	13.95	228	108640	9.855	ng	97
82) Chrysene	13.99	228	99671	9.617	ng	99
85) Indeno(1,2,3-cd)pyrene	16.90	276	56520	5.941	ng	97
87) Benzo(b)fluoranthene	15.00	252	139919m	11.216	ng	
88) Benzo(k)fluoranthene	15.03	252	38727m	3.292	ng	
89) Benzo(a)pyrene	15.37	252	96988	8.655	ng	95
91) Benzo(a,h,i)perylene	17.34	276	51662	5.303	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF102817\
Data File : BF099923.D
Acq On : 28 Oct 2017 17:58
Operator : SJ/JU
Sample : I6077-05 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FILL-PILE-2-COMP-2

Manual Integrations
APPROVED

Sohil
10/30/2017 6:20:29 PM

Quant Time: Oct 30 03:48:04 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 27 15:58:57 2017
Response via : Initial Calibration

