

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102819\
 Data File : BF117575.D
 Acq On : 28 Oct 2019 21:24
 Operator : JU
 Sample : K5153-05
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TR-05-102419

Manual Integrations
 APPROVED

mohammad
 10/30/2019 4:51:41 PM

Quant Time: Oct 29 15:08:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 18:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	44856	20.00	ng	-0.01
21) Naphthalene-d8	8.01	136	183597	20.00	ng	0.00
39) Acenaphthene-d10	9.76	164	90751	20.00	ng	-0.01
64) Phenanthrene-d10	11.24	188	129230	20.00	ng	0.00
76) Chrysene-d12	13.89	240	88327	20.00	ng	0.00
87) Perylene-d12	15.32	264	76475	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	157470	52.23	ng	0.00
7) Phenol-d6	6.38	99	223989	55.31	ng	0.00
23) Nitrobenzene-d5	7.29	82	134383	36.13	ng	-0.01
42) 2,4,6-Tribromophenol	10.56	330	35698	45.46	ng	0.00
45) 2-Fluorobiphenyl	9.08	172	218065	33.86	ng	-0.01
79) Terphenyl-d14	12.83	244	162038	29.92	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.48	163	33810	5.226	ng	99
71) Phenanthrene	11.27	178	19517	2.601	ng	99
75) Fluoranthene	12.46	202	25492	3.394	ng	96
78) Pyrene	12.69	202	25507	3.362	ng	99
81) Benzo(a)anthracene	13.88	228	12489	2.096	ng	96
88) Benzo(b)fluoranthene	14.92	252	12235m	2.709	ng	
90) Benzo(a)pyrene	15.27	252	9606	2.340	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF102819\
Data File : BF117575.D
Acq On : 28 Oct 2019 21:24
Operator : JU
Sample : K5153-05
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TR-05-102419

Manual Integrations
APPROVED

mohammad
10/30/2019 4:51:41 PM

Quant Time: Oct 29 15:08:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 18:53:31 2019
Response via : Initial Calibration

