


Data Path : Z:\HPCHEM1\BNA F\DATA\BF102915\  
 Data File : BF082574.D  
 Acq On : 29 Oct 2015 11:25  
 Operator : UM/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF102615.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Oct 29 19:06:23 2015



AutoFind: Scans 863, 864, 865; Background Corrected with Scan 860

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 37.6      | 17928   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.5       | 249     | PASS             |
| 69          | 198          | 0.00         | 100          | 36.0      | 17139   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 46.5      | 22140   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 195     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 47646   | PASS             |
| 199         | 198          | 5            | 9            | 6.5       | 3106    | PASS             |
| 275         | 198          | 10           | 60           | 22.9      | 10913   | PASS             |
| 365         | 198          | 1            | 100          | 2.5       | 1190    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.8      | 5812    | PASS             |
| 442         | 198          | 50           | 100          | 77.2      | 36760   | PASS             |
| 443         | 442          | 15           | 24           | 19.1      | 7016    | PASS             |