




Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF102918\
 Data File : BF110292.D
 Acq On : 30 Oct 2018 6:44
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 01 08:51:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 31 14:18:16 2018
 Response via : Initial Calibration



(77) Benzidine




12.845min (-0.006) 0.00ng

response 194152

Ion	Exp%	Act%
184.00	100	100
185.00	16.80	14.32
183.00	11.80	11.57
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF102918\
Data File : BF110292.D
Acq On : 30 Oct 2018 6:44
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 01 08:51:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 31 14:18:16 2018
Response via : Initial Calibration



(70) Pentachlorophenol (C)

11.310min (-0.012) 0.00ng

response 13945

Ion	Exp%	Act%
265.70	100	100
268.00	63.20	66.65
264.00	63.20	63.66
0.00	0.00	0.00