


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102918\
 Data File : BF110292.D
 Acq On : 30 Oct 2018 6:44
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102318.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 29 17:29:00 2018



AutoFind: Scans 1640, 1641, 1642; Background Corrected with Scan 1636

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.3	16238	PASS
68	69	0.00	2	1.0	156	PASS
69	198	0.00	100	37.1	14978	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.8	20069	PASS
197	198	0.00	2	0.5	215	PASS
198	198	100	100	100.0	40336	PASS
199	198	5	9	7.2	2919	PASS
275	198	10	60	25.5	10303	PASS
365	198	1	100	2.8	1111	PASS
441	198	0.01	100	13.5	5432	PASS
442	198	50	100	90.9	36674	PASS
443	442	15	24	17.6	6440	PASS