

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102920\
 Data File : BF122160.D
 Acq On : 29 Oct 2020 10:25
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 29 11:23:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 26 17:27:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	36.402	9.0	94	0.00
3	Pyridine	40.000	37.360	6.6	95	0.00
4	n-Nitrosodimethylamine	40.000	38.956	2.6	99	0.01
5 S	2-Fluorophenol	80.000	79.455	0.7	104	0.00
6	Aniline	40.000	37.084	7.3	98	0.00
7 S	Phenol-d6	80.000	74.034	7.5	98	0.00
8	2-Chlorophenol	40.000	39.442	1.4	103	0.00
9	Benzaldehyde	40.000	39.433	1.4	102	0.00
10 C	Phenol	40.000	37.545	6.1	100	0.01
11	bis(2-Chloroethyl)ether	40.000	39.053	2.4	101	0.00
12	1,3-Dichlorobenzene	40.000	38.541	3.6	101	0.00
13 C	1,4-Dichlorobenzene	40.000	39.058	2.4	102	0.00
14	1,2-Dichlorobenzene	40.000	37.794	5.5	99	0.00
15	Benzyl Alcohol	40.000	41.150	-2.9	107	0.01
16	2,2'-oxybis(1-Chloropropane	40.000	37.783	5.5	98	0.00
17	2-Methylphenol	40.000	37.454	6.4	99	0.00
18	Hexachloroethane	40.000	39.997	0.0	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.274	-0.7	108	0.00
20	3+4-Methylphenols	40.000	40.530	-1.3	109	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	39.723	0.7	107	0.00
23 S	Nitrobenzene-d5	80.000	79.892	0.1	105	0.00
24	Nitrobenzene	40.000	39.540	1.2	104	0.00
25	Isophorone	40.000	40.362	-0.9	107	0.00
26 C	2-Nitrophenol	40.000	40.947	-2.4	104	0.00
27	2,4-Dimethylphenol	40.000	38.582	3.5	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.080	2.3	103	0.00
29 C	2,4-Dichlorophenol	40.000	39.803	0.5	104	0.00
30	1,2,4-Trichlorobenzene	40.000	38.968	2.6	103	0.00
31	Naphthalene	40.000	38.405	4.0	102	0.00
32	Benzoic acid	40.000	34.133	14.7	90	0.02
33	4-Chloroaniline	40.000	39.880	0.3	105	0.00
34 C	Hexachlorobutadiene	40.000	41.158	-2.9	106	0.00
35	Caprolactam	40.000	41.804	-4.5	110	0.01
36 C	4-Chloro-3-methylphenol	40.000	41.003	-2.5	108	0.00
37	2-Methylnaphthalene	40.000	38.598	3.5	103	0.00
38	1-Methylnaphthalene	40.000	38.709	3.2	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.265	1.8	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.809	8.0	93	0.00
42 S	2,4,6-Tribromophenol	80.000	83.193	-4.0	111	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.084	2.3	104	0.00
44	2,4,5-Trichlorophenol	40.000	36.997	7.5	99	0.01
45 S	2-Fluorobiphenyl	80.000	74.290	7.1	104	0.00
46	1,1'-Biphenyl	40.000	38.367	4.1	105	0.00
47	2-Chloronaphthalene	40.000	38.751	3.1	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102920\
 Data File : BF122160.D
 Acq On : 29 Oct 2020 10:25
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 29 11:23:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 26 17:27:32 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.296	-3.2	110	0.00
49	Acenaphthylene	40.000	37.362	6.6	102	0.00
50	Dimethylphthalate	40.000	38.955	2.6	105	0.00
51	2,6-Dinitrotoluene	40.000	40.249	-0.6	106	0.00
52 C	Acenaphthene	40.000	38.541	3.6	106	0.00
53	3-Nitroaniline	40.000	39.639	0.9	107	0.00
54 P	2,4-Dinitrophenol	40.000	35.570	11.1	98	0.01
55	Dibenzofuran	40.000	38.142	4.6	105	0.00
56 P	4-Nitrophenol	40.000	32.940	17.7	86	0.01
57	2,4-Dinitrotoluene	40.000	41.755	-4.4	112	0.00
58	Fluorene	40.000	38.924	2.7	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.829	-2.1	106	0.00
60	Diethylphthalate	40.000	40.918	-2.3	111	0.00
61	4-Chlorophenyl-phenylether	40.000	40.209	-0.5	111	0.00
62	4-Nitroaniline	40.000	36.781	8.0	99	0.01
63	Azobenzene	40.000	40.115	-0.3	108	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.816	5.5	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.682	3.3	107	0.00
67	4-Bromophenyl-phenylether	40.000	40.212	-0.5	109	0.00
68	Hexachlorobenzene	40.000	40.023	-0.1	110	0.00
69	Atrazine	40.000	41.596	-4.0	112	0.00
70 C	Pentachlorophenol	40.000	40.368	-0.9	119	0.00
71	Phenanthrene	40.000	38.599	3.5	107	0.00
72	Anthracene	40.000	37.999	5.0	106	0.00
73	Carbazole	40.000	38.120	4.7	105	0.00
74	Di-n-butylphthalate	40.000	40.107	-0.3	109	0.00
75 C	Fluoranthene	40.000	38.125	4.7	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzydine	40.000	36.040	9.9	93	0.00
78	Pyrene	40.000	41.058	-2.6	104	0.00
79 S	Terphenyl-d14	80.000	82.096	-2.6	106	0.00
80	Butylbenzylphthalate	40.000	43.085	-7.7	106	0.00
81	Benzo(a)anthracene	40.000	39.417	1.5	100	0.00
82	3,3'-Dichlorobenzidine	40.000	38.476	3.8	95	0.00
83	Chrysene	40.000	39.262	1.8	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.163	-5.4	104	0.00
85 c	Di-n-octyl phthalate	40.000	40.890	-2.2	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	38.427	3.9	88	0.02
88	Benzo(b)fluoranthene	40.000	44.220	-10.5	96	0.01
89	Benzo(k)fluoranthene	40.000	41.171	-2.9	91	0.00
90 C	Benzo(a)pyrene	40.000	42.227	-5.6	93	0.00
91	Dibenzo(a,h)anthracene	40.000	38.632	3.4	88	0.01
92	Benzo(g,h,i)perylene	40.000	38.329	4.2	88	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102920\
Data File : BF122160.D
Acq On : 29 Oct 2020 10:25
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 29 11:23:30 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 26 17:27:32 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0