



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102920\
 Data File : BF122159.D
 Acq On : 29 Oct 2020 9:54
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 30 10:16:54 2020



AutoFind: Scans 1597, 1598, 1599; Background Corrected with Scan 1592

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.8	43861	PASS
68	69	0.00	2	1.8	723	PASS
69	198	0.00	100	38.2	39211	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.2	49408	PASS
197	198	0.00	2	0.4	402	PASS
198	198	100	100	100.0	102573	PASS
199	198	5	9	6.7	6846	PASS
275	198	10	60	28.2	28888	PASS
365	198	1	100	4.0	4152	PASS
441	198	0.01	100	14.5	14826	PASS
442	198	50	100	97.3	99779	PASS
443	442	15	24	18.1	18089	PASS