

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF102924\  
 Data File : BF140143.D  
 Acq On : 30 Oct 2024 05:58  
 Operator : RC/JU  
 Sample : P4594-12  
 Misc :  
 ALS Vial : 32 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 BP-F16

Quant Time: Oct 30 06:25:24 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF101824.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Oct 18 15:07:50 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.887  | 152  | 126489   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.169  | 136  | 505147   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.922  | 164  | 267797   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.410 | 188  | 404359   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 14.051 | 240  | 249311   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 15.539 | 264  | 183532   | 20.000  | ng    | 0.01     |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.516  | 112  | 837901   | 103.703 | ng    | 0.02     |
| 7) Phenol-d6                | 6.510  | 99   | 1004349  | 95.975  | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.445  | 82   | 800131   | 87.789  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.710 | 330  | 263660   | 105.258 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.245  | 172  | 1401345  | 86.483  | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.998 | 244  | 1233644  | 80.630  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF102924\  
Data File : BF140143.D  
Acq On : 30 Oct 2024 05:58  
Operator : RC/JU  
Sample : P4594-12  
Misc :  
ALS Vial : 32 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
BP-F16

Quant Time: Oct 30 06:25:24 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF101824.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Oct 18 15:07:50 2024  
Response via : Initial Calibration

