

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF103017\
 Data File : BF099953.D
 Acq On : 30 Oct 2017 12:34
 Operator : SJ/JU
 Sample : PB103742BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB103742BL

Integration Parameters: rteint.p
 Integrator: RTE
 Smoothing : ON
 Sampling : 1
 Start Thrs: 0.2
 Stop Thrs : 0
 Filtering: 5
 Min Area: 3 % of largest Peak
 Max Peaks: 100
 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
 Peak separation: 5

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.016	495	498	517	rBV	158045	257732	11.33%	1.453%
2	5.416	563	566	592	rBV	1548123	1689382	74.25%	9.522%
3	6.434	735	739	757	rBV	1574246	1631253	71.69%	9.194%
4	6.557	757	760	771	rVB	1765957	1762151	77.44%	9.932%
5	6.787	796	799	810	rVB	547838	488568	21.47%	2.754%
6	6.940	821	825	835	rBV	1577031	1428311	62.77%	8.050%
7	7.357	892	896	905	rBV	1107973	1050666	46.18%	5.922%
8	8.069	1013	1017	1031	rBV	783182	691585	30.39%	3.898%
9	9.145	1194	1200	1203	rBV	2390895	2264553	99.52%	12.763%
10	9.822	1311	1315	1326	rVB	800432	820778	36.07%	4.626%
11	10.622	1447	1451	1465	rBV	1268343	1250401	54.95%	7.048%
12	11.322	1566	1570	1578	rBV	1000265	861397	37.86%	4.855%
13	12.910	1835	1840	1843	rBV	2333548	2275371	100.00%	12.824%
14	13.986	2019	2023	2033	rBV	750028	723582	31.80%	4.078%
15	15.492	2275	2279	2292	rVB	372006	498726	21.92%	2.811%
16	15.886	2342	2346	2352	rVB3	17300	25215	1.11%	0.142%
17	16.962	2522	2529	2538	rVB9	9414	22779	1.00%	0.128%


Sum of corrected areas: 17742450

Data Path : Z:\HPCHEM1\BNA F\DATA\BF103017\
Data File : BF099953.D
Acq On : 30 Oct 2017 12:34
Operator : SJ/JU
Sample : PB103742BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB103742BL

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\HPCHEM1\BNA F\DATA\BF103017\
Data File : BF099953.D
Acq On : 30 Oct 2017 12:34
Operator : SJ/JU
Sample : PB103742BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB103742BL










Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 3

R.T.	EstConc	Area	Relative to ISTD	R.T.
5.02	10.55 ng	257732	1,4-Dichlorobenzene-d4	6.79

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	56
2		Acetic acid, cyano-, 1,1-dimethyl...	141	C7H11NO2	001116-98-9	25
3		(+)-4-Amino-4,5-dihydro-2(3H)-f...	101	C4H7NO2	016504-58-8	9
4		Acetone	58	C3H6O	000067-64-1	9
5		Morpholine, 4-methyl-	101	C5H11NO	000109-02-4	9



Data Path : Z:\HPCHEM1\BNA F\DATA\BF103017\
Data File : BF099953.D
Acq On : 30 Oct 2017 12:34
Operator : SJ/JU
Sample : PB103742BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB103742BL










Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 unknown6.56 Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
6.56	72.14 ng	1762150	1,4-Dichlorobenzene-d4	6.79

Hit#	of	Tentative ID	MW	MolForm	CAS#	Qual
1	5	3-Chloro-6-fluoro-pyrazine	132	C4H2ClFN2	1000146-10-7	38
2		(E)-3-Chloro-2-methyl-2-pentenal	132	C6H9ClO	031357-76-3	35
3		Tranvlcypropine-propionyl	189	C12H15NO	1000123-86-3	12
4		2H-Cyclopenta[d]pyridazine, 2-me...	132	C8H8N2	022291-85-6	9
5		4-Chloro-2-fluoro-pyrimidine	132	C4H2ClFN2	051422-00-5	9



Data Path : Z:\HPCHEM1\BNA_F\DATA\BF103017\
Data File : BF099953.D
Acq On : 30 Oct 2017 12:34
Operator : SJ/JU
Sample : PB103742BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB103742BL

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
2-Pentanone, 4-hy...	5.02	10.6	ng	257732	1	6.79	488568	20.0
unknown6.56	6.56	72.1	ng	1762150	1	6.79	488568	20.0