

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF103019\
 Data File : BF117644.D
 Acq On : 31 Oct 2019 4:27
 Operator : JU
 Sample : K5583-06
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OB-722

Quant Time: Oct 31 08:15:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 18:53:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	37378	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	151320	20.00	ng	0.00
39) Acenaphthene-d10	9.76	164	72245	20.00	ng	0.00
64) Phenanthrene-d10	11.25	188	104275	20.00	ng	0.00
76) Chrysene-d12	13.90	240	70272	20.00	ng	0.00
87) Perylene-d12	15.33	264	64135	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	128472	51.13	ng	0.00
7) Phenol-d6	6.39	99	122677	36.35	ng	0.01
23) Nitrobenzene-d5	7.29	82	219564	71.63	ng	-0.01
42) 2,4,6-Tribromophenol	10.56	330	39524	63.23	ng	0.00
45) 2-Fluorobiphenyl	9.08	172	364265	71.04	ng	-0.01
79) Terphenyl-d14	12.83	244	277297	64.36	ng	0.00
Target Compounds						
10) Phenol	6.40	94	12310	3.574	ng	91
50) Dimethylphthalate	9.48	163	39378	7.646	ng	98
74) Di-n-butylphthalate	11.80	149	36097	4.758	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF103019\
Data File : BF117644.D
Acq On : 31 Oct 2019 4:27
Operator : JU
Sample : K5583-06
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OB-722

Quant Time: Oct 31 08:15:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF101819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 18:53:31 2019
Response via : Initial Calibration

