

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF103020\  
 Data File : BF122211.D  
 Acq On : 31 Oct 2020 9:39  
 Operator : JU/CG  
 Sample : L4549-02  
 Misc :  
 ALS Vial : 32 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 CHASSIS-WASH

Quant Time: Nov 01 23:51:18 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF101220.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Oct 30 10:16:54 2020  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards          |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 6.710  | 152  | 91638    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.992  | 136  | 328702   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.745  | 164  | 140851   | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.233 | 188  | 223969   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 13.880 | 240  | 208829   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 15.315 | 264  | 198472   | 20.00  | ng    | 0.01     |
| System Monitoring Compounds |        |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.345  | 112  | 726138   | 116.95 | ng    | 0.01     |
| 7) Phenol-d6                | 6.369  | 99   | 807590   | 101.04 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.281  | 82   | 612949   | 95.63  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.539 | 330  | 206966   | 118.79 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.069  | 172  | 948514   | 99.08  | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.816 | 244  | 801534   | 73.15  | ng    | 0.00     |
| Target Compounds            |        |      |          |        |       | Qvalue   |
| 50) Dimethylphthalate       | 9.469  | 163  | 46326    | 4.50   | ng    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF103020\  
Data File : BF122211.D  
Acq On : 31 Oct 2020 9:39  
Operator : JU/CG  
Sample : L4549-02  
Misc :  
ALS Vial : 32 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
CHASSIS-WASH

Quant Time: Nov 01 23:51:18 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF101220.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Oct 30 10:16:54 2020  
Response via : Initial Calibration

