

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF103020\
 Data File : BF122206.D
 Acq On : 31 Oct 2020 6:36
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

mohammad
 11/2/2020 12:36:46 PM

Quant Time: Oct 31 08:19:54 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 10:16:54 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.710	152	103595	20.00	ng	0.00
21) Naphthalene-d8	7.992	136	373869	20.00	ng	0.00
39) Acenaphthene-d10	9.745	164	176632	20.00	ng	0.00
64) Phenanthrene-d10	11.233	188	277421	20.00	ng	0.00
76) Chrysene-d12	13.880	240	212466	20.00	ng	0.00
86) Perylene-d12	15.321	264	188084	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.334	112	562092	80.08	ng	0.00
7) Phenol-d6	6.375	99	679603	75.21	ng	0.00
23) Nitrobenzene-d5	7.287	82	599015	82.17	ng	0.00
42) 2,4,6-Tribromophenol	10.545	330	157524	72.09	ng	0.00
45) 2-Fluorobiphenyl	9.069	172	982403	81.83	ng	0.00
79) Terphenyl-d14	12.816	244	840038	75.35	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.346	88	113980	36.92	ng	99
3) Pyridine	3.069	79	298075	36.72	ng	97
4) n-Nitrosodimethylamine	3.040	42	156415	39.87	ng	93
6) Aniline	6.381	93	415115	37.31	ng	# 68
8) 2-Chlorophenol	6.504	128	282410	39.81	ng	100
9) Benzaldehyde	6.269	77	202363	41.09	ng	99
10) Phenol	6.387	94	352425	37.80	ng	82
11) bis(2-Chloroethyl)ether	6.451	93	278753	38.67	ng	99
12) 1,3-Dichlorobenzene	6.651	146	308332	38.64	ng	97
13) 1,4-Dichlorobenzene	6.728	146	313238	39.10	ng	99
14) 1,2-Dichlorobenzene	6.881	146	283463	38.70	ng	98
15) Benzyl Alcohol	6.875	79	247191	42.30	ng	98
16) 2,2'-oxybis(1-Chloropr...	6.987	45	427404	38.53	ng	# 40
17) 2-Methylphenol	6.992	107	229756	37.88	ng	# 85
18) Hexachloroethane	7.216	117	103916	35.91	ng	93
19) n-Nitroso-di-n-propyla...	7.140	70	213793	41.56	ng	96
20) 3+4-Methylphenols	7.145	107	313484	42.86	ng	# 65
22) Acetophenone	7.134	105	402303	41.82	ng	# 88
24) Nitrobenzene	7.304	77	317967	40.81	ng	99
25) Isophorone	7.545	82	562926	40.84	ng	99
26) 2-Nitrophenol	7.622	139	137946	38.44	ng	94
27) 2,4-Dimethylphenol	7.669	122	240407	39.30	ng	96
28) bis(2-Chloroethoxy)met...	7.751	93	350074	40.59	ng	100
29) 2,4-Dichlorophenol	7.875	162	229994	40.15	ng	98
30) 1,2,4-Trichlorobenzene	7.939	180	240122	39.42	ng	99
31) Naphthalene	8.016	128	790745	39.48	ng	99
32) Benzoic acid	7.834	122	94488	30.38	ng	97
33) 4-Chloroaniline	8.081	127	337584	39.57	ng	99
34) Hexachlorobutadiene	8.128	225	151793	42.03	ng	99
35) Caprolactam	8.463	113	71016	39.04	ng	97
36) 4-Chloro-3-methylphenol	8.575	107	247737	40.27	ng	97
37) 2-Methylnaphthalene	8.710	142	496784	38.30	ng	97
38) 1-Methylnaphthalene	8.804	142	455261	37.90	ng	100
40) 1,2,4,5-Tetrachloroben...	8.875	216	237835	41.72	ng	99
41) Hexachlorocyclopentadiene	8.851	237	7346	15.24	ng	98
43) 2,4,6-Trichlorophenol	8.998	196	149321	42.45	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF103020\
 Data File : BF122206.D
 Acq On : 31 Oct 2020 6:36
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

mohammad
 11/2/2020 12:36:46 PM

Quant Time: Oct 31 08:19:54 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 10:16:54 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.057	196	144807	38.12	ng	98
46) 1,1'-Biphenyl	9.169	154	597214	41.58	ng	99
47) 2-Chloronaphthalene	9.198	162	451623	40.44	ng	99
48) 2-Nitroaniline	9.310	65	157637	42.79	ng	97
49) Acenaphthylene	9.610	152	656078	38.25	ng	100
50) Dimethylphthalate	9.475	163	539784	41.79	ng	100
51) 2,6-Dinitrotoluene	9.551	165	112298	39.64	ng	96
52) Acenaphthene	9.781	154	402625	39.46	ng	99
53) 3-Nitroaniline	9.728	138	128322	39.82	ng	97
54) 2,4-Dinitrophenol	9.863	184	22609m	22.63	ng	
55) Dibenzofuran	9.951	168	586590	39.31	ng	95
56) 4-Nitrophenol	9.922	139	78512	44.79	ng	82
57) 2,4-Dinitrotoluene	9.963	165	139415	39.97	ng	98
58) Fluorene	10.298	166	460316	38.29	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.086	232	113408	38.59	ng	99
60) Diethylphthalate	10.169	149	525236	42.17	ng	98
61) 4-Chlorophenyl-phenyle...	10.286	204	232398	40.31	ng	97
62) 4-Nitroaniline	10.345	138	115512	35.88	ng	91
63) Azobenzene	10.445	77	510068	38.73	ng	98
65) 4,6-Dinitro-2-methylph...	10.380	198	27626	17.92	ng	96
66) n-Nitrosodiphenylamine	10.410	169	420462	44.00	ng	99
67) 4-Bromophenyl-phenylether	10.775	248	141224	44.06	ng	97
68) Hexachlorobenzene	10.845	284	147642	40.70	ng	97
69) Atrazine	10.939	200	87936	37.63	ng	98
70) Pentachlorophenol	11.057	266	53252	41.57	ng	98
71) Phenanthrene	11.263	178	596690	39.23	ng	99
72) Anthracene	11.310	178	623750	39.54	ng	99
73) Carbazole	11.475	167	549192	39.15	ng	98
74) Di-n-butylphthalate	11.792	149	706901	43.80	ng	99
75) Fluoranthene	12.451	202	589074	36.12	ng	98
77) Benzidine	12.580	184	236482	36.10	ng	100
78) Pyrene	12.680	202	598185	36.94	ng	99
80) Butylbenzylphthalate	13.286	149	255574	39.80	ng	97
81) Benzo(a)anthracene	13.874	228	561137	40.30	ng	99
82) 3,3'-Dichlorobenzidine	13.833	252	213659	41.37	ng	98
83) Chrysene	13.910	228	543789	39.79	ng	99
84) Bis(2-ethylhexyl)phtha...	13.845	149	352264	40.41	ng	100
85) Di-n-octyl phthalate	14.457	149	608634	43.16	ng	99
87) Indeno(1,2,3-cd)pyrene	16.733	276	462084	37.84	ng	98
88) Benzo(b)fluoranthene	14.910	252	530302	41.71	ng	99
89) Benzo(k)fluoranthene	14.933	252	508820	42.13	ng	99
90) Benzo(a)pyrene	15.263	252	451697	41.13	ng	100
91) Dibenzo(a,h)anthracene	16.733	278	393329	39.12	ng	97
92) Benzo(g,h,i)perylene	17.157	276	372021	36.97	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF103020\
Data File : BF122206.D
Acq On : 31 Oct 2020 6:36
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM

Quant Time: Oct 31 08:19:54 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 30 10:16:54 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.710 min Scan# 786
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion:152 Resp: 103595
Ion Ratio Lower Upper
152 100
150 158.5 125.3 187.9
115 62.7 51.9 77.9



#2
1,4-Dioxane
Concen: 36.92 ng
RT: 2.346 min Scan# 44
Delta R.T. -0.018 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion: 88 Resp: 113980
Ion Ratio Lower Upper
88 100
58 81.6 65.8 98.8
43 33.8 25.8 38.8




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED




mohammad
11/2/2020 12:36:46 PM




#3
Pyridine
Concen: 36.72 ng
RT: 3.069 min Scan# 167
Delta R.T. -0.012 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion: 79 Resp: 298075
Ion Ratio Lower Upper
79 100
52 80.3 62.2 93.4
51 40.4 31.1 46.7



#4
n-Nitrosodimethylamine
Concen: 39.87 ng
RT: 3.040 min Scan# 162
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion: 42 Resp: 156415
Ion Ratio Lower Upper
42 100
74 116.7 100.2 150.2
44 6.3 5.8 8.6



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM







#5
2-Fluorophenol
Concen: 80.08 ng
RT: 5.334 min Scan# 552
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 112 Resp: 562092
Ion Ratio Lower Upper
112 100
64 63.2 50.4 75.6
63 32.8 25.8 38.6

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#6
Aniline
Concen: 37.31 ng
RT: 6.381 min Scan# 730
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 93 Resp: 415115
Ion Ratio Lower Upper
93 100
66 62.5 32.7 49.1#
65 33.1 16.6 24.8#









#7
Phenol-d6
Concen: 75.21 ng
RT: 6.375 min Scan# 729
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	99	Resp:	679603
Ion	Ratio	Lower	Upper
99	100		
42	21.7	15.9	23.9
71	36.6	27.1	40.7

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#8
2-Chlorophenol
Concen: 39.81 ng
RT: 6.504 min Scan# 751
Delta R.T. 0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	128	Resp:	282410
Ion	Ratio	Lower	Upper
128	100		
130	32.1	12.4	52.4
64	54.9	34.8	74.8









#9
Benzaldehyde
Concen: 41.09 ng
RT: 6.269 min Scan# 711
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 202363
Ion Ratio Lower Upper
77 100
105 97.1 76.7 116.7
106 93.4 71.3 111.3




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#10
Phenol
Concen: 37.80 ng
RT: 6.387 min Scan# 731
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 94 Resp: 352425
Ion Ratio Lower Upper
94 100
65 44.7 17.1 57.1
66 72.1 36.0 76.0









#11
bis(2-Chloroethyl)ether
Concen: 38.67 ng
RT: 6.451 min Scan# 742
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 93 Resp: 278753
Ion Ratio Lower Upper
93 100
63 82.9 62.4 102.4
95 32.3 12.6 52.6

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#12
1,3-Dichlorobenzene
Concen: 38.64 ng
RT: 6.651 min Scan# 776
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 146 Resp: 308332
Ion Ratio Lower Upper
146 100
148 65.2 51.0 76.4
75 31.9 27.6 41.4









#13
1,4-Dichlorobenzene
Concen: 39.10 ng
RT: 6.728 min Scan# 789
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:146 Resp: 313238
Ion Ratio Lower Upper
146 100
148 63.6 50.1 75.1
111 43.1 34.6 52.0

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#14
1,2-Dichlorobenzene
Concen: 38.70 ng
RT: 6.881 min Scan# 815
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:146 Resp: 283463
Ion Ratio Lower Upper
146 100
148 63.2 51.1 76.7
111 45.4 34.2 51.4









#15
Benzyl Alcohol
Concen: 42.30 ng
RT: 6.875 min Scan# 814
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 79 Resp: 247191
Ion Ratio Lower Upper
79 100
108 74.2 62.3 93.5
77 63.2 50.6 75.8

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.53 ng
RT: 6.987 min Scan# 833
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 45 Resp: 427404
Ion Ratio Lower Upper
45 100
77 37.0 0.0 33.3#
79 32.8 0.0 30.4#









#17
2-Methylphenol
Concen: 37.88 ng
RT: 6.992 min Scan# 834
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	107	Resp:	229756
Ion	Ratio	Lower	Upper
107	100		
108	111.8	90.9	136.3
77	70.6	37.5	56.3#
79	65.4	37.4	56.0#

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM




#18
Hexachloroethane
Concen: 35.91 ng
RT: 7.216 min Scan# 872
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	117	Resp:	103916
Ion	Ratio	Lower	Upper
117	100		
119	94.9	76.3	114.5
201	102.9	71.8	107.6






#19
n-Nitroso-di-n-propylamine
Concen: 41.56 ng
RT: 7.140 min Scan# 859
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion: 70 Resp: 213793
Ion Ratio Lower Upper
70 100
42 59.0 50.2 75.4
101 9.7 7.5 11.3
130 20.7 15.5 23.3




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED



mohammad
11/2/2020 12:36:46 PM




#20
3+4-Methylphenols
Concen: 42.86 ng
RT: 7.145 min Scan# 860
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion: 107 Resp: 313484
Ion Ratio Lower Upper
107 100
108 85.0 68.9 108.9
77 54.3 108.0 148.0#
79 27.9 10.5 50.5








#21
Naphthalene-d8
Concen: 20.00 ng
RT: 7.992 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040





Tgt Ion:136 Resp: 373869
Ion Ratio Lower Upper
136 100
137 11.0 8.6 13.0
54 9.9 7.1 10.7
68 6.6 5.0 7.4




#22
Acetophenone
Concen: 41.82 ng
RT: 7.134 min Scan# 858
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:105 Resp: 402303
Ion Ratio Lower Upper
105 100
71 8.8 5.3 7.9#
51 29.9 33.2 49.8#
120 21.7 17.4 26.0









#23
Nitrobenzene-d5
Concen: 82.17 ng
RT: 7.287 min Scan# 884
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion: 82 Resp: 599015
Ion Ratio Lower Upper
82 100
128 41.5 35.0 52.4
54 56.5 43.9 65.9



#24
Nitrobenzene
Concen: 40.81 ng
RT: 7.304 min Scan# 887
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion: 77 Resp: 317967
Ion Ratio Lower Upper
77 100
123 43.3 35.6 53.4
65 13.9 11.2 16.8




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED




mohammad
11/2/2020 12:36:46 PM




#25
Isophorone
Concen: 40.84 ng
RT: 7.545 min Scan# 928
Delta R.T. 0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion: 82 Resp: 562926
Ion Ratio Lower Upper
82 100
95 7.1 5.4 8.0
138 16.9 13.4 20.0



#26
2-Nitrophenol
Concen: 38.44 ng
RT: 7.622 min Scan# 941
Delta R.T. 0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion: 139 Resp: 137946
Ion Ratio Lower Upper
139 100
109 32.4 23.0 34.4
65 53.7 39.9 59.9




Instrument :
BNA_F
ClientSampled :
SSTDCCC040

Manual Integrations
APPROVED




mohammad
11/2/2020 12:36:46 PM




#27
2,4-Dimethylphenol
Concen: 39.30 ng
RT: 7.669 min Scan# 949
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion:122 Resp: 240407
Ion Ratio Lower Upper
122 100
107 119.1 90.8 136.2
121 59.7 47.5 71.3



#28
bis(2-Chloroethoxy)methane
Concen: 40.59 ng
RT: 7.751 min Scan# 963
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion: 93 Resp: 350074
Ion Ratio Lower Upper
93 100
95 32.6 26.2 39.4
123 10.6 8.8 13.2



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM



#29
2,4-Dichlorophenol
Concen: 40.15 ng
RT: 7.875 min Scan# 984
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:162 Resp: 229994
Ion Ratio Lower Upper
162 100
164 65.3 42.5 82.5
98 39.7 20.1 60.1

Manual Integrations
APPROVED



mohammad
11/2/2020 12:36:46 PM




#30
1,2,4-Trichlorobenzene
Concen: 39.42 ng
RT: 7.939 min Scan# 995
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36



Tgt Ion:180 Resp: 240122
Ion Ratio Lower Upper
180 100
182 98.4 79.4 119.2
145 33.2 25.5 38.3









#31
Naphthalene
Concen: 39.48 ng
RT: 8.016 min Scan# 1008
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	128	Resp:	790745
Ion Ratio	Lower	Upper	
128	100		
129	11.3	8.7	13.1
127	13.2	10.8	16.2

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#32
Benzoic acid
Concen: 30.38 ng
RT: 7.834 min Scan# 977
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	122	Resp:	94488
Ion Ratio	Lower	Upper	
122	100		
105	124.5	103.3	143.3
77	99.6	73.9	113.9









#33
4-Chloroaniline
Concen: 39.57 ng
RT: 8.081 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	127	Resp:	337584
Ion Ratio	Lower	Upper	
127	100		
129	32.3	26.3	39.5
65	31.2	26.2	39.2
92	20.0	15.9	23.9

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#34
Hexachlorobutadiene
Concen: 42.03 ng
RT: 8.128 min Scan# 1027
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	225	Resp:	151793
Ion Ratio	Lower	Upper	
225	100		
223	62.7	51.1	76.7
227	64.2	51.0	76.6









#35
Caprolactam
Concen: 39.04 ng
RT: 8.463 min Scan# 1084
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:113 Resp: 71016
Ion Ratio Lower Upper
113 100
55 192.5 167.5 207.5
56 144.6 121.2 161.2

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#36
4-Chloro-3-methylphenol
Concen: 40.27 ng
RT: 8.575 min Scan# 1103
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:107 Resp: 247737
Ion Ratio Lower Upper
107 100
144 23.8 20.1 30.1
142 71.3 59.3 88.9









#37
2-Methylnaphthalene
Concen: 38.30 ng
RT: 8.710 min Scan# 1126
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 496784
Ion Ratio Lower Upper
142 100
141 90.1 70.2 105.4
115 37.4 28.6 43.0




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#38
1-Methylnaphthalene
Concen: 37.90 ng
RT: 8.804 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36


Tgt Ion:142 Resp: 455261
Ion Ratio Lower Upper
142 100
141 92.8 74.5 111.7
116 4.3 3.2 4.8





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.745 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:164 Resp: 176632
Ion Ratio Lower Upper
164 100
162 101.5 83.2 124.8
160 45.2 37.5 56.3

Manual Integrations
APPROVED



mohammad
11/2/2020 12:36:46 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.72 ng
RT: 8.875 min Scan# 1154
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36



Tgt Ion:216 Resp: 237835
Ion Ratio Lower Upper
216 100
214 80.0 63.4 95.2
179 21.2 17.0 25.6
108 18.3 15.5 23.3









#41
Hexachlorocyclopentadiene
Concen: 15.24 ng
RT: 8.851 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	237	Resp:	7346
Ion Ratio	Lower	Upper	
237	100		
235	60.1	41.8	81.8
272	12.7	0.0	33.5

Instrument :
BNA_F
ClientSampled :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#42
2,4,6-Tribromophenol
Concen: 72.09 ng
RT: 10.545 min Scan# 1438
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36


Tgt Ion:	330	Resp:	157524
Ion Ratio	Lower	Upper	
330	100		
332	96.1	77.3	115.9
141	30.1	22.3	33.5





#43
2,4,6-Trichlorophenol
Concen: 42.45 ng
RT: 8.998 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Instrument :
BNA_F
ClientSampled :
SSTDCCC040




Tgt Ion:196 Resp: 149321
Ion Ratio Lower Upper
196 100
198 94.1 75.8 113.6
200 29.9 24.6 36.8

Manual Integrations
APPROVED



mohammad
11/2/2020 12:36:46 PM




#44
2,4,5-Trichlorophenol
Concen: 38.12 ng
RT: 9.057 min Scan# 1185
Delta R.T. 0.012 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:196 Resp: 144807
Ion Ratio Lower Upper
196 100
198 97.6 76.1 114.1
97 51.7 41.0 61.6
132 28.5 22.4 33.6







#45
2-Fluorobiphenyl
Concen: 81.83 ng
RT: 9.069 min Scan# 1187
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:172 Resp: 982403
Ion Ratio Lower Upper
172 100
171 36.2 29.1 43.7
170 24.3 19.8 29.6




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED



mohammad
11/2/2020 12:36:46 PM




#46
1,1'-Biphenyl
Concen: 41.58 ng
RT: 9.169 min Scan# 1204
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36



Tgt Ion:154 Resp: 597214
Ion Ratio Lower Upper
154 100
153 41.0 21.3 61.3
76 14.2 0.0 34.7









#47
2-Chloronaphthalene
Concen: 40.44 ng
RT: 9.198 min Scan# 1209
Delta R.T. 0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	162	Resp:	451623
Ion	Ratio	Lower	Upper
162	100		
127	43.5	34.4	51.6
164	33.1	26.3	39.5

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#48
2-Nitroaniline
Concen: 42.79 ng
RT: 9.310 min Scan# 1228
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	65	Resp:	157637
Ion	Ratio	Lower	Upper
65	100		
92	68.4	55.1	82.7
138	95.9	80.5	120.7





#49

Acenaphthylene

Concen: 38.25 ng

RT: 9.610 min Scan# 1279

Delta R.T. -0.006 min

Lab File: BF122206.D

Acq: 31 Oct 2020 6:36

Tgt Ion:	152	Resp:	656078
Ion Ratio	Lower	Upper	
152	100		
151	20.7	16.4	24.6
153	13.2	10.6	15.8

Instrument :





BNA_F

ClientSampleId :

SSTDCCC040

Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM



#50

Dimethylphthalate

Concen: 41.79 ng




RT: 9.475 min Scan# 1256


Delta R.T. -0.006 min

Lab File: BF122206.D

Acq: 31 Oct 2020 6:36

Tgt Ion:	163	Resp:	539784
Ion Ratio	Lower	Upper	
163	100		
194	3.9	3.3	4.9
164	10.2	8.2	12.4









#51
2,6-Dinitrotoluene
Concen: 39.64 ng
RT: 9.551 min Scan# 1269
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:165 Resp: 112298
Ion Ratio Lower Upper
165 100
63 63.2 51.8 77.6
89 57.1 49.0 73.6

Instrument :
BNA_F
ClientSampled :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#52
Acenaphthene
Concen: 39.46 ng
RT: 9.781 min Scan# 1308
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:154 Resp: 402625
Ion Ratio Lower Upper
154 100
153 111.9 89.8 134.8
152 53.3 43.2 64.8









#53
3-Nitroaniline
Concen: 39.82 ng
RT: 9.728 min Scan# 1299
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:138 Resp: 128322
Ion Ratio Lower Upper
138 100
108 10.8 8.6 13.0
92 116.7 96.6 144.8

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM




#54
2,4-Dinitrophenol
Concen: 22.63 ng m
RT: 9.863 min Scan# 1322
Delta R.T. 0.018 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:184 Resp: 22609
Ion Ratio Lower Upper
184 100
63 71.9 45.9 68.9#
154 64.2 45.8 68.6









#55
Dibenzofuran
Concen: 39.31 ng
RT: 9.951 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion:168 Resp: 586590
Ion Ratio Lower Upper
168 100
139 43.1 31.0 46.6
169 14.0 11.3 16.9



#56
4-Nitrophenol
Concen: 44.79 ng
RT: 9.922 min Scan# 1332
Delta R.T. 0.012 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:139 Resp: 78512
Ion Ratio Lower Upper
139 100
109 78.4 45.4 85.4
65 107.0 68.4 108.4



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM







#57
2,4-Dinitrotoluene
Concen: 39.97 ng
RT: 9.963 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	165	Resp:	139415
Ion Ratio	Lower	Upper	
165	100		
63	52.5	41.3	61.9
89	75.8	61.8	92.8
182	2.7	2.1	3.1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM




#58
Fluorene
Concen: 38.29 ng
RT: 10.298 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:	166	Resp:	460316
Ion Ratio	Lower	Upper	
166	100		
165	99.2	79.1	118.7
167	13.0	11.0	16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 38.59 ng
RT: 10.086 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion:232 Resp: 113408
Ion Ratio Lower Upper
232 100
131 44.3 35.9 53.9
130 2.0 1.8 2.6
166 28.2 23.5 35.3



#60
Diethylphthalate
Concen: 42.17 ng
RT: 10.169 min Scan# 1374
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:149 Resp: 525236
Ion Ratio Lower Upper
149 100
177 21.1 18.1 27.1
150 12.7 10.3 15.5



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM







#61
4-Chlorophenyl-phenylether
Concen: 40.31 ng
RT: 10.286 min Scan# 1394
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:204 Resp: 232398
Ion Ratio Lower Upper
204 100
206 32.2 26.0 39.0
141 61.0 46.2 69.4

Instrument :
BNA_F
ClientSampled :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM




#62
4-Nitroaniline
Concen: 35.88 ng
RT: 10.345 min Scan# 1404
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:138 Resp: 115512
Ion Ratio Lower Upper
138 100
92 54.2 31.0 71.0
108 81.3 51.5 91.5









#63
Azobenzene
Concen: 38.73 ng
RT: 10.445 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion: 77 Resp: 510068
Ion Ratio Lower Upper
77 100
182 22.4 3.4 43.4
105 20.6 1.7 41.7
51 33.2 11.6 51.6



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.233 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion: 188 Resp: 277421
Ion Ratio Lower Upper
188 100
94 10.2 8.3 12.5
80 11.1 9.1 13.7



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM







#65
4,6-Dinitro-2-methylphenol
Concen: 17.92 ng
RT: 10.380 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 27626
Ion Ratio Lower Upper
198 100
51 54.3 29.6 69.6
105 43.2 24.4 64.4




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#66
n-Nitrosodiphenylamine
Concen: 44.00 ng
RT: 10.410 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:169 Resp: 420462
Ion Ratio Lower Upper
169 100
168 63.8 50.0 75.0
167 34.9 27.8 41.6









#67
4-Bromophenyl-phenylether
Concen: 44.06 ng
RT: 10.775 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:248 Resp: 141224
Ion Ratio Lower Upper
248 100
250 98.6 77.0 115.6
141 70.8 53.7 80.5

Instrument :
BNA_F
ClientSampled :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#68
Hexachlorobenzene
Concen: 40.70 ng
RT: 10.845 min Scan# 1489
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:284 Resp: 147642
Ion Ratio Lower Upper
284 100
142 36.7 27.4 41.2
249 31.3 24.2 36.4









#69
Atrazine
Concen: 37.63 ng
RT: 10.939 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:200 Resp: 87936
Ion Ratio Lower Upper
200 100
173 26.1 7.5 47.5
215 46.2 27.1 67.1

Instrument :
BNA_F
ClientSampled :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM




#70
Pentachlorophenol
Concen: 41.57 ng
RT: 11.057 min Scan# 1525
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:266 Resp: 53252
Ion Ratio Lower Upper
266 100
268 66.2 52.4 78.6
264 65.2 49.9 74.9






#71
Phenanthrene
Concen: 39.23 ng
RT: 11.263 min Scan# 1560
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:178 Resp: 596690
Ion Ratio Lower Upper
178 100
176 19.5 15.9 23.9
179 14.9 12.2 18.4




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED



mohammad
11/2/2020 12:36:46 PM




#72
Anthracene
Concen: 39.54 ng
RT: 11.310 min Scan# 1568
Delta R.T. -0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:178 Resp: 623750
Ion Ratio Lower Upper
178 100
176 18.9 15.7 23.5
179 15.4 12.7 19.1





#73
 Carbazole
 Concen: 39.15 ng
 RT: 11.475 min Scan# 1596
 Delta R.T. -0.000 min
 Lab File: BF122206.D
 Acq: 31 Oct 2020 6:36




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040




Tgt Ion:167 Resp: 549192
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.8 26.6
 139 13.8 10.8 16.2

Manual Integrations
 APPROVED



mohammad
 11/2/2020 12:36:46 PM




#74
 Di-n-butylphthalate
 Concen: 43.80 ng
 RT: 11.792 min Scan# 1650
 Delta R.T. 0.000 min
 Lab File: BF122206.D
 Acq: 31 Oct 2020 6:36




Tgt Ion:149 Resp: 706901
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.7 11.5
 104 5.2 4.2 6.2





#75
Fluoranthene
Concen: 36.12 ng
RT: 12.451 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:202 Resp: 589074
Ion Ratio Lower Upper
202 100
101 11.7 0.0 33.4
203 17.5 0.0 37.9

Manual Integrations
APPROVED



mohammad
11/2/2020 12:36:46 PM




#76
Chrysene-d12
Concen: 20.00 ng
RT: 13.880 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:240 Resp: 212466
Ion Ratio Lower Upper
240 100
120 11.7 9.2 13.8
236 26.2 20.9 31.3









#77
Benzidine
Concen: 36.10 ng
RT: 12.580 min Scan# 1784
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion:184 Resp: 236482
Ion Ratio Lower Upper
184 100
185 14.2 11.6 17.4
183 11.7 9.4 14.0



#78
Pyrene
Concen: 36.94 ng
RT: 12.680 min Scan# 1801
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:202 Resp: 598185
Ion Ratio Lower Upper
202 100
200 21.4 17.4 26.2
203 17.7 14.6 21.8



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM



#79
 Terphenyl-d14
 Concen: 75.35 ng
 RT: 12.816 min Scan# 1824
 Delta R.T. -0.000 min
 Lab File: BF122206.D
 Acq: 31 Oct 2020 6:36




Instrument :
 BNA_F
 ClientSampleId :
 SSTCCCC040




Tgt Ion:244 Resp: 840038
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.4 8.2
 122 10.6 8.1 12.1

Manual Integrations
 APPROVED



mohammad
 11/2/2020 12:36:46 PM




#80
 Butylbenzylphthalate
 Concen: 39.80 ng
 RT: 13.286 min Scan# 1904
 Delta R.T. -0.006 min
 Lab File: BF122206.D
 Acq: 31 Oct 2020 6:36




Tgt Ion:149 Resp: 255574
 Ion Ratio Lower Upper
 149 100
 91 78.7 61.0 91.4
 206 18.7 16.7 25.1





#81
Benzo(a)anthracene
Concen: 40.30 ng
RT: 13.874 min Scan# 2004
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Instrument :
BNA_F
ClientSampled :
SSTDCCC040




Tgt Ion:228 Resp: 561137
Ion Ratio Lower Upper
228 100
226 27.7 22.3 33.5
229 19.7 16.2 24.4

Manual Integrations
APPROVED



mohammad
11/2/2020 12:36:46 PM




#82
3,3'-Dichlorobenzidine
Concen: 41.37 ng
RT: 13.833 min Scan# 1997
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:252 Resp: 213659
Ion Ratio Lower Upper
252 100
254 66.0 51.8 77.6
126 10.7 8.8 13.2









#83
Chrysene
Concen: 39.79 ng
RT: 13.910 min Scan# 2010
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion:228 Resp: 543789
Ion Ratio Lower Upper
228 100
226 30.6 24.6 37.0
229 20.2 15.9 23.9



#84
Bis(2-ethylhexyl)phthalate
Concen: 40.41 ng
RT: 13.845 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:149 Resp: 352264
Ion Ratio Lower Upper
149 100
167 26.2 20.9 31.3
279 4.3 3.2 4.8




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED




mohammad
11/2/2020 12:36:46 PM




#85
Di-n-octyl phthalate
Concen: 43.16 ng
RT: 14.457 min Scan# 2103
Delta R.T. 0.000 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36





Tgt Ion:149 Resp: 608634
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 11.8 9.0 13.6



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.321 min Scan# 2250
Delta R.T. 0.018 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36




Tgt Ion:264 Resp: 188084
Ion Ratio Lower Upper
264 100
260 24.0 18.7 28.1
265 21.8 17.3 25.9



Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM







#87
Indeno(1,2,3-cd)pyrene
Concen: 37.84 ng
RT: 16.733 min Scan# 2490
Delta R.T. 0.018 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 276 Resp: 462084
Ion Ratio Lower Upper
276 100
138 22.8 19.9 29.9
277 25.3 20.4 30.6

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#88
Benzo(b)fluoranthene
Concen: 41.71 ng
RT: 14.910 min Scan# 2180
Delta R.T. 0.012 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 252 Resp: 530302
Ion Ratio Lower Upper
252 100
253 22.3 18.2 27.2
125 9.5 8.0 12.0









#89
Benzo(k)fluoranthene
Concen: 42.13 ng
RT: 14.933 min Scan# 2184
Delta R.T. 0.006 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:252 Resp: 508820
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.6
125 10.3 7.7 11.5

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Manual Integrations
APPROVED


mohammad
11/2/2020 12:36:46 PM



#90
Benzo(a)pyrene
Concen: 41.13 ng
RT: 15.263 min Scan# 2240
Delta R.T. 0.012 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion:252 Resp: 451697
Ion Ratio Lower Upper
252 100
253 21.9 17.3 25.9
125 10.7 8.6 13.0









#91
Dibenzo(a,h)anthracene
Concen: 39.12 ng
RT: 16.733 min Scan# 2490
Delta R.T. 0.018 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 278 Resp: 393329
Ion Ratio Lower Upper
278 100
139 14.4 13.0 19.4
279 22.7 19.0 28.4

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

mohammad
11/2/2020 12:36:46 PM



#92
Benzo(g,h,i)perylene
Concen: 36.97 ng
RT: 17.157 min Scan# 2562
Delta R.T. 0.024 min
Lab File: BF122206.D
Acq: 31 Oct 2020 6:36

Tgt Ion: 276 Resp: 372021
Ion Ratio Lower Upper
276 100
277 23.8 19.4 29.2
138 19.7 17.8 26.6

