

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF103122\
 Data File : BF130872.D
 Acq On : 31 Oct 2022 10:55
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 31 11:18:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 29 05:41:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	41.216	-3.0	93	0.01
3	Pyridine	40.000	41.726	-4.3	94	0.00
4	n-Nitrosodimethylamine	40.000	44.782	-12.0	102	0.01
5 S	2-Fluorophenol	80.000	80.372	-0.5	95	0.00
6	Aniline	40.000	40.455	-1.1	93	0.01
7 S	Phenol-d6	80.000	80.946	-1.2	96	0.01
8	2-Chlorophenol	40.000	40.888	-2.2	94	0.00
9	Benzaldehyde	40.000	36.062	9.8	93	0.00
10 C	Phenol	40.000	41.011	-2.5	95	0.01
11	bis(2-Chloroethyl)ether	40.000	41.293	-3.2	96	0.01
12	1,3-Dichlorobenzene	40.000	41.231	-3.1	95	0.00
13 C	1,4-Dichlorobenzene	40.000	41.156	-2.9	95	0.00
14	1,2-Dichlorobenzene	40.000	41.174	-2.9	95	0.00
15	Benzyl Alcohol	40.000	38.872	2.8	89	0.01
16	2,2'-oxybis(1-Chloropropane	40.000	40.624	-1.6	94	0.00
17	2-Methylphenol	40.000	41.434	-3.6	95	0.01
18	Hexachloroethane	40.000	42.091	-5.2	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.336	1.7	93	0.01
20	3+4-Methylphenols	40.000	39.541	1.1	90	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.01
22	Acetophenone	40.000	40.940	-2.3	96	0.00
23 S	Nitrobenzene-d5	80.000	95.833	-19.8	104	0.01
24	Nitrobenzene	40.000	44.987	-12.5	100	0.01
25	Isophorone	40.000	40.866	-2.2	95	0.00
26 C	2-Nitrophenol	40.000	43.115	-7.8	106	0.00
27	2,4-Dimethylphenol	40.000	40.246	-0.6	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.450	-1.1	94	0.00
29 C	2,4-Dichlorophenol	40.000	40.972	-2.4	93	0.01
30	1,2,4-Trichlorobenzene	40.000	41.153	-2.9	96	0.00
31	Naphthalene	40.000	40.399	-1.0	94	0.01
32	Benzoic acid	40.000	30.841	22.9	71	0.02
33	4-Chloroaniline	40.000	40.937	-2.3	94	0.00
34 C	Hexachlorobutadiene	40.000	42.355	-5.9	99	0.01
35	Caprolactam	40.000	40.199	-0.5	91	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.801	-4.5	95	0.01
37	2-Methylnaphthalene	40.000	40.383	-1.0	94	0.00
38	1-Methylnaphthalene	40.000	40.192	-0.5	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	92	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	43.115	-7.8	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.281	-8.2	92	0.01
42 S	2,4,6-Tribromophenol	80.000	92.715	-15.9	97	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.471	-6.2	92	0.00
44	2,4,5-Trichlorophenol	40.000	43.673	-9.2	95	0.00
45 S	2-Fluorobiphenyl	80.000	82.769	-3.5	98	0.00
46	1,1'-Biphenyl	40.000	41.627	-4.1	93	0.01
47	2-Chloronaphthalene	40.000	42.268	-5.7	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF103122\
 Data File : BF130872.D
 Acq On : 31 Oct 2022 10:55
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 31 11:18:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 29 05:41:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.843	-14.6	107	0.00
49	Acenaphthylene	40.000	41.606	-4.0	92	0.00
50	Dimethylphthalate	40.000	40.871	-2.2	91	0.01
51	2,6-Dinitrotoluene	40.000	44.763	-11.9	101	0.00
52 C	Acenaphthene	40.000	41.930	-4.8	93	0.01
53	3-Nitroaniline	40.000	49.675	-24.2	102	0.01
54 P	2,4-Dinitrophenol	40.000	45.599	-14.0	110	0.00
55	Dibenzofuran	40.000	41.255	-3.1	93	0.01
56 P	4-Nitrophenol	40.000	45.871	-14.7	92	0.00
57	2,4-Dinitrotoluene	40.000	45.908	-14.8	104	0.00
58	Fluorene	40.000	41.828	-4.6	94	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.935	-2.3	89	0.00
60	Diethylphthalate	40.000	41.570	-3.9	92	0.01
61	4-Chlorophenyl-phenylether	40.000	41.367	-3.4	94	0.00
62	4-Nitroaniline	40.000	50.586	-26.5#	102	0.00
63	Azobenzene	40.000	41.473	-3.7	91	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.729	-19.3	118	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.403	-6.0	92	0.01
67	4-Bromophenyl-phenylether	40.000	43.144	-7.9	93	0.00
68	Hexachlorobenzene	40.000	42.946	-7.4	94	0.00
69	Atrazine	40.000	43.264	-8.2	91	0.00
70 C	Pentachlorophenol	40.000	36.132	9.7	74	0.01
71	Phenanthrene	40.000	42.136	-5.3	92	0.01
72	Anthracene	40.000	42.387	-6.0	92	0.01
73	Carbazole	40.000	41.540	-3.8	91	0.01
74	Di-n-butylphthalate	40.000	41.863	-4.7	91	0.01
75 C	Fluoranthene	40.000	41.965	-4.9	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	90	0.01
77	Benzydine	40.000	36.023	9.9	78	0.01
78	Pyrene	40.000	43.212	-8.0	92	0.01
79 S	Terphenyl-d14	80.000	89.237	-11.5	97	0.01
80	Butylbenzylphthalate	40.000	46.557	-16.4	96	0.01
81	Benzo(a)anthracene	40.000	42.268	-5.7	92	0.01
82	3,3'-Dichlorobenzidine	40.000	42.810	-7.0	93	0.00
83	Chrysene	40.000	41.898	-4.7	93	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	46.033	-15.1	98	0.01
85 c	Di-n-octyl phthalate	40.000	45.394	-13.5	97	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	42.671	-6.7	87	0.02
88	Benzo(b)fluoranthene	40.000	42.112	-5.3	95	0.01
89	Benzo(k)fluoranthene	40.000	40.490	-1.2	88	0.01
90 C	Benzo(a)pyrene	40.000	41.990	-5.0	91	0.01
91	Dibenzo(a,h)anthracene	40.000	42.667	-6.7	86	0.02
92	Benzo(g,h,i)perylene	40.000	42.849	-7.1	85	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF103122\
Data File : BF130872.D
Acq On : 31 Oct 2022 10:55
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 31 11:18:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 29 05:41:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0