

Data Path : Z:\HPCHEM1\BNA F\DATA\BF110116\
 Data File : BF090965.D
 Acq On : 1 Nov 2016 21:49
 Operator : UM/SJ
 Sample : H5463-05
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-3(0-2)

Manual Integrations
 APPROVED

Umangi
 11/2/2016 5:42:04 PM

Quant Time: Nov 02 01:58:11 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 01 11:53:50 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	173441	20.00	ng	-0.01
21) Naphthalene-d8	8.02	136	746297	20.00	ng	-0.02
38) Acenaphthene-d10	9.78	164	391979	20.00	ng	-0.01
63) Phenanthrene-d10	11.25	188	545376	20.00	ng	-0.01
75) Chrysene-d12	13.89	240	342409	20.00	ng	-0.01
86) Perylene-d12	15.29	264	309678	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	847696	85.45	ng	0.00
7) Phenol-d6	6.38	99	1246284	93.12	ng	-0.01
23) Nitrobenzene-d5	7.30	82	704400	61.79	ng	-0.02
41) 2,4,6-Tribromophenol	10.57	330	156447	38.84	ng	-0.01
44) 2-Fluorobiphenyl	9.10	172	1350576	60.50	ng	-0.01
78) Terphenyl-d14	12.84	244	963056	70.86	ng	-0.01
Target Compounds						
10) Phenol	6.40	94	36140	2.45	ng	# 65
49) Dimethylphthalate	9.49	163	234169	8.94	ng	# 97
77) Pyrene	12.69	202	49394m	2.28	ng	
91) Benzo(g,h,i)perylene	17.01	276	40297	2.82	ng	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF110116\
Data File : BF090965.D
Acq On : 1 Nov 2016 21:49
Operator : UM/SJ
Sample : H5463-05
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-3(0-2)

Manual Integrations
APPROVED

Umangi
11/2/2016 5:42:04 PM

Quant Time: Nov 02 01:58:11 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 01 11:53:50 2016
Response via : Initial Calibration

