

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF110117\
 Data File : BF100044.D
 Acq On : 1 Nov 2017 15:33
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 01 18:21:21 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 01 18:13:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2	1,4-Dioxane	0.563	0.581	-3.2	75	0.00
3	Pyridine	1.353	1.368	-1.1	69	0.00
4	n-Nitrosodimethylamine	0.483	0.505	-4.6	71	0.00
5 S	2-Fluorophenol	1.274	1.423	-11.7	80	0.00
6	Aniline	1.959	2.096	-7.0	78	0.00
7 S	Phenol-d6	1.511	1.635	-8.2	79	0.00
8	2-Chlorophenol	1.358	1.500	-10.5	79	0.00
9	Benzaldehyde	0.903	0.885	2.0	71	0.00
10 C	Phenol	1.658	1.782	-7.5	77	0.00
11	bis(2-Chloroethyl)ether	1.281	1.356	-5.9	75	0.00
12	1,3-Dichlorobenzene	1.524	1.626	-6.7	78	0.00
13 C	1,4-Dichlorobenzene	1.547	1.671	-8.0	78	0.00
14	1,2-Dichlorobenzene	1.410	1.527	-8.3	79	0.00
15	Benzyl Alcohol	0.961	1.035	-7.7	80	0.00
16	2,2'-oxybis(1-Chloropropane	1.423	1.452	-2.0	74	0.00
17	2-Methylphenol	1.136	1.218	-7.2	78	0.00
18	Hexachloroethane	0.516	0.545	-5.6	77	0.00
19 P	n-Nitroso-di-n-propylamine	0.885	0.944	-6.7	79	0.00
20	3+4-Methylphenols	1.322	1.509	-14.1	85	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
22	Acetophenone	0.455	0.467	-2.6	80	0.00
23 S	Nitrobenzene-d5	0.311	0.310	0.3	77	0.00
24	Nitrobenzene	0.313	0.318	-1.6	76	0.00
25	Isophorone	0.564	0.563	0.2	76	0.00
26 C	2-Nitrophenol	0.179	0.194	-8.4	79	0.00
27	2,4-Dimethylphenol	0.264	0.281	-6.4	81	0.00
28	bis(2-Chloroethoxy)methane	0.395	0.399	-1.0	78	0.00
29 C	2,4-Dichlorophenol	0.274	0.295	-7.7	81	0.00
30	1,2,4-Trichlorobenzene	0.293	0.300	-2.4	78	0.00
31	Naphthalene	0.965	0.986	-2.2	79	0.00
32	Benzoic acid	0.233	0.238	-2.1	79	0.00
33	4-Chloroaniline	0.399	0.428	-7.3	81	0.00
34 C	Hexachlorobutadiene	0.151	0.153	-1.3	79	0.00
35	Caprolactam	0.086	0.092	-7.0	80	0.00
36 C	4-Chloro-3-methylphenol	0.280	0.290	-3.6	79	0.00
37	2-Methylnaphthalene	0.647	0.665	-2.8	80	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
39	1,2,4,5-Tetrachlorobenzene	0.646	0.661	-2.3	80	0.00
40 P	Hexachlorocyclopentadiene	0.093	0.086	7.5	69	0.00
41 S	2,4,6-Tribromophenol	0.182	0.196	-7.7	84	0.00
42 C	2,4,6-Trichlorophenol	0.391	0.396	-1.3	79	0.00
43	2,4,5-Trichlorophenol	0.415	0.416	-0.2	75	0.00
44 S	2-Fluorobiphenyl	1.296	1.341	-3.5	79	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF110117\
 Data File : BF100044.D
 Acq On : 1 Nov 2017 15:33
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 01 18:21:21 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 01 18:13:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.809	1.823	-0.8	79	0.00
46	2-Chloronaphthalene	1.314	1.321	-0.5	78	0.00
47	2-Nitroaniline	0.345	0.353	-2.3	78	0.00
48	Acenaphthylene	2.024	2.057	-1.6	80	0.00
49	Dimethylphthalate	1.584	1.535	3.1	75	0.00
50	2,6-Dinitrotoluene	0.333	0.346	-3.9	79	0.00
51 C	Acenaphthene	1.237	1.274	-3.0	82	0.00
52	3-Nitroaniline	0.392	0.414	-5.6	81	0.00
53 P	2,4-Dinitrophenol	0.114	0.106	7.0	69	0.00
54	Dibenzofuran	1.746	1.809	-3.6	82	0.00
55 P	4-Nitrophenol	0.205	0.186	9.3	67	0.00
56	2,4-Dinitrotoluene	0.401	0.444	-10.7	85	0.00
57	Fluorene	1.445	1.486	-2.8	81	0.00
58	2,3,4,6-Tetrachlorophenol	0.291	0.297	-2.1	78	0.00
59	Diethylphthalate	1.547	1.541	0.4	78	0.00
60	4-Chlorophenyl-phenylether	0.680	0.699	-2.8	82	0.00
61	4-Nitroaniline	0.374	0.391	-4.5	79	0.00
62	Azobenzene	1.297	1.290	0.5	77	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
64	4,6-Dinitro-2-methylphenol	0.126	0.115	8.7	69	0.00
65 c	n-Nitrosodiphenylamine	0.738	0.752	-1.9	82	0.00
66	4-Bromophenyl-phenylether	0.211	0.215	-1.9	81	0.00
67	Hexachlorobenzene	0.215	0.219	-1.9	80	0.00
68	Atrazine	0.222	0.223	-0.5	78	0.00
69 C	Pentachlorophenol	0.092	0.100	-8.7	86	0.00
70	Phenanthrene	1.129	1.142	-1.2	81	0.00
71	Anthracene	1.146	1.156	-0.9	81	0.00
72	Carbazole	1.077	1.118	-3.8	83	0.00
73	Di-n-butylphthalate	1.374	1.357	1.2	78	0.00
74 C	Fluoranthene	1.173	1.190	-1.4	81	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
76	Benzidine	0.875	0.953	-8.9	87	0.00
77	Pyrene	1.521	1.577	-3.7	81	0.00
78 S	Terphenyl-d14	0.915	0.929	-1.5	81	0.00
79	Butylbenzylphthalate	0.749	0.758	-1.2	77	0.00
80	Benzo(a)anthracene	1.268	1.251	1.3	76	0.00
81	3,3'-Dichlorobenzidine	0.460	0.450	2.2	75	0.00
82	Chrysene	1.192	1.188	0.3	78	0.00
83	Bis(2-ethylhexyl)phthalate	1.052	0.990	5.9	75	0.00
84 c	Di-n-octyl phthalate	1.805	1.792	0.7	76	0.00
85	Indeno(1,2,3-cd)pyrene	1.094	1.000	8.6	75	0.00
86 I	Perylene-d12	1.000	1.000	0.0	73	0.00
87	Benzo(b)fluoranthene	1.263	1.272	-0.7	76	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF110117\
Data File : BF100044.D
Acq On : 1 Nov 2017 15:33
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 01 18:21:21 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 01 18:13:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.191	1.281	-7.6	74	0.00
89 C	Benzo(a)pyrene	1.134	1.171	-3.3	74	0.00
90	Dibenzo(a,h)anthracene	1.008	1.006	0.2	75	0.00
91	Benzo(a,h,i)perylene	0.986	1.000	-1.4	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0