

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110124\
Data File : BF140201.D
Acq On : 01 Nov 2024 09:49
Operator : RC/JU
Sample : PB164558BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164558BL

Quant Time: Nov 01 11:18:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	130306	20.000	ng	-0.01
21) Naphthalene-d8	8.163	136	526301	20.000	ng	-0.01
39) Acenaphthene-d10	9.922	164	304865	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	589539	20.000	ng	0.00
76) Chrysene-d12	14.068	240	439060	20.000	ng	0.02
86) Perylene-d12	15.574	264	357043	20.000	ng	0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1023395	122.950	ng	0.02
7) Phenol-d6	6.510	99	1331053	123.469	ng	0.00
23) Nitrobenzene-d5	7.445	82	897794	94.545	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	417935	146.561	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	1626431	88.170	ng	-0.01
79) Terphenyl-d14	13.016	244	2143125	79.538	ng	0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110124\
Data File : BF140201.D
Acq On : 01 Nov 2024 09:49
Operator : RC/JU
Sample : PB164558BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164558BL

Quant Time: Nov 01 11:18:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

