




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110124\
Data File : BF140199.D
Acq On : 01 Nov 2024 08:56
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 01 11:17:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration



TIC: BF140199.D\data.ms

(70) Pentachlorophenol (C)




11.221min (+ 0.012) 111304.73 ng

response 195666

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	62.21
264.00	63.20	62.38
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110124\
Data File : BF140199.D
Acq On : 01 Nov 2024 08:56
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 01 11:17:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration



TIC: BF140199.D\data.ms

(77) Benzidine

12.762min (+ 0.017) 118399.71 ng

response 767076

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.36
183.00	11.80	11.79
0.00	0.00	0.00