

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110216\
 Data File : BF090989.D
 Acq On : 2 Nov 2016 19:35
 Operator : UM/SJ
 Sample : H5509-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-91-3.0-3.5

Manual Integrations
 APPROVED

Umangi
 11/3/2016 7:13:38 PM

Quant Time: Nov 03 01:34:42 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 01 11:53:50 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	164032	20.00	ng	-0.03
21) Naphthalene-d8	8.01	136	652324	20.00	ng	-0.03
38) Acenaphthene-d10	9.77	164	301760	20.00	ng	-0.02
63) Phenanthrene-d10	11.24	188	459530	20.00	ng	-0.02
75) Chrysene-d12	13.88	240	339210	20.00	ng	-0.02
86) Perylene-d12	15.29	264	325586	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.32	112	1281241	136.57	ng	-0.01
7) Phenol-d6	6.37	99	1733523	136.96	ng	-0.02
23) Nitrobenzene-d5	7.29	82	1058557	106.23	ng	-0.03
41) 2,4,6-Tribromophenol	10.56	330	407250	131.32	ng	-0.02
44) 2-Fluorobiphenyl	9.09	172	1428282	83.11	ng	-0.02
78) Terphenyl-d14	12.83	244	930944	69.14	ng	-0.02
Target Compounds						Qvalue
10) Phenol	6.38	94	41054	2.94	ng	# 62
14) 1,2-Dichlorobenzene	6.89	146	103766	8.70	ng	# 86
31) Naphthalene	8.03	128	279116	9.15	ng	# 99
37) 2-Methylnaphthalene	8.73	142	87660	4.45	ng	# 93
49) Dimethylphthalate	9.49	163	545996	27.08	ng	# 97
51) Acenaphthene	9.79	154	65547	3.80	ng	# 88
54) Dibenzofuran	9.96	168	85875	3.86	ng	# 83
57) Fluorene	10.30	166	92660	5.08	ng	# 91
70) Phenanthrene	11.26	178	946678	40.45	ng	# 96
71) Anthracene	11.32	178	230716	9.36	ng	# 98
72) Carbazole	11.48	167	101504	4.66	ng	# 94
73) Di-n-butylphthalate	11.81	149	195824	6.98	ng	# 98
74) Fluoranthene	12.46	202	1404595	54.81	ng	# 85
77) Pyrene	12.68	202	1086192	50.60	ng	# 95
80) Benzo(a)anthracene	13.87	228	680817	37.82	ng	# 93
82) Chrysene	13.91	228	495413m	27.21	ng	# 85
85) Indeno(1,2,3-cd)pyrene	16.60	276	336647	20.86	ng	# 85
87) Benzo(b)fluoranthene	14.90	252	872669m	39.86	ng	# 83
88) Benzo(k)fluoranthene	14.91	252	158624m	9.17	ng	# 80
89) Benzo(a)pyrene	15.23	252	555062	29.51	ng	# 83
90) Dibenzo(a,h)anthracene	16.61	278	99531	6.25	ng	# 80
91) Benzo(g,h,i)perylene	17.00	276	284237	18.95	ng	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110216\
Data File : BF090989.D
Acq On : 2 Nov 2016 19:35
Operator : UM/SJ
Sample : H5509-01
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SB-91-3.0-3.5

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM

Quant Time: Nov 03 01:34:42 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 01 11:53:50 2016
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.72 min Scan# 405
Delta R.T. -0.03 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Instrument :
BNA_F
ClientSampleId :
SB-91-3.0-3.5

Tgt Ion	Ratio	Lower	Upper
152	100		
150	164.0	124.7	187.1
115	77.2	39.0	58.4

Manual Integrations
APPROVED




Umangi
11/3/2016 7:13:38 PM






#5

2-Fluorophenol
Concen: 136.57 ng
RT: 5.32 min Scan# 283
Delta R.T. -0.01 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
112	100		
64	58.4	39.7	59.5
63	29.2	17.4	26.0





#7

Phenol-d6

Concen: 136.96 ng

RT: 6.37 min Scan# 375

Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
99	100		
42	16.7	13.7	20.5
71	34.2	14.2	21.2

Instrument :





BNA_F

ClientSampleId :

SB-91-3.0-3.5

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM



#10

Phenol

Concen: 2.94 ng


RT: 6.38 min Scan# 376


Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
94	100		
65	30.5	0.7	40.7
66	45.9	0.8	40.8






#14
1,2-Dichlorobenzene
Concen: 8.70 ng
RT: 6.89 min Scan# 420
Delta R.T. -0.03 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Instrument :
BNA_F
ClientSampled :
SB-91-3.0-3.5




Tgt Ion	Ratio	Resp	Lower	Upper
146	100	103766		
148	63.3	52.7	79.1	
111	55.6	28.8	43.2	

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM






Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E




#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.01 min Scan# 518
Delta R.T. -0.03 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Resp	Lower	Upper
136	100	652324		
137	11.4	9.0	13.6	
54	8.5	8.2	12.2	
68	7.5	5.8	8.6	

Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0









#23
 Nitrobenzene-d5
 Concen: 106.23 ng
 RT: 7.29 min Scan# 455
 Delta R.T. -0.03 min
 Lab File: BF090989.D
 Acq: 2 Nov 2016 19:35

Instrument :
 BNA_F
 ClientSampleId :
 SB-91-3.0-3.5

Tgt Ion	Ratio	Lower	Upper
82	100		
128	39.1	51.0	76.6#
54	49.5	47.5	71.3




Manual Integrations
 APPROVED


Umangi
 11/3/2016 7:13:38 PM



#31
 Naphthalene
 Concen: 9.15 ng
 RT: 8.03 min Scan# 520
 Delta R.T. -0.03 min
 Lab File: BF090989.D
 Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.4	12.6
127	12.9	9.9	14.9









#37
2-Methylnaphthalene
Concen: 4.45 ng
RT: 8.73 min Scan# 581
Delta R.T. -0.02 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Instrument :
BNA_F
ClientSampleId :
SB-91-3.0-3.5

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.6	67.5	101.3
115	29.6	18.2	27.2




Manual Integrations
APPROVED


Umangi
11/3/2016 7:13:38 PM



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.77 min Scan# 672
Delta R.T. -0.02 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.0	75.2	112.8
160	44.4	31.5	47.3









#41
 2,4,6-Tribromophenol
 Concen: 131.32 ng
 RT: 10.56 min Scan# 741
 Delta R.T. -0.02 min
 Lab File: BF090989.D
 Acq: 2 Nov 2016 19:35

Instrument :
 BNA_F
 ClientSampleId :
 SB-91-3.0-3.5

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	407250		
332	95.4	76.6	114.8	
141	25.6	29.7	44.5#	




Manual Integrations
 APPROVED


Umangi
 11/3/2016 7:13:38 PM



#44
 2-Fluorobiphenyl
 Concen: 83.11 ng
 RT: 9.09 min Scan# 613
 Delta R.T. -0.02 min
 Lab File: BF090989.D
 Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	1428282		
171	36.4	26.1	39.1	
170	24.5	17.4	26.2	









#49
Dimethylphthalate
Concen: 27.08 ng
RT: 9.49 min Scan# 648
Delta R.T. -0.02 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Instrument :
BNA_F
ClientSampleId :
SB-91-3.0-3.5

Tgt Ion	Ratio	Resp Lower	Resp Upper
163	100		
194	4.0	4.4	6.6#
164	9.7	8.3	12.5




Manual Integrations
APPROVED


Umangi
11/3/2016 7:13:38 PM



#51
Acenaphthene
Concen: 3.80 ng
RT: 9.79 min Scan# 674
Delta R.T. -0.03 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Resp Lower	Resp Upper
154	100		
153	114.9	82.1	123.1
152	54.6	37.9	56.9









#54
Dibenzofuran
Concen: 3.86 ng
RT: 9.96 min Scan# 689
Delta R.T. -0.03 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Instrument :
BNA_F
ClientSampleId :
SB-91-3.0-3.5

Tgt Ion	Ratio	Lower	Upper
168	100		
139	41.1	24.2	36.2#
169	17.9	10.6	15.8#




Manual Integrations
APPROVED


Umangi
11/3/2016 7:13:38 PM



#57
Fluorene
Concen: 5.08 ng
RT: 10.30 min Scan# 719
Delta R.T. -0.03 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.1	72.6	109.0
167	16.9	10.6	16.0#





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.24 min Scan# 801

Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Instrument :

BNA_F





ClientSampleId :

SB-91-3.0-3.5

Tgt Ion:	188	Resp:	459530
Ion Ratio	Lower	Upper	
188	100		
94	10.9	11.1	16.7#
80	11.8	11.0	16.4

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM



#70

Phenanthrene

Concen: 40.45 ng




RT: 11.26 min Scan# 803


Delta R.T. -0.03 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Tgt Ion:	178	Resp:	946678
Ion Ratio	Lower	Upper	
178	100		
176	19.8	13.8	20.8
179	16.5	12.2	18.2





#71

Anthracene

Concen: 9.36 ng

RT: 11.32 min Scan# 808

Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Instrument :

BNA_F





ClientSampleId :

SB-91-3.0-3.5

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.5	14.1	21.1
179	15.9	12.2	18.2

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM



#72

Carbazole

Concen: 4.66 ng




RT: 11.48 min Scan# 822


Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
167	100		
166	22.3	15.7	23.5
139	14.3	9.5	14.3#





#73

Di-n-butylphthalate

Concen: 6.98 ng

RT: 11.81 min Scan# 851

Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Instrument :

BNA_F





ClientSampleId :

SB-91-3.0-3.5

Tgt Ion:	149	Resp:	195824
Ion Ratio	Lower	Upper	
149	100		
150	9.6	7.4	11.2
104	4.6	2.4	3.6#

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM



#74

Fluoranthene

Concen: 54.81 ng




RT: 12.46 min Scan# 908


Delta R.T. -0.01 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Tgt Ion:	202	Resp:	1404595
Ion Ratio	Lower	Upper	
202	100		
101	6.2	0.0	38.3
203	17.9	0.0	37.3





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.88 min Scan# 1032

Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Instrument :

BNA_F





ClientSampleId :

SB-91-3.0-3.5

Tgt Ion:	240	Resp:	339210
Ion Ratio	Lower	Upper	
240	100		
120	12.2	16.2	24.2#
236	26.8	18.4	27.6

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM



#77

Pyrene

Concen: 50.60 ng




RT: 12.68 min Scan# 927


Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Tgt Ion:	202	Resp:	1086192
Ion Ratio	Lower	Upper	
202	100		
200	21.5	15.0	22.4
203	19.2	14.1	21.1









#78
 Terphenyl-d14
 Concen: 69.14 ng
 RT: 12.83 min Scan# 940
 Delta R.T. -0.02 min
 Lab File: BF090989.D
 Acq: 2 Nov 2016 19:35

Instrument :
 BNA_F
 ClientSampleId :
 SB-91-3.0-3.5

Tgt Ion	Ratio	Lower	Upper
244	100		
212	6.7	4.3	6.5#
122	11.5	17.4	26.2#




Manual Integrations
 APPROVED


Umangi
 11/3/2016 7:13:38 PM



#80
 Benzo(a)anthracene
 Concen: 37.82 ng
 RT: 13.87 min Scan# 1031
 Delta R.T. -0.02 min
 Lab File: BF090989.D
 Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.6	20.0	30.0
229	22.2	15.4	23.2





#82

Chrysene

Concen: 27.21 ng m

RT: 13.91 min Scan# 1034

Delta R.T. -0.02 min

Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Instrument :

BNA_F


ClientSampleId :

SB-91-3.0-3.5




Tgt Ion:	228	Resp:	495413
Ion Ratio	Lower	Upper	
228	100		
226	30.6	21.0	31.4
229	24.3	15.6	23.4#

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 20.86 ng

RT: 16.60 min Scan# 1270




Delta R.T. -0.03 min


Lab File: BF090989.D

Acq: 2 Nov 2016 19:35

Tgt Ion:	276	Resp:	336647
Ion Ratio	Lower	Upper	
276	100		
138	22.6	25.7	38.5#
277	24.9	15.6	23.4#

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#86
Perylene-d12
Concen: 20.00 ng
RT: 15.29 min Scan# 1155
Delta R.T. -0.01 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Instrument :
BNA_F
ClientSampleId :
SB-91-3.0-3.5

Tgt Ion	Ratio	Resp	Lower	Upper
264	100			
260	24.1	16.7	25.1	
265	23.0	17.2	25.8	




Manual Integrations
APPROVED


Umangi
11/3/2016 7:13:38 PM



#87
Benzo(b)fluoranthene
Concen: 39.86 ng m
RT: 14.90 min Scan# 1121
Delta R.T. -0.01 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Resp	Lower	Upper
252	100			
253	23.4	17.5	26.3	
125	10.7	17.1	25.7#	









#88
Benzo(k)fluoranthene
Concen: 9.17 ng m
RT: 14.91 min Scan# 1122
Delta R.T. -0.02 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Instrument :
BNA_F
ClientSampleId :
SB-91-3.0-3.5

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.7	17.0	25.4
125	12.4	16.0	24.0




Manual Integrations
APPROVED


Umangi
11/3/2016 7:13:38 PM



#89
Benzo(a)pyrene
Concen: 29.51 ng
RT: 15.23 min Scan# 1150
Delta R.T. -0.01 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.6	17.2	25.8
125	8.2	18.0	27.0









#90
Dibenzo(a,h)anthracene
Concen: 6.25 ng
RT: 16.61 min Scan# 1271
Delta R.T. -0.03 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Instrument :
BNA_F
ClientSampled :
SB-91-3.0-3.5

Tgt Ion	Ratio	Lower	Upper
278	100		
139	19.2	26.3	39.5#
279	29.6	18.2	27.4#

Manual Integrations
APPROVED

Umangi
11/3/2016 7:13:38 PM



#91
Benzo(g,h,i)perylene
Concen: 18.95 ng
RT: 17.00 min Scan# 1305
Delta R.T. -0.03 min
Lab File: BF090989.D
Acq: 2 Nov 2016 19:35

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.7	17.8	26.6
138	22.3	34.6	51.8#

