

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110217\
 Data File : BF100096.D
 Acq On : 2 Nov 2017 20:22
 Operator : SJ/JU
 Sample : I6249-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-2

Quant Time: Nov 03 02:38:21 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 01 18:13:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	84575	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	338840	20.00	ng	-0.01
38) Acenaphthene-d10	9.80	164	147897	20.00	ng	-0.01
63) Phenanthrene-d10	11.30	188	235106	20.00	ng	-0.01
75) Chrysene-d12	13.94	240	156343	20.00	ng	-0.03
86) Perylene-d12	15.41	264	153982	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	517882	96.13	ng	0.00
7) Phenol-d6	6.42	99	603969	94.55	ng	-0.01
23) Nitrobenzene-d5	7.34	82	333785	63.42	ng	-0.01
41) 2,4,6-Tribromophenol	10.60	330	119534	88.69	ng	-0.01
44) 2-Fluorobiphenyl	9.12	172	617794	64.45	ng	-0.01
78) Terphenyl-d14	12.87	244	411746	57.55	ng	-0.02
Target Compounds						
10) Phenol	6.43	94	17404	2.482	ng	Qvalue # 68
49) Dimethylphthalate	9.52	163	88623	7.567	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF110217\
Data File : BF100096.D
Acq On : 2 Nov 2017 20:22
Operator : SJ/JU
Sample : I6249-03
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP-2

Quant Time: Nov 03 02:38:21 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 01 18:13:08 2017
Response via : Initial Calibration

