




Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110217\
 Data File : BF100106.D
 Acq On : 3 Nov 2017 1:48
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 03 05:56:30 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 01 18:13:08 2017
 Response via : Initial Calibration



TIC: BF100106.D

(69) Pentachlorophenol (C)

11.116min (-0.012) 0.00ng

response 22105

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	66.33
264.00	61.90	61.79
0.00	0.00	0.00

