


Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110217\
 Data File : BF100080.D
 Acq On : 2 Nov 2017 13:07
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 01 18:13:08 2017



AutoFind: Scans 1606, 1607, 1608; Background Corrected with Scan 1602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.0	15488	PASS
68	69	0.00	2	0.9	148	PASS
69	198	0.00	100	30.4	16227	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.0	25054	PASS
197	198	0.00	2	0.5	281	PASS
198	198	100	100	100.0	53344	PASS
199	198	5	9	6.6	3495	PASS
275	198	10	60	23.5	12554	PASS
365	198	1	100	2.6	1367	PASS
441	443	0.01	100	83.0	7147	PASS
442	198	50	100	84.0	44817	PASS
443	442	15	24	19.2	8613	PASS