


Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110217\
 Data File : BF100106.D
 Acq On : 3 Nov 2017 1:48
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 01 18:13:08 2017



AutoFind: Scans 1605, 1606, 1607; Background Corrected with Scan 1601

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.8	17140	PASS
68	69	0.00	2	1.0	172	PASS
69	198	0.00	100	30.6	17005	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.4	26387	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	55640	PASS
199	198	5	9	6.8	3791	PASS
275	198	10	60	23.1	12860	PASS
365	198	1	100	2.2	1231	PASS
441	443	0.01	100	82.8	7763	PASS
442	198	50	100	88.2	49049	PASS
443	442	15	24	19.1	9379	PASS