

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110220\
 Data File : BF122219.D
 Acq On : 2 Nov 2020 11:22
 Operator : JU/CG
 Sample : PB132741TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB132741TB

Quant Time: Nov 02 14:08:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 10:16:54 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.704	152	102589	20.00	ng	-0.01
21) Naphthalene-d8	7.992	136	402253	20.00	ng	0.00
39) Acenaphthene-d10	9.745	164	207777	20.00	ng	0.00
64) Phenanthrene-d10	11.233	188	398894	20.00	ng	0.00
76) Chrysene-d12	13.880	240	326542	20.00	ng	0.00
86) Perylene-d12	15.321	264	286823	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.340	112	706926	101.70	ng	0.00
7) Phenol-d6	6.369	99	888114	99.26	ng	0.00
23) Nitrobenzene-d5	7.281	82	579326	73.86	ng	0.00
42) 2,4,6-Tribromophenol	10.545	330	260958	101.53	ng	0.00
45) 2-Fluorobiphenyl	9.069	172	999075	70.75	ng	0.00
79) Terphenyl-d14	12.821	244	1096321	63.98	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110220\
Data File : BF122219.D
Acq On : 2 Nov 2020 11:22
Operator : JU/CG
Sample : PB132741TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB132741TB

Quant Time: Nov 02 14:08:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 30 10:16:54 2020
Response via : Initial Calibration

