

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110220\
 Data File : BF122223.D
 Acq On : 2 Nov 2020 13:40
 Operator : JU/CG
 Sample : L4217-11MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HR-01-102820MSD

Manual Integrations
 APPROVED

mohammad
 11/3/2020 10:02:02 AM

Quant Time: Nov 02 14:10:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 10:16:54 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.710	152	77883	20.00	ng	0.00
21) Naphthalene-d8	7.992	136	308025	20.00	ng	0.00
39) Acenaphthene-d10	9.745	164	159181	20.00	ng	0.00
64) Phenanthrene-d10	11.233	188	312653	20.00	ng	0.00
76) Chrysene-d12	13.886	240	264173	20.00	ng	0.00
86) Perylene-d12	15.321	264	240025	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.345	112	670329	127.03	ng	0.01
7) Phenol-d6	6.375	99	748449	110.18	ng	0.00
23) Nitrobenzene-d5	7.286	82	571996	95.24	ng	0.00
42) 2,4,6-Tribromophenol	10.545	330	282835	143.64	ng	0.00
45) 2-Fluorobiphenyl	9.069	172	964982	89.20	ng	0.00
79) Terphenyl-d14	12.821	244	1154951	83.32	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	2.504	88	88651	38.20	ng	# 83
3) Pyridine	3.246	79	208400	34.15	ng	99
4) n-Nitrosodimethylamine	3.140	42	139839	47.41	ng	94
6) Aniline	6.381	93	261107	31.21	ng	# 40
8) 2-Chlorophenol	6.510	128	267123	50.08	ng	96
9) Benzaldehyde	6.281	77	38613	7.31	ng	98
10) Phenol	6.387	94	298186	42.54	ng	87
11) bis(2-Chloroethyl)ether	6.451	93	271771	50.15	ng	97
12) 1,3-Dichlorobenzene	6.645	146	276372	46.07	ng	99
13) 1,4-Dichlorobenzene	6.728	146	282940	46.97	ng	98
14) 1,2-Dichlorobenzene	6.881	146	263811	47.91	ng	99
15) Benzyl Alcohol	6.875	79	238996	54.40	ng	99
16) 2,2'-oxybis(1-Chloropr...	6.987	45	403109	48.33	ng	54
17) 2-Methylphenol	6.992	107	215242	47.20	ng	# 85
18) Hexachloroethane	7.210	117	103574	47.60	ng	98
19) n-Nitroso-di-n-propyla...	7.134	70	204912	52.98	ng	99
20) 3+4-Methylphenols	7.151	107	294981	53.64	ng	# 61
22) Acetophenone	7.128	105	384022	48.45	ng	# 90
24) Nitrobenzene	7.304	77	311729	48.56	ng	98
25) Isophorone	7.539	82	539867	47.54	ng	99
26) 2-Nitrophenol	7.622	139	144599	48.91	ng	97
27) 2,4-Dimethylphenol	7.669	122	237995	47.22	ng	99
28) bis(2-Chloroethoxy)met...	7.745	93	337433	47.49	ng	99
29) 2,4-Dichlorophenol	7.875	162	229106	48.55	ng	99
30) 1,2,4-Trichlorobenzene	7.934	180	231117	46.06	ng	99
31) Naphthalene	8.016	128	761225	46.13	ng	99
32) Benzoic acid	7.845	122	110662	39.95	ng	97
33) 4-Chloroaniline	8.081	127	205059	29.17	ng	98
34) Hexachlorobutadiene	8.122	225	140382	47.18	ng	99
35) Caprolactam	8.469	113	59554m	39.73	ng	
36) 4-Chloro-3-methylphenol	8.581	107	254286	50.17	ng	98
37) 2-Methylnaphthalene	8.704	142	504828	47.24	ng	97
38) 1-Methylnaphthalene	8.804	142	465032	46.99	ng	100
40) 1,2,4,5-Tetrachloroben...	8.875	216	243153	47.33	ng	99
41) Hexachlorocyclopentadiene	8.851	237	72346	55.46	ng	97
43) 2,4,6-Trichlorophenol	8.998	196	157004	49.53	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110220\
 Data File : BF122223.D
 Acq On : 2 Nov 2020 13:40
 Operator : JU/CG
 Sample : L4217-11MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HR-01-102820MSD


Manual Integrations
 APPROVED

mohammad
 11/3/2020 10:02:02 AM

Quant Time: Nov 02 14:10:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 10:16:54 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.057	196	154449	45.12	ng	99
46) 1,1'-Biphenyl	9.169	154	618815	47.81	ng	99
47) 2-Chloronaphthalene	9.198	162	484011	48.09	ng	99
48) 2-Nitroaniline	9.310	65	174859	52.67	ng	99
49) Acenaphthylene	9.610	152	731107	47.29	ng	99
50) Dimethylphthalate	9.475	163	615018	52.84	ng	99
51) 2,6-Dinitrotoluene	9.551	165	127256	49.84	ng	97
52) Acenaphthene	9.780	154	447712	48.69	ng	98
53) 3-Nitroaniline	9.727	138	106449	36.66	ng	96
54) 2,4-Dinitrophenol	9.851	184	128924	96.89	ng	89
55) Dibenzofuran	9.951	168	661043	49.16	ng	92
56) 4-Nitrophenol	9.928	139	64039m	40.54	ng	
57) 2,4-Dinitrotoluene	9.963	165	174580	55.54	ng	96
58) Fluorene	10.298	166	543135	50.13	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.086	232	149715	56.53	ng	95
60) Diethylphthalate	10.175	149	585353	52.15	ng	98
61) 4-Chlorophenyl-phenyle...	10.286	204	258797	49.81	ng	95
62) 4-Nitroaniline	10.351	138	136437	47.03	ng	95
63) Azobenzene	10.445	77	606940	51.13	ng	98
65) 4,6-Dinitro-2-methylph...	10.380	198	98198	46.98	ng	95
66) n-Nitrosodiphenylamine	10.410	169	504191	46.81	ng	97
67) 4-Bromophenyl-phenylether	10.775	248	172988	47.89	ng	97
68) Hexachlorobenzene	10.845	284	193841	47.41	ng	98
69) Atrazine	10.945	200	131699	50.01	ng	98
70) Pentachlorophenol	11.057	266	147081	90.35	ng	98
71) Phenanthrene	11.263	178	815768	47.59	ng	99
72) Anthracene	11.316	178	856213	48.16	ng	99
73) Carbazole	11.480	167	788608	49.88	ng	99
74) Di-n-butylphthalate	11.792	149	917562	50.45	ng	99
75) Fluoranthene	12.451	202	920425	50.07	ng	98
77) Benzidine	12.580	184	340121	41.76	ng	100
78) Pyrene	12.680	202	939086	46.65	ng	99
80) Butylbenzylphthalate	13.292	149	409954	51.35	ng	97
81) Benzo(a)anthracene	13.874	228	858108	49.57	ng	99
82) 3,3'-Dichlorobenzidine	13.833	252	231520	36.05	ng	98
83) Chrysene	13.915	228	837374	49.27	ng	99
84) Bis(2-ethylhexyl)phtha...	13.845	149	555921	51.29	ng	98
85) Di-n-octyl phthalate	14.457	149	954728	54.46	ng	99
87) Indeno(1,2,3-cd)pyrene	16.745	276	723671	46.44	ng	97
88) Benzo(b)fluoranthene	14.915	252	835077	51.47	ng	98
89) Benzo(k)fluoranthene	14.939	252	730450	47.39	ng	99
90) Benzo(a)pyrene	15.262	252	680906	48.59	ng	100
91) Dibenzo(a,h)anthracene	16.739	278	599076	46.69	ng	98
92) Benzo(g,h,i)perylene	17.162	276	599045	46.65	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.710 min Scan# 786
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:152 Resp: 77883
Ion Ratio Lower Upper
152 100
150 156.8 125.3 187.9
115 63.8 51.9 77.9

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#2
1,4-Dioxane
Concen: 38.20 ng
RT: 2.504 min Scan# 71
Delta R.T. 0.141 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 88 Resp: 88651
Ion Ratio Lower Upper
88 100
58 81.7 65.8 98.8
43 0.0 25.8 38.8#





#3
Pyridine
Concen: 34.15 ng
RT: 3.246 min Scan# 197
Delta R.T. 0.165 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40





Tgt Ion: 79 Resp: 208400

Ion	Ratio	Lower	Upper
79	100		
52	78.4	62.2	93.4
51	39.3	31.1	46.7

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Manual Integrations
APPROVED




mohammad
11/3/2020 10:02:02 AM




#4
n-Nitrosodimethylamine
Concen: 47.41 ng
RT: 3.140 min Scan# 179
Delta R.T. 0.094 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 42 Resp: 139839

Ion	Ratio	Lower	Upper
42	100		
74	117.6	100.2	150.2
44	6.8	5.8	8.6









#5
2-Fluorophenol
Concen: 127.03 ng
RT: 5.345 min Scan# 554
Delta R.T. 0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Tgt Ion:112 Resp: 670329
Ion Ratio Lower Upper
112 100
64 65.6 50.4 75.6
63 34.4 25.8 38.6




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#6
Aniline
Concen: 31.21 ng
RT: 6.381 min Scan# 730
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 93 Resp: 261107
Ion Ratio Lower Upper
93 100
66 80.8 32.7 49.1#
65 45.4 16.6 24.8#









#7
Phenol-d6
Concen: 110.18 ng
RT: 6.375 min Scan# 729
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 99 Resp: 748449
Ion Ratio Lower Upper
99 100
42 20.9 15.9 23.9
71 36.5 27.1 40.7

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#8
2-Chlorophenol
Concen: 50.08 ng
RT: 6.510 min Scan# 752
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 128 Resp: 267123
Ion Ratio Lower Upper
128 100
130 32.6 12.4 52.4
64 49.9 34.8 74.8









#9
Benzaldehyde
Concen: 7.31 ng
RT: 6.281 min Scan# 713
Delta R.T. 0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Tgt Ion: 77 Resp: 38613
Ion Ratio Lower Upper
77 100
105 96.6 76.7 116.7
106 95.9 71.3 111.3




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM




#10
Phenol
Concen: 42.54 ng
RT: 6.387 min Scan# 731
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 94 Resp: 298186
Ion Ratio Lower Upper
94 100
65 42.6 17.1 57.1
66 66.9 36.0 76.0








#11
bis(2-Chloroethyl)ether
Concen: 50.15 ng
RT: 6.451 min Scan# 742
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion: 93 Resp: 271771

Ion	Ratio	Lower	Upper
93	100		
63	79.4	62.4	102.4
95	32.2	12.6	52.6





#12
1,3-Dichlorobenzene
Concen: 46.07 ng
RT: 6.645 min Scan# 775
Delta R.T. -0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40



Tgt Ion: 146 Resp: 276372


Ion	Ratio	Lower	Upper
146	100		
148	62.9	51.0	76.4
75	35.9	27.6	41.4



Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Manual Integrations
APPROVED

mohammad
11/3/2020 10:02:02 AM







#13
1,4-Dichlorobenzene
Concen: 46.97 ng
RT: 6.728 min Scan# 789
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:146 Resp: 282940
Ion Ratio Lower Upper
146 100
148 64.4 50.1 75.1
111 42.2 34.6 52.0

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#14
1,2-Dichlorobenzene
Concen: 47.91 ng
RT: 6.881 min Scan# 815
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40


Tgt Ion:146 Resp: 263811
Ion Ratio Lower Upper
146 100
148 63.5 51.1 76.7
111 44.6 34.2 51.4





#15
Benzyl Alcohol
Concen: 54.40 ng
RT: 6.875 min Scan# 814
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion: 79 Resp: 238996
Ion Ratio Lower Upper
79 100
108 75.8 62.3 93.5
77 62.9 50.6 75.8

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 48.33 ng
RT: 6.987 min Scan# 833
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40



Tgt Ion: 45 Resp: 403109
Ion Ratio Lower Upper
45 100
77 31.6 0.0 33.3
79 27.9 0.0 30.4









#17
2-Methylphenol
Concen: 47.20 ng
RT: 6.992 min Scan# 834
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	107	Resp:	215242
Ion	Ratio	Lower	Upper
107	100		
108	111.0	90.9	136.3
77	69.2	37.5	56.3#
79	64.8	37.4	56.0#

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#18
Hexachloroethane
Concen: 47.60 ng
RT: 7.210 min Scan# 871
Delta R.T. -0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40


Tgt Ion:	117	Resp:	103574
Ion	Ratio	Lower	Upper
117	100		
119	92.8	76.3	114.5
201	91.4	71.8	107.6





#19
n-Nitroso-di-n-propylamine
Concen: 52.98 ng
RT: 7.134 min Scan# 858
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion: 70 Resp: 204912
Ion Ratio Lower Upper
70 100
42 63.9 50.2 75.4
101 9.3 7.5 11.3
130 19.1 15.5 23.3

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#20
3+4-Methylphenols
Concen: 53.64 ng
RT: 7.151 min Scan# 861
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40



Tgt Ion: 107 Resp: 294981
Ion Ratio Lower Upper
107 100
108 89.4 68.9 108.9
77 41.1 108.0 148.0#
79 29.3 10.5 50.5









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 7.992 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	136	Resp:	308025
Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.6	13.0
54	9.4	7.1	10.7
68	6.2	5.0	7.4

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#22
Acetophenone
Concen: 48.45 ng
RT: 7.128 min Scan# 857
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	105	Resp:	384022
Ion	Ratio	Lower	Upper
105	100		
71	6.2	5.3	7.9
51	31.7	33.2	49.8
120	21.4	17.4	26.0









#23
Nitrobenzene-d5
Concen: 95.24 ng
RT: 7.286 min Scan# 884
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 82 Resp: 571996
Ion Ratio Lower Upper
82 100
128 42.5 35.0 52.4
54 52.7 43.9 65.9

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#24
Nitrobenzene
Concen: 48.56 ng
RT: 7.304 min Scan# 887
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 77 Resp: 311729
Ion Ratio Lower Upper
77 100
123 46.5 35.6 53.4
65 13.9 11.2 16.8









#25
Isophorone
Concen: 47.54 ng
RT: 7.539 min Scan# 927
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 82 Resp: 539867
Ion Ratio Lower Upper
82 100
95 7.1 5.4 8.0
138 16.1 13.4 20.0

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM




#26
2-Nitrophenol
Concen: 48.91 ng
RT: 7.622 min Scan# 941
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:139 Resp: 144599
Ion Ratio Lower Upper
139 100
109 31.5 23.0 34.4
65 51.4 39.9 59.9









#27
2,4-Dimethylphenol
Concen: 47.22 ng
RT: 7.669 min Scan# 949
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40





Tgt Ion:122 Resp: 237995
Ion Ratio Lower Upper
122 100
107 115.7 90.8 136.2
121 59.2 47.5 71.3



#28
bis(2-Chloroethoxy)methane
Concen: 47.49 ng
RT: 7.745 min Scan# 962
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion: 93 Resp: 337433
Ion Ratio Lower Upper
93 100
95 32.0 26.2 39.4
123 10.4 8.8 13.2



Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD


Manual Integrations
APPROVED

mohammad
11/3/2020 10:02:02 AM



#29
2,4-Dichlorophenol
Concen: 48.55 ng
RT: 7.875 min Scan# 984
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion:162 Resp: 229106
Ion Ratio Lower Upper
162 100
164 63.8 42.5 82.5
98 40.7 20.1 60.1

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#30
1,2,4-Trichlorobenzene
Concen: 46.06 ng
RT: 7.934 min Scan# 994
Delta R.T. -0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:180 Resp: 231117
Ion Ratio Lower Upper
180 100
182 98.4 79.4 119.2
145 33.8 25.5 38.3









#31
Naphthalene
Concen: 46.13 ng
RT: 8.016 min Scan# 1008
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40





Tgt Ion:128 Resp: 761225
Ion Ratio Lower Upper
128 100
129 11.3 8.7 13.1
127 13.9 10.8 16.2



#32
Benzoic acid
Concen: 39.95 ng
RT: 7.845 min Scan# 979
Delta R.T. 0.018 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:122 Resp: 110662
Ion Ratio Lower Upper
122 100
105 125.6 103.3 143.3
77 98.5 73.9 113.9



Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Manual Integrations
APPROVED

mohammad
11/3/2020 10:02:02 AM







#33
4-Chloroaniline
Concen: 29.17 ng
RT: 8.081 min Scan# 1019
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	127	Resp:	205059
Ion Ratio	Lower	Upper	
127	100		
129	31.7	26.3	39.5
65	31.3	26.2	39.2
92	19.5	15.9	23.9

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#34
Hexachlorobutadiene
Concen: 47.18 ng
RT: 8.122 min Scan# 1026
Delta R.T. -0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	225	Resp:	140382
Ion Ratio	Lower	Upper	
225	100		
223	63.5	51.1	76.7
227	62.5	51.0	76.6









#35
Caprolactam
Concen: 39.73 ng m
RT: 8.469 min Scan# 1085
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:113 Resp: 59554
Ion Ratio Lower Upper
113 100
55 195.1 167.5 207.5
56 145.7 121.2 161.2

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#36
4-Chloro-3-methylphenol
Concen: 50.17 ng
RT: 8.581 min Scan# 1104
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40


Tgt Ion:107 Resp: 254286
Ion Ratio Lower Upper
107 100
144 24.6 20.1 30.1
142 71.9 59.3 88.9





#37
2-Methylnaphthalene
Concen: 47.24 ng
RT: 8.704 min Scan# 1125
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion:142 Resp: 504828
Ion Ratio Lower Upper
142 100
141 89.3 70.2 105.4
115 38.6 28.6 43.0

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#38
1-Methylnaphthalene
Concen: 46.99 ng
RT: 8.804 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:142 Resp: 465032
Ion Ratio Lower Upper
142 100
141 93.1 74.5 111.7
116 4.3 3.2 4.8





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.745 min Scan# 1302
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD



Tgt Ion:164 Resp: 159181
Ion Ratio Lower Upper
164 100
162 103.9 83.2 124.8
160 47.1 37.5 56.3




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#40
1,2,4,5-Tetrachlorobenzene
Concen: 47.33 ng
RT: 8.875 min Scan# 1154
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:216 Resp: 243153
Ion Ratio Lower Upper
216 100
214 78.2 63.4 95.2
179 21.3 17.0 25.6
108 19.1 15.5 23.3









#41
Hexachlorocyclopentadiene
Concen: 55.46 ng
RT: 8.851 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	237	Resp:	72346
Ion	Ratio	Lower	Upper
237	100		
235	64.5	41.8	81.8
272	13.1	0.0	33.5

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#42
2,4,6-Tribromophenol
Concen: 143.64 ng
RT: 10.545 min Scan# 1438
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	330	Resp:	282835
Ion	Ratio	Lower	Upper
330	100		
332	96.4	77.3	115.9
141	28.6	22.3	33.5









#43
2,4,6-Trichlorophenol
Concen: 49.53 ng
RT: 8.998 min Scan# 1175
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:196 Resp: 157004
Ion Ratio Lower Upper
196 100
198 96.5 75.8 113.6
200 29.9 24.6 36.8

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#44
2,4,5-Trichlorophenol
Concen: 45.12 ng
RT: 9.057 min Scan# 1185
Delta R.T. 0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:196 Resp: 154449
Ion Ratio Lower Upper
196 100
198 96.2 76.1 114.1
97 52.0 41.0 61.6
132 29.0 22.4 33.6









#45
2-Fluorobiphenyl
Concen: 89.20 ng
RT: 9.069 min Scan# 1187
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:172 Resp: 964982
Ion Ratio Lower Upper
172 100
171 36.4 29.1 43.7
170 24.2 19.8 29.6

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#46
1,1'-Biphenyl
Concen: 47.81 ng
RT: 9.169 min Scan# 1204
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:154 Resp: 618815
Ion Ratio Lower Upper
154 100
153 41.2 21.3 61.3
76 13.4 0.0 34.7









#47
2-Chloronaphthalene
Concen: 48.09 ng
RT: 9.198 min Scan# 1209
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 162 Resp: 484011
Ion Ratio Lower Upper
162 100
127 41.7 34.4 51.6
164 33.0 26.3 39.5

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#48
2-Nitroaniline
Concen: 52.67 ng
RT: 9.310 min Scan# 1228
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 65 Resp: 174859
Ion Ratio Lower Upper
65 100
92 67.9 55.1 82.7
138 99.1 80.5 120.7









#49
Acenaphthylene
Concen: 47.29 ng
RT: 9.610 min Scan# 1279
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:152 Resp: 731107
Ion Ratio Lower Upper
152 100
151 20.1 16.4 24.6
153 13.2 10.6 15.8

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#50
Dimethylphthalate
Concen: 52.84 ng
RT: 9.475 min Scan# 1256
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40


Tgt Ion:163 Resp: 615018
Ion Ratio Lower Upper
163 100
194 3.9 3.3 4.9
164 10.5 8.2 12.4





#51
2,6-Dinitrotoluene
Concen: 49.84 ng
RT: 9.551 min Scan# 1269
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion:165 Resp: 127256
Ion Ratio Lower Upper
165 100
63 67.2 51.8 77.6
89 59.3 49.0 73.6

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#52
Acenaphthene
Concen: 48.69 ng
RT: 9.780 min Scan# 1308
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40



Tgt Ion:154 Resp: 447712
Ion Ratio Lower Upper
154 100
153 114.9 89.8 134.8
152 55.1 43.2 64.8









#53
3-Nitroaniline
Concen: 36.66 ng
RT: 9.727 min Scan# 1299
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:138 Resp: 106449
Ion Ratio Lower Upper
138 100
108 10.3 8.6 13.0
92 115.4 96.6 144.8

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM




#54
2,4-Dinitrophenol
Concen: 96.89 ng
RT: 9.851 min Scan# 1320
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:184 Resp: 128924
Ion Ratio Lower Upper
184 100
63 67.8 45.9 68.9
154 63.4 45.8 68.6









#55
Dibenzenofuran
Concen: 49.16 ng
RT: 9.951 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40





Tgt Ion:168 Resp: 661043
Ion Ratio Lower Upper
168 100
139 45.1 31.0 46.6
169 14.5 11.3 16.9



#56
4-Nitrophenol
Concen: 40.54 ng m
RT: 9.928 min Scan# 1333
Delta R.T. 0.018 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:139 Resp: 64039
Ion Ratio Lower Upper
139 100
109 75.5 45.4 85.4
65 101.1 68.4 108.4



Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Manual Integrations
APPROVED

mohammad
11/3/2020 10:02:02 AM







#57
2,4-Dinitrotoluene
Concen: 55.54 ng
RT: 9.963 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	165	Resp:	174580
Ion	Ratio	Lower	Upper
165	100		
63	55.8	41.3	61.9
89	79.3	61.8	92.8
182	2.9	2.1	3.1

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM




#58
Fluorene
Concen: 50.13 ng
RT: 10.298 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:	166	Resp:	543135
Ion	Ratio	Lower	Upper
166	100		
165	99.3	79.1	118.7
167	13.3	11.0	16.6









#59
2,3,4,6-Tetrachlorophenol
Concen: 56.53 ng
RT: 10.086 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40





Tgt Ion:232 Resp: 149715
Ion Ratio Lower Upper
232 100
131 41.5 35.9 53.9
130 2.3 1.8 2.6
166 27.0 23.5 35.3



#60
Diethylphthalate
Concen: 52.15 ng
RT: 10.175 min Scan# 1375
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:149 Resp: 585353
Ion Ratio Lower Upper
149 100
177 21.6 18.1 27.1
150 12.6 10.3 15.5



Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD


Manual Integrations
APPROVED

mohammad
11/3/2020 10:02:02 AM



#61
4-Chlorophenyl-phenylether
Concen: 49.81 ng
RT: 10.286 min Scan# 1394
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion:204 Resp: 258797
Ion Ratio Lower Upper
204 100
206 33.0 26.0 39.0
141 62.8 46.2 69.4

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#62
4-Nitroaniline
Concen: 47.03 ng
RT: 10.351 min Scan# 1405
Delta R.T. 0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:138 Resp: 136437
Ion Ratio Lower Upper
138 100
92 50.1 31.0 71.0
108 77.4 51.5 91.5









#63
Azobenzene
Concen: 51.13 ng
RT: 10.445 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40





Tgt Ion: 77 Resp: 606940
Ion Ratio Lower Upper
77 100
182 22.5 3.4 43.4
105 20.5 1.7 41.7
51 32.7 11.6 51.6



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.233 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion: 188 Resp: 312653
Ion Ratio Lower Upper
188 100
94 10.5 8.3 12.5
80 11.2 9.1 13.7



Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD


Manual Integrations
APPROVED

mohammad
11/3/2020 10:02:02 AM



#65
4,6-Dinitro-2-methylphenol
Concen: 46.98 ng
RT: 10.380 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion:198 Resp: 98198
Ion Ratio Lower Upper
198 100
51 53.9 29.6 69.6
105 46.4 24.4 64.4

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#66
n-Nitrosodiphenylamine
Concen: 46.81 ng
RT: 10.410 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40



Tgt Ion:169 Resp: 504191
Ion Ratio Lower Upper
169 100
168 65.4 50.0 75.0
167 35.7 27.8 41.6









#67
4-Bromophenyl-phenylether
Concen: 47.89 ng
RT: 10.775 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:248 Resp: 172988
Ion Ratio Lower Upper
248 100
250 96.4 77.0 115.6
141 72.3 53.7 80.5

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#68
Hexachlorobenzene
Concen: 47.41 ng
RT: 10.845 min Scan# 1489
Delta R.T. -0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:284 Resp: 193841
Ion Ratio Lower Upper
284 100
142 35.5 27.4 41.2
249 31.7 24.2 36.4









#69
Atrazine
Concen: 50.01 ng
RT: 10.945 min Scan# 1506
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:200 Resp: 131699
Ion Ratio Lower Upper
200 100
173 26.2 7.5 47.5
215 46.4 27.1 67.1

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#70
Pentachlorophenol
Concen: 90.35 ng
RT: 11.057 min Scan# 1525
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:266 Resp: 147081
Ion Ratio Lower Upper
266 100
268 66.5 52.4 78.6
264 63.8 49.9 74.9









#71
Phenanthrene
Concen: 47.59 ng
RT: 11.263 min Scan# 1560
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:178 Resp: 815768
Ion Ratio Lower Upper
178 100
176 20.3 15.9 23.9
179 15.5 12.2 18.4

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#72
Anthracene
Concen: 48.16 ng
RT: 11.316 min Scan# 1569
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40


Tgt Ion:178 Resp: 856213
Ion Ratio Lower Upper
178 100
176 18.9 15.7 23.5
179 15.5 12.7 19.1





#73
 Carbazole
 Concen: 49.88 ng
 RT: 11.480 min Scan# 1597
 Delta R.T. 0.006 min
 Lab File: BF122223.D
 Acq: 2 Nov 2020 13:40




Instrument :
 BNA_F
 ClientSampleId :
 HR-01-102820MSD




Tgt Ion:167 Resp: 788608
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.8 26.6
 139 13.3 10.8 16.2

Manual Integrations
 APPROVED



mohammad
 11/3/2020 10:02:02 AM




#74
 Di-n-butylphthalate
 Concen: 50.45 ng
 RT: 11.792 min Scan# 1650
 Delta R.T. -0.000 min
 Lab File: BF122223.D
 Acq: 2 Nov 2020 13:40




Tgt Ion:149 Resp: 917562
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.7 11.5
 104 5.5 4.2 6.2





#75
Fluoranthene
Concen: 50.07 ng
RT: 12.451 min Scan# 1762
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion:202 Resp: 920425
Ion Ratio Lower Upper
202 100
101 11.8 0.0 33.4
203 17.7 0.0 37.9

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#76
Chrysene-d12
Concen: 20.00 ng
RT: 13.886 min Scan# 2006
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40



Tgt Ion:240 Resp: 264173
Ion Ratio Lower Upper
240 100
120 9.8 9.2 13.8
236 25.3 20.9 31.3









#77
Benzidine
Concen: 41.76 ng
RT: 12.580 min Scan# 1784
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:184 Resp: 340121
Ion Ratio Lower Upper
184 100
185 14.2 11.6 17.4
183 11.7 9.4 14.0

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM



#78
Pyrene
Concen: 46.65 ng
RT: 12.680 min Scan# 1801
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40


Tgt Ion:202 Resp: 939086
Ion Ratio Lower Upper
202 100
200 21.9 17.4 26.2
203 17.8 14.6 21.8





#79
Terphenyl-d14
Concen: 83.32 ng
RT: 12.821 min Scan# 1825
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion:244 Resp: 1154951
Ion Ratio Lower Upper
244 100
212 7.2 5.4 8.2
122 9.8 8.1 12.1

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#80
Butylbenzylphthalate
Concen: 51.35 ng
RT: 13.292 min Scan# 1905
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:149 Resp: 409954
Ion Ratio Lower Upper
149 100
91 73.4 61.0 91.4
206 20.7 16.7 25.1





#81
Benzo(a)anthracene
Concen: 49.57 ng
RT: 13.874 min Scan# 2004
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Tgt Ion:228 Resp: 858108
Ion Ratio Lower Upper
228 100
226 27.7 22.3 33.5
229 19.8 16.2 24.4

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#82
3,3'-Dichlorobenzidine
Concen: 36.05 ng
RT: 13.833 min Scan# 1997
Delta R.T. 0.006 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40



Tgt Ion:252 Resp: 231520
Ion Ratio Lower Upper
252 100
254 63.2 51.8 77.6
126 11.9 8.8 13.2









#83
Chrysene
Concen: 49.27 ng
RT: 13.915 min Scan# 2011
Delta R.T. 0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:228 Resp: 837374
Ion Ratio Lower Upper
228 100
226 31.3 24.6 37.0
229 20.0 15.9 23.9

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM




#84
Bis(2-ethylhexyl)phthalate
Concen: 51.29 ng
RT: 13.845 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:149 Resp: 555921
Ion Ratio Lower Upper
149 100
167 25.1 20.9 31.3
279 4.1 3.2 4.8









#85
Di-n-octyl phthalate
Concen: 54.46 ng
RT: 14.457 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40





Tgt Ion:149 Resp: 954728
Ion Ratio Lower Upper
149 100
167 1.6 1.2 1.8
43 11.5 9.0 13.6



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.321 min Scan# 2250
Delta R.T. 0.018 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:264 Resp: 240025
Ion Ratio Lower Upper
264 100
260 23.2 18.7 28.1
265 21.7 17.3 25.9



Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Manual Integrations
APPROVED

mohammad
11/3/2020 10:02:02 AM







#87
Indeno(1,2,3-cd)pyrene
Concen: 46.44 ng
RT: 16.745 min Scan# 2492
Delta R.T. 0.029 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 276 Resp: 723671
Ion Ratio Lower Upper
276 100
138 22.6 19.9 29.9
277 25.0 20.4 30.6

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD




Manual Integrations
APPROVED


mohammad
11/3/2020 10:02:02 AM




#88
Benzo(b)fluoranthene
Concen: 51.47 ng
RT: 14.915 min Scan# 2181
Delta R.T. 0.018 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion: 252 Resp: 835077
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.2
125 9.0 8.0 12.0







#89
Benzo(k)fluoranthene
Concen: 47.39 ng
RT: 14.939 min Scan# 2185
Delta R.T. 0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40




Tgt Ion:252 Resp: 730450
Ion Ratio Lower Upper
252 100
253 22.6 17.8 26.6
125 9.8 7.7 11.5




Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Manual Integrations
APPROVED



mohammad
11/3/2020 10:02:02 AM




#90
Benzo(a)pyrene
Concen: 48.59 ng
RT: 15.262 min Scan# 2240
Delta R.T. 0.012 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40



Tgt Ion:252 Resp: 680906
Ion Ratio Lower Upper
252 100
253 21.6 17.3 25.9
125 11.3 8.6 13.0









#91
Dibenzo(a,h)anthracene
Concen: 46.69 ng
RT: 16.739 min Scan# 2491
Delta R.T. 0.023 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:278 Resp: 599076
Ion Ratio Lower Upper
278 100
139 15.1 13.0 19.4
279 23.0 19.0 28.4

Instrument :
BNA_F
ClientSampleId :
HR-01-102820MSD

Manual Integrations
APPROVED

mohammad
11/3/2020 10:02:02 AM



#92
Benzo(g,h,i)perylene
Concen: 46.65 ng
RT: 17.162 min Scan# 2563
Delta R.T. 0.029 min
Lab File: BF122223.D
Acq: 2 Nov 2020 13:40

Tgt Ion:276 Resp: 599045
Ion Ratio Lower Upper
276 100
277 23.8 19.4 29.2
138 19.2 17.8 26.6

