

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110317\
 Data File : BF100148.D
 Acq On : 3 Nov 2017 21:26
 Operator : SJ/JU
 Sample : I6093-07
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TEC-B8-B9-COMP-1-4

Manual Integrations
 APPROVED

Sohil
 11/4/2017 2:19:07 PM

Quant Time: Nov 03 23:55:30 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 03 15:46:10 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	94751	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	386500	20.00	ng	0.00
38) Acenaphthene-d10	9.80	164	170735	20.00	ng	0.00
63) Phenanthrene-d10	11.30	188	271720	20.00	ng	0.00
75) Chrysene-d12	13.95	240	153368	20.00	ng	0.00
86) Perylene-d12	15.41	264	163673	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	609023	100.91	ng	0.00
7) Phenol-d6	6.42	99	730087	102.02	ng	0.00
23) Nitrobenzene-d5	7.34	82	411335	68.52	ng	-0.01
41) 2,4,6-Tribromophenol	10.60	330	128619	82.66	ng	0.00
44) 2-Fluorobiphenyl	9.12	172	758537	68.54	ng	0.00
78) Terphenyl-d14	12.88	244	495116	70.55	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.43	94	43449	5.530	ng	82
49) Dimethylphthalate	9.52	163	139774	10.338	ng	99
51) Acenaphthene	9.84	154	32842	3.111	ng	98
57) Fluorene	10.36	166	35006	2.837	ng	97
70) Phenanthrene	11.32	178	453704	29.587	ng	98
71) Anthracene	11.37	178	65833	4.229	ng	95
72) Carbazole	11.55	167	40237	2.749	ng	99
74) Fluoranthene	12.52	202	561255	35.219	ng	99
77) Pyrene	12.75	202	439836	37.715	ng	100
80) Benzo(a)anthracene	13.94	228	161482	16.609	ng	98
82) Chrysene	13.97	228	175550	19.206	ng	98
85) Indeno(1,2,3-cd)pyrene	16.89	276	102496	12.215	ng	97
87) Benzo(b)fluoranthene	14.99	252	242595m	23.473	ng	
88) Benzo(k)fluoranthene	15.02	252	93006m	9.543	ng	
89) Benzo(a)pyrene	15.35	252	160940	17.335	ng	# 97
90) Dibenzo(a,h)anthracene	16.88	278	24128m	2.925	ng	
91) Benzo(g,h,i)perylene	17.33	276	98784	12.240	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110317\
Data File : BF100148.D
Acq On : 3 Nov 2017 21:26
Operator : SJ/JU
Sample : I6093-07
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TEC-B8-B9-COMP-1-4

Manual Integrations
APPROVED

Sohil
11/4/2017 2:19:07 PM

Quant Time: Nov 03 23:55:30 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 03 15:46:10 2017
Response via : Initial Calibration

