

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110317\
 Data File : BF100153.D
 Acq On : 3 Nov 2017 23:41
 Operator : SJ/JU
 Sample : I6189-01 5X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LEONIA-GEN

Quant Time: Nov 04 00:43:13 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 03 15:46:10 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	97382	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	389750	20.00	ng	0.00
38) Acenaphthene-d10	9.80	164	167647	20.00	ng	0.00
63) Phenanthrene-d10	11.30	188	265082	20.00	ng	0.00
75) Chrysene-d12	13.95	240	152850	20.00	ng	0.00
86) Perylene-d12	15.41	264	161792	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	130018	20.96	ng	0.01
7) Phenol-d6	6.42	99	163751	22.26	ng	0.00
23) Nitrobenzene-d5	7.35	82	86789	14.34	ng	0.00
41) 2,4,6-Tribromophenol	10.60	330	29731	19.46	ng	0.00
44) 2-Fluorobiphenyl	9.12	172	192945	17.76	ng	0.00
78) Terphenyl-d14	12.87	244	110580	15.81	ng	0.00


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110317\
Data File : BF100153.D
Acq On : 3 Nov 2017 23:41
Operator : SJ/JU
Sample : I6189-01 5X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
LEONIA-GEN

Quant Time: Nov 04 00:43:13 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 03 15:46:10 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.77 min Scan# 796
Delta R.T. 0.00 min
Lab File: BF100153.D
Acq: 3 Nov 2017 23:41

Instrument :
BNA_F
ClientSampleId :
LEONIA-GEN

Tgt Ion:152 Resp: 97382
Ion Ratio Lower Upper
152 100
150 157.4 124.6 186.8
115 56.9 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 20.961 ng
RT: 5.40 min Scan# 564
Delta R.T. 0.01 min
Lab File: BF100153.D
Acq: 3 Nov 2017 23:41

Tgt Ion:112 Resp: 130018
Ion Ratio Lower Upper
112 100
64 49.1 47.9 71.9
63 23.7 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 22.265 ng

RT: 6.42 min Scan# 737

Delta R.T. -0.01 min

Lab File: BF100153.D

Acq: 3 Nov 2017 23:41

Instrument :

BNA_F

ClientSampleId :

LEONIA-GEN


Tgt Ion: 99 Resp: 163751

Ion Ratio Lower Upper

99 100

42 13.6 14.5 21.7#




71 27.5 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.05 min Scan# 1014

Delta R.T. 0.00 min

Lab File: BF100153.D

Acq: 3 Nov 2017 23:41

Tgt Ion: 136 Resp: 389750


Ion Ratio Lower Upper

136 100

137 11.0 8.9 13.3

54 6.4 6.6 9.8#

68 5.4 5.0 7.4






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1









#23
 Nitrobenzene-d5
 Concen: 14.336 ng
 RT: 7.35 min Scan# 894
 Delta R.T. -0.01 min
 Lab File: BF100153.D
 Acq: 3 Nov 2017 23:41

Instrument :
 BNA_F
 ClientSampleId :
 LEONIA-GEN

Tgt Ion: 82 Resp: 86789
 Ion Ratio Lower Upper
 82 100
 128 51.6 36.1 54.1
 54 42.5 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.80 min Scan# 1312
 Delta R.T. -0.01 min
 Lab File: BF100153.D
 Acq: 3 Nov 2017 23:41

Tgt Ion: 164 Resp: 167647
 Ion Ratio Lower Upper
 164 100
 162 104.1 86.1 129.1
 160 45.2 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): BF1
 Ion 162.00 (161.70 to 162.70): BF1
 Ion 160.00 (159.70 to 160.70): BF1









#41
 2,4,6-Tribromophenol
 Concen: 19.460 ng
 RT: 10.60 min Scan# 1448
 Delta R.T. -0.01 min
 Lab File: BF100153.D
 Acq: 3 Nov 2017 23:41

Instrument :
 BNA_F
 ClientSampleId :
 LEONIA-GEN

Tgt Ion:330 Resp: 29731
 Ion Ratio Lower Upper
 330 100
 332 95.8 77.0 115.6
 141 35.0 29.9 44.9




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 17.757 ng
 RT: 9.12 min Scan# 1196
 Delta R.T. -0.01 min
 Lab File: BF100153.D
 Acq: 3 Nov 2017 23:41

Tgt Ion:172 Resp: 192945
 Ion Ratio Lower Upper
 172 100
 171 35.4 29.7 44.5
 170 23.0 19.6 29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.30 min Scan# 1566
Delta R.T. 0.00 min
Lab File: BF100153.D
Acq: 3 Nov 2017 23:41

Instrument :
BNA_F
ClientSampleId :
LEONIA-GEN

Tgt Ion:188 Resp: 265082
Ion Ratio Lower Upper
188 100
94 9.5 8.5 12.7
80 10.0 9.2 13.8




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF1
Ion 80.00 (79.70 to 80.70): BF1






#75
Chrysene-d12
Concen: 20.000 ng
RT: 13.95 min Scan# 2017
Delta R.T. 0.00 min
Lab File: BF100153.D
Acq: 3 Nov 2017 23:41

Tgt Ion:240 Resp: 152850
Ion Ratio Lower Upper
240 100
120 12.6 9.5 14.3
236 23.9 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 15.810 ng

RT: 12.87 min Scan# 1834

Delta R.T. -0.01 min

Lab File: BF100153.D

Acq: 3 Nov 2017 23:41

Instrument :

BNA_F

ClientSampleId :

LEONIA-GEN


Tgt Ion:244 Resp: 110580

Ion Ratio Lower Upper

244 100

212 6.9 5.4 8.0




122 13.5 11.0 16.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.41 min Scan# 2265

Delta R.T. 0.00 min

Lab File: BF100153.D

Acq: 3 Nov 2017 23:41


Tgt Ion:264 Resp: 161792

Ion Ratio Lower Upper

264 100

260 23.5 19.2 28.8

265 21.6 17.5 26.3



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

