


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110317\
 Data File : BF100155.D
 Acq On : 4 Nov 2017 1:03
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 03 15:46:10 2017



AutoFind: Scans 1605, 1606, 1607; Background Corrected with Scan 1600

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.5	21301	PASS
68	69	0.00	2	1.7	376	PASS
69	198	0.00	100	30.1	22496	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.4	34717	PASS
197	198	0.00	2	0.7	490	PASS
198	198	100	100	100.0	74834	PASS
199	198	5	9	7.0	5209	PASS
275	198	10	60	23.5	17560	PASS
365	198	1	100	2.7	2030	PASS
441	443	0.01	100	82.9	11163	PASS
442	198	50	100	96.1	71922	PASS
443	442	15	24	18.7	13460	PASS