

Data Path : Z:\HPCHEM1\BNA F\DATA\BF110415\
 Data File : BF082681.D
 Acq On : 4 Nov 2015 3:40
 Operator : UM/IZ
 Sample : G4235-02
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 20151026MGP207BRV40

Quant Time: Nov 04 06:45:47 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF103015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 03 12:27:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	202562	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	747223	20.00	ng	0.00
38) Acenaphthene-d10	9.92	164	357182	20.00	ng	-0.01
63) Phenanthrene-d10	11.40	188	699376	20.00	ng	-0.01
75) Chrysene-d12	14.04	240	490386	20.00	ng	-0.01
86) Perylene-d12	15.48	264	423945	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	867797	64.52	ng	0.01
7) Phenol-d6	6.50	99	650872	36.43	ng	-0.01
23) Nitrobenzene-d5	7.44	82	1437471	78.56	ng	-0.01
41) 2,4,6-Tribromophenol	10.72	330	589243	150.74	ng	0.00
44) 2-Fluorobiphenyl	9.24	172	2328641	92.56	ng	-0.01
78) Terphenyl-d14	12.99	244	2019465	90.12	ng	-0.01
Target Compounds						Qvalue
10) Phenol	6.51	94	62678	3.10	ng	99
15) Benzyl Alcohol	7.01	79	34402m	2.06	ng	
31) Naphthalene	8.19	128	501430	12.40	ng	97
37) 2-Methylnaphthalene	8.88	142	518728	19.80	ng	98
45) 1,1'-Biphenyl	9.34	154	153092	5.07	ng	95
48) Acenaphthylene	9.78	152	740847	19.66	ng	99
49) Dimethylphthalate	9.63	163	58950	2.02	ng	99
51) Acenaphthene	9.95	154	222209	9.29	ng	97
57) Fluorene	10.46	166	282376	10.78	ng	100
70) Phenanthrene	11.42	178	294582	7.47	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF110415\
Data File : BF082681.D
Acq On : 4 Nov 2015 3:40
Operator : UM/IZ
Sample : G4235-02
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
20151026MGP207BRV40

Quant Time: Nov 04 06:45:47 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF103015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 03 12:27:47 2015
Response via : Initial Calibration

