

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110417\  
 Data File : BF100175.D  
 Acq On : 4 Nov 2017 13:49  
 Operator : SJ/JU  
 Sample : I6099-01  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 ERM-P-1

Quant Time: Nov 06 01:08:06 2017  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Nov 03 15:46:10 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.76  | 152  | 87832    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.05  | 136  | 364987   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 9.80  | 164  | 172125   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.29 | 188  | 307697   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 13.94 | 240  | 156357   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 15.40 | 264  | 136263   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.39  | 112  | 627862   | 112.23 | ng    | 0.00     |
| 7) Phenol-d6                | 6.42  | 99   | 740919   | 111.69 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.34  | 82   | 409654   | 72.26  | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 10.60 | 330  | 170691   | 108.82 | ng    | -0.01    |
| 44) 2-Fluorobiphenyl        | 9.12  | 172  | 740315   | 66.36  | ng    | -0.01    |
| 78) Terphenyl-d14           | 12.87 | 244  | 534680   | 74.73  | ng    | 0.00     |
| Target Compounds            |       |      |          |        |       |          |
| 10) Phenol                  | 6.43  | 94   | 30016    | 4.121  | ng    | # 75     |
| 49) Dimethylphthalate       | 9.52  | 163  | 32855    | 2.410  | ng    | # 99     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110417\  
Data File : BF100175.D  
Acq On : 4 Nov 2017 13:49  
Operator : SJ/JU  
Sample : I6099-01  
Misc :  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
ERM-P-1

Quant Time: Nov 06 01:08:06 2017  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Nov 03 15:46:10 2017  
Response via : Initial Calibration

