

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110417\
 Data File : BF100183.D
 Acq On : 4 Nov 2017 17:23
 Operator : SJ/JU
 Sample : I6208-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EFF-WASTEWATER

Quant Time: Nov 06 01:39:22 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 03 15:46:10 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.76	152	82165	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	338479	20.00	ng	0.00
38) Acenaphthene-d10	9.80	164	148803	20.00	ng	-0.01
63) Phenanthrene-d10	11.29	188	245640	20.00	ng	0.00
75) Chrysene-d12	13.95	240	140237	20.00	ng	0.00
86) Perylene-d12	15.40	264	150289	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	289875	55.39	ng	0.02
7) Phenol-d6	6.48	99	281478	45.36	ng	0.05
23) Nitrobenzene-d5	7.34	82	505455	96.14	ng	-0.01
41) 2,4,6-Tribromophenol	10.60	330	117977	87.00	ng	0.00
44) 2-Fluorobiphenyl	9.12	172	1046909	108.55	ng	0.00
78) Terphenyl-d14	12.87	244	688909	107.36	ng	0.00
Target Compounds						
32) Benzoic acid	7.97	122	129224	32.840	ng	93
49) Dimethylphthalate	9.52	163	45671	3.876	ng	99
83) Bis(2-ethylhexyl)phthalate	13.90	149	20122	2.729	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF110417\
Data File : BF100183.D
Acq On : 4 Nov 2017 17:23
Operator : SJ/JU
Sample : I6208-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EFF-WASTEWATER

Quant Time: Nov 06 01:39:22 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 03 15:46:10 2017
Response via : Initial Calibration

