


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF110417\
 Data File : BF100170.D
 Acq On : 4 Nov 2017 10:33
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 03 15:46:10 2017



AutoFind: Scans 1604, 1605, 1606; Background Corrected with Scan 1599

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.1	16438	PASS
68	69	0.00	2	1.7	291	PASS
69	198	0.00	100	28.9	16947	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.8	26227	PASS
197	198	0.00	2	0.4	244	PASS
198	198	100	100	100.0	58568	PASS
199	198	5	9	6.5	3812	PASS
275	198	10	60	23.4	13681	PASS
365	198	1	100	2.6	1512	PASS
441	443	0.01	100	81.7	8966	PASS
442	198	50	100	97.9	57333	PASS
443	442	15	24	19.1	10978	PASS