

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110420\
 Data File : BF122258.D
 Acq On : 4 Nov 2020 12:24
 Operator : JU/CG
 Sample : PB132766BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB132766BL

Quant Time: Nov 05 12:32:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 10:16:54 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.710	152	99479	20.00	ng	0.00
21) Naphthalene-d8	7.992	136	397935	20.00	ng	0.00
39) Acenaphthene-d10	9.745	164	207466	20.00	ng	0.00
64) Phenanthrene-d10	11.239	188	404130	20.00	ng	0.00
76) Chrysene-d12	13.886	240	297145	20.00	ng	0.00
86) Perylene-d12	15.339	264	238729	20.00	ng	0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.345	112	870864	129.20	ng	0.01
7) Phenol-d6	6.369	99	1094051	126.09	ng	0.00
23) Nitrobenzene-d5	7.286	82	731420	94.26	ng	0.00
42) 2,4,6-Tribromophenol	10.545	330	320327	124.82	ng	0.00
45) 2-Fluorobiphenyl	9.069	172	1276034	90.50	ng	0.00
79) Terphenyl-d14	12.821	244	1381348	88.59	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110420\
Data File : BF122258.D
Acq On : 4 Nov 2020 12:24
Operator : JU/CG
Sample : PB132766BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB132766BL

Quant Time: Nov 05 12:32:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF101220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 30 10:16:54 2020
Response via : Initial Calibration

