

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110422\
 Data File : BF131003.D
 Acq On : 04 Nov 2022 21:58
 Operator : CG\JU
 Sample : N5340-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-9-102722

Quant Time: Nov 05 00:40:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 04 06:20:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	56740	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	237658	20.000	ng	-0.01
39) Acenaphthene-d10	9.957	164	103053	20.000	ng	0.00
64) Phenanthrene-d10	11.451	188	185048	20.000	ng	0.00
76) Chrysene-d12	14.098	240	146066	20.000	ng	0.00
86) Perylene-d12	15.604	264	119020	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	21177	6.470	ng	0.00
7) Phenol-d6	6.528	99	99226	23.330	ng	-0.01
23) Nitrobenzene-d5	7.475	82	424623	107.736	ng	-0.01
42) 2,4,6-Tribromophenol	10.745	330	2873	3.312	ng	-0.01
45) 2-Fluorobiphenyl	9.275	172	797509	117.002	ng	0.00
79) Terphenyl-d14	13.039	244	648023	81.297	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110422\
Data File : BF131003.D
Acq On : 04 Nov 2022 21:58
Operator : CG\JU
Sample : N5340-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-9-102722

Quant Time: Nov 05 00:40:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 04 06:20:24 2022
Response via : Initial Calibration

