

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110422\
 Data File : BF131006.D
 Acq On : 04 Nov 2022 23:37
 Operator : CG\JU
 Sample : N5310-18
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BIN 1-2-(2840)-COMP-2

Quant Time: Nov 05 00:40:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 04 06:20:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	72775	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	271521	20.000	ng	-0.01
39) Acenaphthene-d10	9.957	164	133567	20.000	ng	0.00
64) Phenanthrene-d10	11.445	188	248662	20.000	ng	-0.01
76) Chrysene-d12	14.098	240	171615	20.000	ng	0.00
86) Perylene-d12	15.604	264	168692	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.534	112	432997	103.146	ng	0.00
7) Phenol-d6	6.534	99	538002	98.623	ng	0.00
23) Nitrobenzene-d5	7.475	82	373520	82.951	ng	-0.01
42) 2,4,6-Tribromophenol	10.745	330	115821	103.010	ng	-0.01
45) 2-Fluorobiphenyl	9.275	172	634682	71.841	ng	0.00
79) Terphenyl-d14	13.039	244	638901	68.220	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110422\
Data File : BF131006.D
Acq On : 04 Nov 2022 23:37
Operator : CG\JU
Sample : N5310-18
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BIN 1-2-(2840)-COMP-2

Quant Time: Nov 05 00:40:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 04 06:20:24 2022
Response via : Initial Calibration

