

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110424\
 Data File : BF140222.D
 Acq On : 04 Nov 2024 16:30
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF110424

Quant Time: Nov 04 17:14:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 04 17:12:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	37.739	5.7	98	0.00
3	Pyridine	40.000	38.166	4.6	98	0.00
4	n-Nitrosodimethylamine	40.000	34.990	12.5	89	0.00
5 S	2-Fluorophenol	80.000	76.112	4.9	100	0.00
6	Aniline	40.000	39.169	2.1	102	0.00
7 S	Phenol-d6	80.000	74.725	6.6	99	0.00
8	2-Chlorophenol	40.000	38.401	4.0	101	0.00
9	Benzaldehyde	40.000	36.136	9.7	99	0.00
10 C	Phenol	40.000	37.761	5.6	99	0.00
11	bis(2-Chloroethyl)ether	40.000	37.306	6.7	98	0.00
12	1,3-Dichlorobenzene	40.000	37.917	5.2	99	0.00
13 C	1,4-Dichlorobenzene	40.000	37.754	5.6	100	0.00
14	1,2-Dichlorobenzene	40.000	37.677	5.8	100	0.00
15	Benzyl Alcohol	40.000	37.179	7.1	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.342	11.6	93	0.00
17	2-Methylphenol	40.000	38.079	4.8	99	0.00
18	Hexachloroethane	40.000	37.859	5.4	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.724	10.7	95	0.00
20	3+4-Methylphenols	40.000	37.805	5.5	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	37.137	7.2	97	0.00
23 S	Nitrobenzene-d5	80.000	74.179	7.3	96	0.00
24	Nitrobenzene	40.000	36.762	8.1	95	0.00
25	Isophorone	40.000	37.401	6.5	97	0.00
26 C	2-Nitrophenol	40.000	41.280	-3.2	103	0.00
27	2,4-Dimethylphenol	40.000	39.414	1.5	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.889	5.3	98	0.00
29 C	2,4-Dichlorophenol	40.000	39.922	0.2	102	0.00
30	1,2,4-Trichlorobenzene	40.000	38.635	3.4	101	0.00
31	Naphthalene	40.000	38.433	3.9	100	0.00
32	Benzoic acid	40.000	44.034	-10.1	108	0.00
33	4-Chloroaniline	40.000	38.431	3.9	102	0.00
34 C	Hexachlorobutadiene	40.000	38.853	2.9	100	0.00
35	Caprolactam	40.000	39.373	1.6	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.250	4.4	99	0.00
37	2-Methylnaphthalene	40.000	37.911	5.2	99	0.00
38	1-Methylnaphthalene	40.000	38.030	4.9	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.094	2.3	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.507	-6.3	98	0.00
42 S	2,4,6-Tribromophenol	80.000	81.350	-1.7	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.677	-1.7	101	0.00
44	2,4,5-Trichlorophenol	40.000	39.387	1.5	103	0.00
45 S	2-Fluorobiphenyl	80.000	75.698	5.4	100	0.00
46	1,1'-Biphenyl	40.000	37.919	5.2	100	0.00
47	2-Chloronaphthalene	40.000	38.792	3.0	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110424\
 Data File : BF140222.D
 Acq On : 04 Nov 2024 16:30
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF110424

Quant Time: Nov 04 17:14:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 04 17:12:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.081	7.3	92	0.00
49	Acenaphthylene	40.000	38.450	3.9	101	0.00
50	Dimethylphthalate	40.000	38.418	4.0	100	0.00
51	2,6-Dinitrotoluene	40.000	39.590	1.0	102	0.00
52 C	Acenaphthene	40.000	38.413	4.0	100	0.00
53	3-Nitroaniline	40.000	39.605	1.0	100	0.00
54 P	2,4-Dinitrophenol	40.000	39.148	2.1	102	0.00
55	Dibenzofuran	40.000	38.050	4.9	99	0.00
56 P	4-Nitrophenol	40.000	40.458	-1.1	98	0.00
57	2,4-Dinitrotoluene	40.000	39.582	1.0	100	0.00
58	Fluorene	40.000	37.564	6.1	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.675	-1.7	103	0.00
60	Diethylphthalate	40.000	38.245	4.4	100	0.00
61	4-Chlorophenyl-phenylether	40.000	38.434	3.9	102	0.00
62	4-Nitroaniline	40.000	39.567	1.1	100	0.00
63	Azobenzene	40.000	36.558	8.6	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.617	-4.0	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.288	4.3	99	0.00
67	4-Bromophenyl-phenylether	40.000	39.582	1.0	103	0.00
68	Hexachlorobenzene	40.000	39.460	1.3	102	0.00
69	Atrazine	40.000	41.052	-2.6	101	0.00
70 C	Pentachlorophenol	40.000	43.268	-8.2	102	0.00
71	Phenanthrene	40.000	37.872	5.3	99	0.00
72	Anthracene	40.000	37.815	5.5	99	0.00
73	Carbazole	40.000	37.610	6.0	97	0.00
74	Di-n-butylphthalate	40.000	38.436	3.9	99	0.00
75 C	Fluoranthene	40.000	37.839	5.4	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	56.158	-40.4#	162	0.00
78	Pyrene	40.000	39.265	1.8	99	0.00
79 S	Terphenyl-d14	80.000	78.524	1.8	99	0.00
80	Butylbenzylphthalate	40.000	40.595	-1.5	100	0.00
81	Benzo(a)anthracene	40.000	38.926	2.7	99	0.00
82	3,3'-Dichlorobenzidine	40.000	40.877	-2.2	103	0.00
83	Chrysene	40.000	38.252	4.4	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.176	-0.4	101	0.00
85 c	Di-n-octyl phthalate	40.000	40.532	-1.3	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.960	-2.4	108	0.00
88	Benzo(b)fluoranthene	40.000	36.201	9.5	95	0.00
89	Benzo(k)fluoranthene	40.000	34.351	14.1	99	0.00
90 C	Benzo(a)pyrene	40.000	39.336	1.7	106	0.00
91	Dibenzo(a,h)anthracene	40.000	41.415	-3.5	108	0.00
92	Benzo(g,h,i)perylene	40.000	41.120	-2.8	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110424\
Data File : BF140222.D
Acq On : 04 Nov 2024 16:30
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF110424

Quant Time: Nov 04 17:14:36 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 04 17:12:57 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0