

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110524\
 Data File : BF140232.D
 Acq On : 05 Nov 2024 12:52
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Quant Time: Nov 05 16:05:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 05 15:55:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	166612	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	633777	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	371519	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	675726	20.000	ng	0.00
76) Chrysene-d12	14.057	240	440779	20.000	ng	0.00
86) Perylene-d12	15.562	264	369484	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	113220	11.261	ng	0.00
7) Phenol-d6	6.487	99	147074	11.373	ng	-0.02
23) Nitrobenzene-d5	7.428	82	137719	10.804	ng	-0.02
42) 2,4,6-Tribromophenol	10.698	330	37004	10.077	ng	-0.01
45) 2-Fluorobiphenyl	9.228	172	275075	11.823	ng	-0.01
79) Terphenyl-d14	12.986	244	304890	11.331	ng	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.669	88	25542	5.526	ng	99
3) Pyridine	3.440	79	60966	5.406	ng	99
4) n-Nitrosodimethylamine	3.363	42	33282	5.111	ng	# 96
6) Aniline	6.534	93	67005	5.735	ng	98
8) 2-Chlorophenol	6.657	128	60035	5.707	ng	94
9) Benzaldehyde	6.428	77	48266	6.059	ng	96
10) Phenol	6.504	94	77615	5.651	ng	93
11) bis(2-Chloroethyl)ether	6.604	93	59042	5.659	ng	99
12) 1,3-Dichlorobenzene	6.816	146	70377	5.709	ng	99
13) 1,4-Dichlorobenzene	6.892	146	71647	5.763	ng	98
14) 1,2-Dichlorobenzene	7.045	146	67578	5.820	ng	97
15) Benzyl Alcohol	7.010	79	53385	5.355	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.145	45	96009	5.685	ng	98
17) 2-Methylphenol	7.116	107	48094	5.481	ng	99
18) Hexachloroethane	7.386	117	25475	5.501	ng	97
19) n-Nitroso-di-n-propyla...	7.275	70	46525	5.640	ng	100
20) 3+4-Methylphenols	7.269	107	62607	5.681	ng	93
22) Acetophenone	7.275	105	89338	5.685	ng	# 97
24) Nitrobenzene	7.451	77	71621	5.355	ng	99
25) Isophorone	7.686	82	119458	5.442	ng	99
26) 2-Nitrophenol	7.769	139	25340	4.763	ng	99
27) 2,4-Dimethylphenol	7.798	122	38041	5.255	ng	98
28) bis(2-Chloroethoxy)met...	7.898	93	72777	5.541	ng	96
29) 2,4-Dichlorophenol	8.004	162	47065	5.256	ng	99
30) 1,2,4-Trichlorobenzene	8.092	180	59052	5.503	ng	99
31) Naphthalene	8.175	128	191018	5.837	ng	98
33) 4-Chloroaniline	8.222	127	59005	5.476	ng	99
34) Hexachlorobutadiene	8.292	225	39831	5.484	ng	100
35) Caprolactam	8.551	113	14213	5.218	ng	97
36) 4-Chloro-3-methylphenol	8.698	107	53339	5.350	ng	98
37) 2-Methylnaphthalene	8.869	142	122302	5.726	ng	99
38) 1-Methylnaphthalene	8.963	142	121198	5.789	ng	99
40) 1,2,4,5-Tetrachloroben...	9.033	216	61028	5.611	ng	99
41) Hexachlorocyclopentadiene	9.022	237	11506	3.775	ng	99
43) 2,4,6-Trichlorophenol	9.145	196	34991	5.192	ng	97
44) 2,4,5-Trichlorophenol	9.180	196	36304	5.123	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110524\
 Data File : BF140232.D
 Acq On : 05 Nov 2024 12:52
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Quant Time: Nov 05 16:05:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 05 15:55:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	9.328	154	154417	5.750	ng	98
47) 2-Chloronaphthalene	9.357	162	110964	5.490	ng	98
48) 2-Nitroaniline	9.451	65	33158	4.906	ng	100
49) Acenaphthylene	9.775	152	160677	5.615	ng	99
50) Dimethylphthalate	9.628	163	126555	5.531	ng	98
51) 2,6-Dinitrotoluene	9.692	165	28015	5.209	ng	98
52) Acenaphthene	9.945	154	111840	5.510	ng	99
53) 3-Nitroaniline	9.857	138	27313	5.393	ng	98
55) Dibenzofuran	10.116	168	164665	5.802	ng	99
56) 4-Nitrophenol	10.016	139	14822	4.163	ng	99
57) 2,4-Dinitrotoluene	10.092	165	35682	5.113	ng	96
58) Fluorene	10.463	166	134275	5.943	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.233	232	29985	5.165	ng	97
60) Diethylphthalate	10.327	149	126272	5.554	ng	100
61) 4-Chlorophenyl-phenyle...	10.457	204	66377	5.830	ng	95
62) 4-Nitroaniline	10.469	138	25854	5.137	ng	98
63) Azobenzene	10.616	77	135576	5.752	ng	98
66) n-Nitrosodiphenylamine	10.569	169	109166	5.617	ng	99
67) 4-Bromophenyl-phenylether	10.945	248	38865	5.341	ng	97
68) Hexachlorobenzene	11.010	284	44138	5.322	ng	98
69) Atrazine	11.092	200	34413	6.472	ng	98
70) Pentachlorophenol	11.204	266	17820	4.001	ng	98
71) Phenanthrene	11.427	178	182820	5.669	ng	99
72) Anthracene	11.480	178	180850	5.719	ng	98
73) Carbazole	11.633	167	164314	5.687	ng	99
74) Di-n-butylphthalate	11.963	149	189064	5.528	ng	99
75) Fluoranthene	12.616	202	195471	5.797	ng	100
77) Benzidine	12.739	184	55905	6.487	ng	99
78) Pyrene	12.845	202	203999	5.605	ng	98
80) Butylbenzylphthalate	13.468	149	63884	4.875	ng	97
81) Benzo(a)anthracene	14.045	228	152702	5.398	ng	98
82) 3,3'-Dichlorobenzidine	14.010	252	42666	4.803	ng	98
83) Chrysene	14.080	228	145493	5.492	ng	98
84) Bis(2-ethylhexyl)phtha...	14.033	149	95669	5.209	ng	98
85) Di-n-octyl phthalate	14.674	149	132302	4.786	ng	99
87) Indeno(1,2,3-cd)pyrene	17.062	276	102265	4.715	ng	98
88) Benzo(b)fluoranthene	15.121	252	120060	5.372	ng	98
89) Benzo(k)fluoranthene	15.151	252	112119	5.467	ng	99
90) Benzo(a)pyrene	15.498	252	94454	5.126	ng	100
91) Dibenzo(a,h)anthracene	17.086	278	84741	4.749	ng	99
92) Benzo(g,h,i)perylene	17.515	276	87630	4.814	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 05 16:05:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 05 15:55:36 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.875 min Scan# 81
 Delta R.T. -0.006 min
 Lab File: BF140232.D
 Acq: 05 Nov 2024 12:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005


Tgt Ion:152 Resp: 166612
 Ion Ratio Lower Upper
 152 100
 150 156.9 129.4 194.2
 115 65.6 51.5 77.3



#2
 1,4-Dioxane
 Concen: 5.526 ng
 RT: 2.669 min Scan# 98
 Delta R.T. 0.000 min
 Lab File: BF140232.D
 Acq: 05 Nov 2024 12:52

Tgt Ion: 88 Resp: 25542
 Ion Ratio Lower Upper
 88 100
 58 82.3 66.7 100.1
 43 35.3 27.8 41.8





#3

Pyridine

Concen: 5.406 ng

RT: 3.440 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion: 79 Resp: 60966

Ion Ratio Lower Upper

79 100

52 81.7 64.6 97.0

51 41.6 32.5 48.7



#4

n-Nitrosodimethylamine

Concen: 5.111 ng

RT: 3.363 min Scan# 216

Delta R.T. -0.029 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion: 42 Resp: 33282


Ion Ratio Lower Upper

42 100

74 104.4 86.9 130.3

44 7.8 5.0 7.6









#5
2-Fluorophenol
Concen: 11.261 ng
RT: 5.487 min Scan# 511
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:112 Resp: 113220
Ion Ratio Lower Upper
112 100
64 65.3 53.4 80.2
63 34.2 28.9 43.3



#6
Aniline
Concen: 5.735 ng
RT: 6.534 min Scan# 755
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion: 93 Resp: 67005
Ion Ratio Lower Upper
93 100
66 40.1 33.6 50.4
65 21.4 17.5 26.3









#7
Phenol-d6
Concen: 11.373 ng
RT: 6.487 min Scan# 747
Delta R.T. -0.018 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 99 Resp: 147074
Ion Ratio Lower Upper
99 100
42 22.1 18.2 27.2
71 37.3 30.9 46.3



#8
2-Chlorophenol
Concen: 5.707 ng
RT: 6.657 min Scan# 776
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion: 128 Resp: 60035
Ion Ratio Lower Upper
128 100
130 30.8 12.9 52.9
64 50.8 36.3 76.3









#9
Benzaldehyde
Concen: 6.059 ng
RT: 6.428 min Scan# 71
Delta R.T. -0.000 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 77 Resp: 48266
Ion Ratio Lower Upper
77 100
105 95.9 72.1 112.1
106 91.8 68.8 108.8



#10
Phenol
Concen: 5.651 ng
RT: 6.504 min Scan# 750
Delta R.T. -0.018 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion: 94 Resp: 77615
Ion Ratio Lower Upper
94 100
65 31.6 15.1 55.1
66 43.3 28.8 68.8





#11

bis(2-Chloroethyl)ether

Concen: 5.659 ng

RT: 6.604 min Scan# 769

Delta R.T. -0.012 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion: 93 Resp: 59042

Ion Ratio Lower Upper

93 100

63 82.8 63.6 103.6

95 31.6 12.7 52.7



#12

1,3-Dichlorobenzene

Concen: 5.709 ng

RT: 6.816 min Scan# 803

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:146 Resp: 70377


Ion Ratio Lower Upper

146 100

148 64.3 51.7 77.5

75 34.8 28.4 42.6









#13
1,4-Dichlorobenzene
Concen: 5.763 ng
RT: 6.892 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:146 Resp: 71647
Ion Ratio Lower Upper
146 100
148 62.2 51.8 77.6
111 43.3 35.4 53.2



#14
1,2-Dichlorobenzene
Concen: 5.820 ng
RT: 7.045 min Scan# 842
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:146 Resp: 67578
Ion Ratio Lower Upper
146 100
148 62.9 51.8 77.6
111 44.8 37.9 56.9









#15
Benzyl Alcohol
Concen: 5.355 ng
RT: 7.010 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 79 Resp: 53385
Ion Ratio Lower Upper
79 100
108 71.8 55.8 83.6
77 65.0 51.1 76.7



#16
2,2'-oxybis(1-chloropropane)
Concen: 5.685 ng
RT: 7.145 min Scan# 859
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion: 45 Resp: 96009
Ion Ratio Lower Upper
45 100
77 12.4 0.0 33.3
79 9.5 0.0 29.9









#17
2-Methylphenol
Concen: 5.481 ng
RT: 7.116 min Scan# 81
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:	107	Resp:	48094
Ion	Ratio	Lower	Upper
107	100		
108	113.7	90.5	135.7
77	49.8	41.8	62.6
79	50.2	40.6	61.0



#18
Hexachloroethane
Concen: 5.501 ng
RT: 7.386 min Scan# 900
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:	117	Resp:	25475
Ion	Ratio	Lower	Upper
117	100		
119	96.6	77.3	115.9
201	99.6	85.1	127.7









#19
n-Nitroso-di-n-propylamine
Concen: 5.640 ng
RT: 7.275 min Scan# 81
Delta R.T. -0.018 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 70 Resp: 46525




Ion	Ratio	Lower	Upper
70	100		
42	66.0	52.8	79.2
101	9.7	7.4	11.2
130	19.1	15.1	22.7




#20
3+4-Methylphenols
Concen: 5.681 ng
RT: 7.269 min Scan# 880
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion: 107 Resp: 62607

Ion	Ratio	Lower	Upper
107	100		
108	93.8	69.4	109.4
77	122.8	92.7	132.7
79	33.0	13.9	53.9









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:136 Resp: 633777




Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.6
54	9.2	7.7	11.5
68	6.3	5.2	7.8




#22
Acetophenone
Concen: 5.685 ng
RT: 7.275 min Scan# 881
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:105 Resp: 89338

Ion	Ratio	Lower	Upper
105	100		
71	9.2	5.6	8.4
51	39.4	33.2	49.8
120	21.4	16.8	25.2





#23

Nitrobenzene-d5

Concen: 10.804 ng

RT: 7.428 min Scan# 90

Delta R.T. -0.018 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion: 82 Resp: 137719

Ion Ratio Lower Upper

82 100

128 38.8 32.8 49.2

54 55.9 44.3 66.5



#24

Nitrobenzene

Concen: 5.355 ng

RT: 7.451 min Scan# 911

Delta R.T. -0.012 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion: 77 Resp: 71621


Ion Ratio Lower Upper

77 100

123 40.8 33.0 49.4

65 14.5 11.7 17.5









#25
Isophorone
Concen: 5.442 ng
RT: 7.686 min Scan# 91
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion: 82 Resp: 119458
Ion Ratio Lower Upper
82 100
95 7.0 6.0 9.0
138 16.0 12.9 19.3



#26
2-Nitrophenol
Concen: 4.763 ng
RT: 7.769 min Scan# 965
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:139 Resp: 25340
Ion Ratio Lower Upper
139 100
109 33.8 27.8 41.8
65 60.0 48.2 72.2





#27

2,4-Dimethylphenol

Concen: 5.255 ng

RT: 7.798 min Scan# 91

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:122 Resp: 38041

Ion Ratio Lower Upper

122 100

107 121.7 99.4 149.0

121 57.9 47.1 70.7



#28

bis(2-Chloroethoxy)methane

Concen: 5.541 ng

RT: 7.898 min Scan# 987

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion: 93 Resp: 72777


Ion Ratio Lower Upper

93 100

95 30.4 26.1 39.1

123 10.4 9.4 14.0





#29

2,4-Dichlorophenol

Concen: 5.256 ng

RT: 8.004 min Scan# 1005

Delta R.T. -0.012 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:162 Resp: 47065

Ion Ratio Lower Upper

162 100

164 66.1 47.5 87.5

98 39.6 19.5 59.5



#30

1,2,4-Trichlorobenzene

Concen: 5.503 ng

RT: 8.092 min Scan# 1020

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:180 Resp: 59052


Ion Ratio Lower Upper

180 100

182 95.8 75.9 113.9

145 32.5 26.1 39.1









#31
Naphthalene
Concen: 5.837 ng
RT: 8.175 min Scan# 1034
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:128 Resp: 191018
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.6
127 12.8 10.7 16.1



#33
4-Chloroaniline
Concen: 5.476 ng
RT: 8.222 min Scan# 1042
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:127 Resp: 59005
Ion Ratio Lower Upper
127 100
129 32.6 25.8 38.6
65 37.1 29.0 43.6
92 20.7 17.4 26.0









#34
Hexachlorobutadiene
Concen: 5.484 ng
RT: 8.292 min Scan# 1055
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:225 Resp: 39831
Ion Ratio Lower Upper
225 100
223 63.8 50.8 76.2
227 64.2 51.0 76.6



#35
Caprolactam
Concen: 5.218 ng
RT: 8.551 min Scan# 1098
Delta R.T. -0.059 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:113 Resp: 14213
Ion Ratio Lower Upper
113 100
55 182.5 165.2 205.2
56 135.6 122.1 162.1









#36
4-Chloro-3-methylphenol
Concen: 5.350 ng
RT: 8.698 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:107 Resp: 53339
Ion Ratio Lower Upper
107 100
144 24.1 18.6 27.8
142 74.7 58.1 87.1



#37
2-Methylnaphthalene
Concen: 5.726 ng
RT: 8.869 min Scan# 1152
Delta R.T. -0.000 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:142 Resp: 122302
Ion Ratio Lower Upper
142 100
141 84.3 67.8 101.8





#38

1-Methylnaphthalene

Concen: 5.789 ng

RT: 8.963 min Scan# 1168

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :





SSTDICC005

Tgt Ion:142 Resp: 121198

Ion Ratio Lower Upper

142 100

141 86.4 69.7 104.5



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.916 min Scan# 1330

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:164 Resp: 371519


Ion Ratio Lower Upper

164 100

162 96.8 77.0 115.4

160 43.1 34.1 51.1





#40

1,2,4,5-Tetrachlorobenzene

Concen: 5.611 ng

RT: 9.033 min Scan# 1181

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion:216 Resp: 61028





Ion Ratio Lower Upper

216 100

214 77.9 62.8 94.2

179 20.9 17.3 25.9

108 17.2 15.3 22.9



#41

Hexachlorocyclopentadiene

Concen: 3.775 ng

RT: 9.022 min Scan# 1178

Delta R.T. -0.000 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:237 Resp: 11506


Ion Ratio Lower Upper

237 100

235 60.8 41.8 81.8

272 11.4 0.0 31.8





#42

2,4,6-Tribromophenol

Concen: 10.077 ng

RT: 10.698 min Scan# 1199

Delta R.T. -0.012 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:330 Resp: 37004

Ion Ratio Lower Upper

330 100

332 97.3 76.4 114.6

141 36.3 27.0 40.4



#43

2,4,6-Trichlorophenol

Concen: 5.192 ng

RT: 9.145 min Scan# 1199

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:196 Resp: 34991


Ion Ratio Lower Upper

196 100

198 97.4 80.8 121.2

200 32.2 26.0 39.0









#44
2,4,5-Trichlorophenol
Concen: 5.123 ng
RT: 9.180 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:196	Resp:	36304
Ion Ratio	Lower	Upper
196	100	
198	98.6	80.0 120.0
97	56.4	45.7 68.5
132	28.6	23.0 34.6



#45
2-Fluorobiphenyl
Concen: 11.823 ng
RT: 9.228 min Scan# 1213
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:172	Resp:	275075
Ion Ratio	Lower	Upper
172	100	
171	35.2	28.8 43.2
170	23.5	19.6 29.4









#46
1,1'-Biphenyl
Concen: 5.750 ng
RT: 9.328 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:154 Resp: 154417
Ion Ratio Lower Upper
154 100
153 39.6 20.8 60.8
76 12.4 0.0 32.6



#47
2-Chloronaphthalene
Concen: 5.490 ng
RT: 9.357 min Scan# 1235
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:162 Resp: 110964
Ion Ratio Lower Upper
162 100
127 42.8 34.8 52.2
164 35.3 27.3 40.9





#48

2-Nitroaniline

Concen: 4.906 ng

RT: 9.451 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion: 65 Resp: 33158

Ion Ratio Lower Upper

65 100

92 64.9 51.6 77.4

138 91.2 73.0 109.6



#49

Acenaphthylene

Concen: 5.615 ng

RT: 9.775 min Scan# 1306

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion: 152 Resp: 160677


Ion Ratio Lower Upper

152 100

151 20.1 16.4 24.6

153 13.5 11.1 16.7









#50
Dimethylphthalate
Concen: 5.531 ng
RT: 9.628 min Scan# 1281
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:163 Resp: 126555
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 10.0 8.6 13.0



#51
2,6-Dinitrotoluene
Concen: 5.209 ng
RT: 9.692 min Scan# 1292
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:165 Resp: 28015
Ion Ratio Lower Upper
165 100
63 59.7 48.8 73.2
89 58.8 48.1 72.1









#52
Acenaphthene
Concen: 5.510 ng
RT: 9.945 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:154 Resp: 111840
Ion Ratio Lower Upper
154 100
153 112.6 91.3 136.9
152 51.3 41.4 62.2



#53
3-Nitroaniline
Concen: 5.393 ng
RT: 9.857 min Scan# 1320
Delta R.T. -0.018 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:138 Resp: 27313
Ion Ratio Lower Upper
138 100
108 11.7 8.8 13.2
92 123.1 100.0 150.0









#55
Dibenzenofuran
Concen: 5.802 ng
RT: 10.116 min Scan# 11
Delta R.T. -0.011 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:168 Resp: 164665
Ion Ratio Lower Upper
168 100
139 38.7 31.5 47.3
169 13.4 10.6 15.8



#56
4-Nitrophenol
Concen: 4.163 ng
RT: 10.016 min Scan# 1347
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:139 Resp: 14822
Ion Ratio Lower Upper
139 100
109 83.4 63.8 103.8
65 116.3 94.7 134.7









#57
2,4-Dinitrotoluene
Concen: 5.113 ng
RT: 10.092 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:	165	Resp:	35682
Ion	Ratio	Lower	Upper
165	100		
63	49.0	41.3	61.9
89	74.2	62.6	94.0
182	2.6	2.0	3.0



#58
Fluorene
Concen: 5.943 ng
RT: 10.463 min Scan# 1423
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:	166	Resp:	134275
Ion	Ratio	Lower	Upper
166	100		
165	97.1	78.7	118.1
167	12.8	10.8	16.2









#59
2,3,4,6-Tetrachlorophenol
Concen: 5.165 ng
RT: 10.233 min Scan# 1384
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:232 Resp: 29985
Ion Ratio Lower Upper
232 100
131 44.9 37.2 55.8
130 1.9 1.8 2.8
166 32.2 24.3 36.5



#60
Diethylphthalate
Concen: 5.554 ng
RT: 10.327 min Scan# 1400
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:149 Resp: 126272
Ion Ratio Lower Upper
149 100
177 22.6 18.1 27.1
150 12.7 10.2 15.4









#61
4-Chlorophenyl-phenylether
Concen: 5.830 ng
RT: 10.457 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:204 Resp: 66377
Ion Ratio Lower Upper
204 100
206 31.1 26.2 39.4
141 58.5 50.3 75.5



#62
4-Nitroaniline
Concen: 5.137 ng
RT: 10.469 min Scan# 1424
Delta R.T. -0.024 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:138 Resp: 25854
Ion Ratio Lower Upper
138 100
92 59.3 35.3 75.3
108 89.7 70.1 110.1





#63

Azobenzene

Concen: 5.752 ng

RT: 10.616 min Scan# 1449

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion: 77 Resp: 135576





Ion Ratio Lower Upper

77 100

182 23.9 4.7 44.7

105 19.0 0.0 40.0

51 31.5 13.2 53.2



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.404 min Scan# 1583

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion: 188 Resp: 675726


Ion Ratio Lower Upper

188 100

94 8.9 6.9 10.3

80 10.0 7.5 11.3









#66
n-Nitrosodiphenylamine
Concen: 5.617 ng
RT: 10.569 min Scan# 1441
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:169 Resp: 109166
Ion Ratio Lower Upper
169 100
168 66.2 53.0 79.6
167 35.5 29.4 44.2



#67
4-Bromophenyl-phenylether
Concen: 5.341 ng
RT: 10.945 min Scan# 1505
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:248 Resp: 38865
Ion Ratio Lower Upper
248 100
250 99.0 77.9 116.9
141 77.2 58.6 88.0









#68
Hexachlorobenzene
Concen: 5.322 ng
RT: 11.010 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:284 Resp: 44138
Ion Ratio Lower Upper
284 100
142 35.6 27.5 41.3
249 32.0 26.2 39.2



#69
Atrazine
Concen: 6.472 ng
RT: 11.092 min Scan# 1530
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:200 Resp: 34413
Ion Ratio Lower Upper
200 100
173 25.3 4.4 44.4
215 46.9 25.2 65.2





#70

Pentachlorophenol

Concen: 4.001 ng

RT: 11.204 min Scan# 1549

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:266 Resp: 17820

Ion Ratio Lower Upper

266 100

268 62.5 50.2 75.4

264 63.1 52.2 78.4



#71

Phenanthrene

Concen: 5.669 ng

RT: 11.427 min Scan# 1587

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:178 Resp: 182820


Ion Ratio Lower Upper

178 100

176 19.8 15.9 23.9

179 14.8 12.6 18.8





#72

Anthracene

Concen: 5.719 ng

RT: 11.480 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:178 Resp: 180850

Ion Ratio Lower Upper

178 100

176 18.0 15.4 23.0

179 15.0 12.7 19.1



#73

Carbazole

Concen: 5.687 ng

RT: 11.633 min Scan# 1622

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:167 Resp: 164314


Ion Ratio Lower Upper

167 100

166 22.0 17.3 25.9

139 12.7 10.7 16.1





#74

Di-n-butylphthalate

Concen: 5.528 ng

RT: 11.963 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:149 Resp: 189064

Ion Ratio Lower Upper

149 100

150 9.7 7.4 11.2

104 5.6 4.5 6.7



#75

Fluoranthene

Concen: 5.797 ng

RT: 12.616 min Scan# 1789

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:202 Resp: 195471


Ion Ratio Lower Upper

202 100

101 10.5 0.0 30.4

203 17.9 0.0 38.0





#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.057 min Scan# 2034

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:240 Resp: 440779

Ion Ratio Lower Upper

240 100

120 9.5 7.5 11.3

236 26.4 20.9 31.3



#77

Benzidine

Concen: 6.487 ng

RT: 12.739 min Scan# 1810

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:184 Resp: 55905


Ion Ratio Lower Upper

184 100

185 15.5 12.5 18.7

183 11.0 9.3 13.9





#78

Pyrene

Concen: 5.605 ng

RT: 12.845 min Scan# 1828

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:202 Resp: 203999

Ion Ratio Lower Upper

202 100

200 20.8 17.5 26.3

203 17.0 14.4 21.6



#79

Terphenyl-d14

Concen: 11.331 ng

RT: 12.986 min Scan# 1852

Delta R.T. -0.012 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:244 Resp: 304890


Ion Ratio Lower Upper

244 100

212 7.3 6.0 9.0

122 10.1 7.6 11.4





#80

Butylbenzylphthalate

Concen: 4.875 ng

RT: 13.468 min Scan# 1934

Delta R.T. 0.000 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:149 Resp: 63884

Ion Ratio Lower Upper

149 100

91 75.7 63.0 94.4

206 21.3 16.4 24.6



#81

Benzo(a)anthracene

Concen: 5.398 ng

RT: 14.045 min Scan# 2032

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:228 Resp: 152702


Ion Ratio Lower Upper

228 100

226 26.7 22.3 33.5

229 19.8 16.1 24.1









#82
3,3'-Dichlorobenzidine
Concen: 4.803 ng
RT: 14.010 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 42666
Ion Ratio Lower Upper
252 100
254 63.2 51.8 77.8
126 10.1 8.6 12.8



#83
Chrysene
Concen: 5.492 ng
RT: 14.080 min Scan# 2038
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:228 Resp: 145493
Ion Ratio Lower Upper
228 100
226 29.5 24.6 36.8
229 19.2 16.1 24.1









#84
Bis(2-ethylhexyl)phthalate
Concen: 5.209 ng
RT: 14.033 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 95669
Ion Ratio Lower Upper
149 100
167 26.2 21.8 32.8
279 3.5 3.2 4.8



#85
Di-n-octyl phthalate
Concen: 4.786 ng
RT: 14.674 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:149 Resp: 132302
Ion Ratio Lower Upper
149 100
167 1.7 1.3 1.9
43 11.7 9.7 14.5





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.562 min Scan# 21

Delta R.T. -0.000 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:264 Resp: 369484

Ion Ratio Lower Upper

264 100

260 23.9 18.8 28.2

265 21.7 17.3 25.9



#87

Indeno(1,2,3-cd)pyrene

Concen: 4.715 ng

RT: 17.062 min Scan# 2545

Delta R.T. -0.018 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:276 Resp: 102265


Ion Ratio Lower Upper

276 100

138 22.5 17.6 26.4

277 24.3 20.2 30.2





#88

Benzo(b)fluoranthene

Concen: 5.372 ng

RT: 15.121 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:252 Resp: 120060

Ion Ratio Lower Upper

252 100

253 21.6 17.9 26.9

125 9.3 7.0 10.6



#89

Benzo(k)fluoranthene

Concen: 5.467 ng

RT: 15.151 min Scan# 2220

Delta R.T. -0.006 min

Lab File: BF140232.D

Acq: 05 Nov 2024 12:52




Tgt Ion:252 Resp: 112119


Ion Ratio Lower Upper

252 100

253 21.6 17.8 26.8

125 9.1 7.2 10.8









#90
Benzo(a)pyrene
Concen: 5.126 ng
RT: 15.498 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 94454
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.0
125 9.5 7.6 11.4



#91
Dibenzo(a,h)anthracene
Concen: 4.749 ng
RT: 17.086 min Scan# 2549
Delta R.T. -0.012 min
Lab File: BF140232.D
Acq: 05 Nov 2024 12:52

Tgt Ion:278 Resp: 84741
Ion Ratio Lower Upper
278 100
139 13.7 11.1 16.7
279 23.2 19.1 28.7





#92
 Benzo(g,h,i)perylene
 Concen: 4.814 ng
 RT: 17.515 min Scan# 20
 Delta R.T. -0.024 min
 Lab File: BF140232.D
 Acq: 05 Nov 2024 12:52

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Tgt Ion:	276	Resp:	87630
Ion	Ratio	Lower	Upper
276	100		
277	22.5	18.8	28.2
138	20.4	15.3	22.9

