

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110617\
 Data File : BF100260.D
 Acq On : 6 Nov 2017 19:33
 Operator : SJ/JU
 Sample : I6181-01 5X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WASTE

Manual Integrations
 APPROVED

Sohil
 11/7/2017 5:39:51 PM

Quant Time: Nov 07 01:35:22 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 06 13:31:35 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.76	152	101613	20.00	ng	0.00
21) Naphthalene-d8	8.05	136	385144	20.00	ng	0.00
38) Acenaphthene-d10	9.80	164	176109	20.00	ng	0.00
63) Phenanthrene-d10	11.29	188	297765	20.00	ng	0.00
75) Chrysene-d12	13.94	240	180764	20.00	ng	0.00
86) Perylene-d12	15.40	264	168159	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	57388	8.87	ng	0.06
7) Phenol-d6	6.45	99	80935	10.55	ng	0.03
23) Nitrobenzene-d5	7.35	82	53781	8.99	ng	0.01
41) 2,4,6-Tribromophenol	10.60	330	16499	10.28	ng	0.00
44) 2-Fluorobiphenyl	9.12	172	108399	9.50	ng	0.00
78) Terphenyl-d14	12.87	244	91991	11.12	ng	0.00
Target Compounds						
31) Naphthalene	8.07	128	111909	6.019	ng	Qvalue # 94
37) 2-Methylnaphthalene	8.76	142	414083	33.236	ng	98
45) 1,1'-Biphenyl	9.22	154	79947	5.018	ng	96
51) Acenaphthene	9.83	154	26847	2.466	ng	# 61
57) Fluorene	10.35	166	74803	5.878	ng	# 90
70) Phenanthrene	11.32	178	195303	11.622	ng	95
77) Pyrene	12.75	202	30316	2.206	ng	# 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF110617\
Data File : BF100260.D
Acq On : 6 Nov 2017 19:33
Operator : SJ/JU
Sample : I6181-01 5X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WASTE

Manual Integrations
APPROVED

Sohil
11/7/2017 5:39:51 PM

Quant Time: Nov 07 01:35:22 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 06 13:31:35 2017
Response via : Initial Calibration

