


Data Path : Z:\HPCHEM1\BNA_F\Data\BF110617\
 Data File : BF100253.D
 Acq On : 6 Nov 2017 16:10
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF102317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 07 14:09:40 2017



AutoFind: Scans 1603, 1604, 1605; Background Corrected with Scan 1598

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.9	23192	PASS
68	69	0.00	2	1.7	404	PASS
69	198	0.00	100	30.5	24457	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.4	36410	PASS
197	198	0.00	2	0.3	217	PASS
198	198	100	100	100.0	80280	PASS
199	198	5	9	7.0	5614	PASS
275	198	10	60	23.1	18541	PASS
365	198	1	100	2.5	2035	PASS
441	443	0.01	100	83.6	11845	PASS
442	198	50	100	95.8	76869	PASS
443	442	15	24	18.4	14170	PASS