

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110624\
 Data File : BF140253.D
 Acq On : 06 Nov 2024 12:42
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 06 13:15:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 05 16:47:56 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	84	0.00
2	1,4-Dioxane	40.000	39.760	0.6	85	0.00
3	Pyridine	40.000	39.819	0.5	85	0.00
4	n-Nitrosodimethylamine	40.000	37.023	7.4	78	-0.01
5 S	2-Fluorophenol	80.000	76.338	4.6	84	0.00
6	Aniline	40.000	38.971	2.6	84	0.00
7 S	Phenol-d6	80.000	76.154	4.8	83	0.00
8	2-Chlorophenol	40.000	39.068	2.3	86	0.00
9	Benzaldehyde	40.000	36.419	9.0	77	0.00
10 C	Phenol	40.000	38.490	3.8	84	0.00
11	bis(2-Chloroethyl)ether	40.000	38.928	2.7	85	0.00
12	1,3-Dichlorobenzene	40.000	38.683	3.3	85	0.00
13 C	1,4-Dichlorobenzene	40.000	38.716	3.2	86	0.00
14	1,2-Dichlorobenzene	40.000	38.692	3.3	87	0.00
15	Benzyl Alcohol	40.000	38.159	4.6	81	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.458	6.4	82	0.00
17	2-Methylphenol	40.000	38.345	4.1	83	0.00
18	Hexachloroethane	40.000	39.126	2.2	86	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.321	6.7	82	0.00
20	3+4-Methylphenols	40.000	38.697	3.3	85	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	37.214	7.0	85	0.00
23 S	Nitrobenzene-d5	80.000	73.892	7.6	83	0.00
24	Nitrobenzene	40.000	37.129	7.2	83	0.00
25	Isophorone	40.000	36.912	7.7	83	0.00
26 C	2-Nitrophenol	40.000	38.369	4.1	82	0.00
27	2,4-Dimethylphenol	40.000	38.177	4.6	85	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.529	6.2	85	0.00
29 C	2,4-Dichlorophenol	40.000	38.536	3.7	86	0.00
30	1,2,4-Trichlorobenzene	40.000	38.120	4.7	86	0.00
31	Naphthalene	40.000	37.499	6.3	85	0.00
32	Benzoic acid	40.000	37.052	7.4	78	-0.01
33	4-Chloroaniline	40.000	38.494	3.8	87	0.00
34 C	Hexachlorobutadiene	40.000	37.901	5.2	86	0.00
35	Caprolactam	40.000	38.009	5.0	84	-0.01
36 C	4-Chloro-3-methylphenol	40.000	38.202	4.5	84	0.00
37	2-Methylnaphthalene	40.000	37.706	5.7	86	0.00
38	1-Methylnaphthalene	40.000	37.574	6.1	86	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.279	4.3	88	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.209	-5.5	83	0.00
42 S	2,4,6-Tribromophenol	80.000	79.541	0.6	90	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.267	1.8	87	0.00
44	2,4,5-Trichlorophenol	40.000	38.532	3.7	87	0.00
45 S	2-Fluorobiphenyl	80.000	73.527	8.1	86	0.00
46	1,1'-Biphenyl	40.000	37.380	6.5	87	0.00
47	2-Chloronaphthalene	40.000	37.782	5.5	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110624\
 Data File : BF140253.D
 Acq On : 06 Nov 2024 12:42
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 06 13:15:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 05 16:47:56 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.443	6.4	82	0.00
49	Acenaphthylene	40.000	37.661	5.8	86	0.00
50	Dimethylphthalate	40.000	37.797	5.5	86	0.00
51	2,6-Dinitrotoluene	40.000	38.985	2.5	87	0.00
52 C	Acenaphthene	40.000	38.118	4.7	87	0.00
53	3-Nitroaniline	40.000	38.356	4.1	86	0.00
54 P	2,4-Dinitrophenol	40.000	35.454	11.4	78	0.00
55	Dibenzofuran	40.000	37.451	6.4	86	0.00
56 P	4-Nitrophenol	40.000	39.566	1.1	82	0.00
57	2,4-Dinitrotoluene	40.000	39.491	1.3	87	0.00
58	Fluorene	40.000	37.094	7.3	86	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.998	-2.5	93	0.00
60	Diethylphthalate	40.000	37.959	5.1	86	0.00
61	4-Chlorophenyl-phenylether	40.000	37.988	5.0	88	0.00
62	4-Nitroaniline	40.000	39.176	2.1	87	0.00
63	Azobenzene	40.000	37.258	6.9	86	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.725	3.2	83	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.709	5.7	88	0.00
67	4-Bromophenyl-phenylether	40.000	38.892	2.8	90	0.00
68	Hexachlorobenzene	40.000	38.601	3.5	90	0.00
69	Atrazine	40.000	43.239	-8.1	98	0.00
70 C	Pentachlorophenol	40.000	41.162	-2.9	88	0.00
71	Phenanthrene	40.000	37.264	6.8	87	0.00
72	Anthracene	40.000	37.760	5.6	89	0.00
73	Carbazole	40.000	37.915	5.2	89	0.00
74	Di-n-butylphthalate	40.000	38.733	3.2	90	0.00
75 C	Fluoranthene	40.000	38.300	4.3	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	91	0.00
77	Benzydine	40.000	39.504	1.2	72	0.00
78	Pyrene	40.000	38.073	4.8	90	0.00
79 S	Terphenyl-d14	80.000	75.752	5.3	91	0.00
80	Butylbenzylphthalate	40.000	39.013	2.5	88	0.00
81	Benzo(a)anthracene	40.000	39.020	2.4	92	0.00
82	3,3'-Dichlorobenzidine	40.000	38.178	4.6	84	0.00
83	Chrysene	40.000	37.634	5.9	89	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.940	2.7	90	0.00
85 c	Di-n-octyl phthalate	40.000	38.763	3.1	87	0.01
86 I	Perylene-d12	20.000	20.000	0.0	88	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	38.928	2.7	84	0.01
88	Benzo(b)fluoranthene	40.000	40.424	-1.1	86	0.01
89	Benzo(k)fluoranthene	40.000	36.756	8.1	87	0.01
90 C	Benzo(a)pyrene	40.000	39.145	2.1	86	0.01
91	Dibenzo(a,h)anthracene	40.000	38.278	4.3	83	0.01
92	Benzo(g,h,i)perylene	40.000	37.898	5.3	82	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110624\
Data File : BF140253.D
Acq On : 06 Nov 2024 12:42
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 06 13:15:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 05 16:47:56 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0