

Data Path : Z:\HPCHEM1\BNA F\DATA\BF110816\
 Data File : BF091110.D
 Acq On : 9 Nov 2016 5:05
 Operator : UM/SJ
 Sample : H5593-06
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EAST-SIDEWALL

Quant Time: Nov 09 06:19:46 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 12:59:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	183124	20.00	ng	0.00
21) Naphthalene-d8	7.95	136	721200	20.00	ng	0.00
38) Acenaphthene-d10	9.70	164	352033	20.00	ng	0.00
63) Phenanthrene-d10	11.17	188	451712	20.00	ng	0.00
75) Chrysene-d12	13.80	240	345274	20.00	ng	-0.01
86) Perylene-d12	15.19	264	308735	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.25	112	1113980	100.47	ng	0.01
7) Phenol-d6	6.32	99	1405511	93.39	ng	0.00
23) Nitrobenzene-d5	7.23	82	905293	68.48	ng	-0.01
41) 2,4,6-Tribromophenol	10.49	330	258770	81.31	ng	0.00
44) 2-Fluorobiphenyl	9.02	172	1157340	56.60	ng	-0.01
78) Terphenyl-d14	12.76	244	854325	61.75	ng	0.00
Target Compounds						
10) Phenol	6.33	94	80308	4.82	ng	Qvalue # 66
49) Dimethylphthalate	9.42	163	214957	9.31	ng	98
70) Phenanthrene	11.20	178	107313	4.47	ng	97
74) Fluoranthene	12.38	202	85287	3.42	ng	92
77) Pyrene	12.61	202	75560	3.34	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF110816\
Data File : BF091110.D
Acq On : 9 Nov 2016 5:05
Operator : UM/SJ
Sample : H5593-06
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EAST-SIDEWALL

Quant Time: Nov 09 06:19:46 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 08 12:59:29 2016
Response via : Initial Calibration

