

Data Path : Z:\HPCHEM1\BNA F\DATA\BF110816\
 Data File : BF091114.D
 Acq On : 9 Nov 2016 6:55
 Operator : UM/SJ
 Sample : H5547-01
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WATER-QUALITYDRY-1

Quant Time: Nov 09 12:17:52 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 12:59:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	160761	20.00	ng	0.00
21) Naphthalene-d8	7.95	136	631147	20.00	ng	0.00
38) Acenaphthene-d10	9.70	164	302004	20.00	ng	0.00
63) Phenanthrene-d10	11.17	188	444547	20.00	ng	0.00
75) Chrysene-d12	13.80	240	316780	20.00	ng	-0.01
86) Perylene-d12	15.17	264	269927	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	488520	50.19	ng	0.00
7) Phenol-d6	6.30	99	381861	28.90	ng	-0.01
23) Nitrobenzene-d5	7.23	82	1003895	86.77	ng	-0.01
41) 2,4,6-Tribromophenol	10.49	330	343968	125.98	ng	0.00
44) 2-Fluorobiphenyl	9.02	172	1597746	91.08	ng	-0.01
78) Terphenyl-d14	12.76	244	1276944	100.59	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.42	163	69525	3.51	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF110816\
Data File : BF091114.D
Acq On : 9 Nov 2016 6:55
Operator : UM/SJ
Sample : H5547-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WATER-QUALITYDRY-1

Quant Time: Nov 09 12:17:52 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 08 12:59:29 2016
Response via : Initial Calibration

