

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110817\  
 Data File : BF100294.D  
 Acq On : 8 Nov 2017 16:28  
 Operator : SJ/JU  
 Sample : I6341-01  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 A6835DW-A6865DW

Quant Time: Nov 08 17:29:24 2017  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Nov 08 14:33:31 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.90  | 152  | 186002   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.18  | 136  | 741885   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 9.93  | 164  | 358057   | 20.00  | ng    | 0.00     |
| 63) Phenanthrene-d10        | 11.42 | 188  | 693191   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 14.06 | 240  | 503405   | 20.00  | ng    | -0.01    |
| 86) Perylene-d12            | 15.54 | 264  | 400061   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.51  | 112  | 642482   | 55.37  | ng    | 0.00     |
| 7) Phenol-d6                | 6.51  | 99   | 515671   | 36.04  | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.46  | 82   | 1052831  | 93.82  | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 10.73 | 330  | 507380   | 139.06 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.26  | 172  | 2055031  | 87.39  | ng    | 0.00     |
| 78) Terphenyl-d14           | 13.02 | 244  | 2083647  | 95.10  | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF110817\  
Data File : BF100294.D  
Acq On : 8 Nov 2017 16:28  
Operator : SJ/JU  
Sample : I6341-01  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
A6835DW-A6865DW

Quant Time: Nov 08 17:29:24 2017  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Nov 08 14:33:31 2017  
Response via : Initial Calibration

